

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072221\
 Data File : BN015643.D
 Acq On : 21 Jul 2021 13:39
 Operator : CG/JU
 Sample : SSTDICC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Jul 21 15:32:26 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 21 14:10:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.744	152	12316	0.400	ng	0.00
7) Naphthalene-d8	10.521	136	54909	0.400	ng	0.00
13) Acenaphthene-d10	14.370	164	29432	0.400	ng	0.00
19) Phenanthrene-d10	17.127	188	56335	0.400	ng	0.00
29) Chrysene-d12	21.323	240	55203	0.400	ng	0.00
35) Perylene-d12	23.635	264	51288	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.346	112	14628	0.414	ng	0.00
5) Phenol-d6	6.914	99	17465	0.416	ng	0.00
8) Nitrobenzene-d5	8.886	82	14671	0.371	ng	0.00
11) 2-Methylnaphthalene-d10	12.118	152	34851	0.386	ng	0.00
14) 2,4,6-Tribromophenol	15.867	330	4671	0.438	ng	0.00
15) 2-Fluorobiphenyl	13.000	172	45463	0.383	ng	0.00
27) Fluoranthene-d10	19.162	212	64973	0.382	ng	0.00
31) Terphenyl-d14	19.768	244	48986	0.374	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.281	88	1555	0.258	ng	96
3) n-Nitrosodimethylamine	3.650	42	3970	0.391	ng	100
6) bis(2-Chloroethyl)ether	7.172	93	14961	0.398	ng	100
9) Naphthalene	10.571	128	64143	0.402	ng	100
10) Hexachlorobutadiene	10.863	225	11452	0.385	ng	# 100
12) 2-Methylnaphthalene	12.189	142	40254	0.398	ng	100
16) Acenaphthylene	14.091	152	50633	0.370	ng	100
17) Acenaphthene	14.434	154	35052	0.383	ng	100
18) Fluorene	15.423	166	42893	0.391	ng	100
20) 4,6-Dinitro-2-methylph...	15.512	198	1076	0.474	ng	100
21) 4-Bromophenyl-phenylether	16.324	248	13159	0.368	ng	100
22) Hexachlorobenzene	16.433	284	14852	0.369	ng	100
23) Atrazine	16.592	200	9434	0.365	ng	100
24) Pentachlorophenol	16.786	266	1987	0.231	ng	100
25) Phenanthrene	17.164	178	69605	0.382	ng	100
26) Anthracene	17.261	178	58464	0.374	ng	100
28) Fluoranthene	19.192	202	79040	0.401	ng	100
30) Pyrene	19.554	202	78016	0.368	ng	100
32) Benzo(a)anthracene	21.302	228	75172	0.390	ng	100
33) Chrysene	21.355	228	78271	0.383	ng	100
34) Bis(2-ethylhexyl)phtha...	21.249	149	36682	0.444	ng	100
36) Indeno(1,2,3-cd)pyrene	26.007	276	80180	0.438	ng	100
37) Benzo(b)fluoranthene	22.937	252	79346	0.394	ng	100
38) Benzo(k)fluoranthene	22.981	252	76522	0.377	ng	100
39) Benzo(a)pyrene	23.532	252	66514	0.384	ng	100
40) Dibenzo(a,h)anthracene	26.024	278	65217	0.453	ng	100
41) Benzo(g,h,i)perylene	26.733	276	70015	0.451	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072221\
 Data File : BN015643.D
 Acq On : 21 Jul 2021 13:39
 Operator : CG/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Jul 21 15:32:26 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 21 14:10:52 2021
 Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.744 min Scan# 506
 Delta R.T. -0.000 min
 Lab File: BN015643.D
 Acq: 21 Jul 2021 13:39




Instrument :
 BNA_N
 ClientSampled :


Tgt Ion:	152	Resp:	12316
Ion	Ratio	Lower	Upper
152	100		
150	154.3	123.4	185.2
115	58.8	47.0	70.6



1,4-Dioxane
 Concen: 0.258 ng
 RT: 3.281 min Scan# 22
 Delta R.T. -0.000 min
 Lab File: BN015643.D
 Acq: 21 Jul 2021 13:39

Tgt Ion:	88	Resp:	1555
Ion	Ratio	Lower	Upper
88	100		
43	131.8	105.4	158.0
58	80.8	72.6	108.8









n-Nitrosodimethylamine
Concen: 0.391 ng
RT: 3.650 min Scan# 62
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39

Instrument :
BNA_N
ClientSampled :

Tgt Ion: 42 Resp: 3970




Ion	Ratio	Lower	Upper
42	100		
74	154.4	123.8	185.8
44	7.3	5.9	8.9




2-Fluorophenol
Concen: 0.414 ng
RT: 5.346 min Scan# 246
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39

Tgt Ion: 112 Resp: 14628

Ion	Ratio	Lower	Upper
112	100		
64	55.3	44.2	66.4
63	27.6	22.1	33.1









Phenol-d6
 Concen: 0.416 ng
 RT: 6.914 min Scan# 416
 Delta R.T. 0.000 min
 Lab File: BN015643.D
 Acq: 21 Jul 2021 13:39




Instrument :
 BNA_N
 ClientSampled :

Tgt Ion:	99	Resp:	17465
Ion	Ratio	Lower	Upper
99	100		
42	16.3	13.0	19.6
71	30.5	24.4	36.6



bis(2-Chloroethyl)ether
 Concen: 0.398 ng
 RT: 7.172 min Scan# 444
 Delta R.T. 0.000 min
 Lab File: BN015643.D
 Acq: 21 Jul 2021 13:39


Tgt Ion:	93	Resp:	14961
Ion	Ratio	Lower	Upper
93	100		
63	73.9	59.1	88.7
95	32.9	26.3	39.5






Abundance Scan 742 (10.521 min): BN015643.D\data.ms (-735) (-)

Naphthalene-d8
Concen: 0.400 ng
RT: 10.521 min Scan# 742
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39

Instrument :
BNA_N
ClientSampled :






Tgt Ion:	136	Resp:	54909
Ion Ratio	Lower	Upper	
136	100		
137	11.0	8.8	13.2
54	7.1	5.7	8.5
68	6.0	4.8	7.2




Abundance Scan 613 (8.886 min): BN015643.D\data.ms (-610) (-)

Nitrobenzene-d5
Concen: 0.371 ng
RT: 8.886 min Scan# 613
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39

Tgt Ion:	82	Resp:	14671
Ion Ratio	Lower	Upper	
82	100		
128	40.3	32.2	48.4
54	48.8	39.0	58.6









Naphthalene
 Concen: 0.402 ng
 RT: 10.571 min Scan# 746
 Delta R.T. 0.000 min
 Lab File: BN015643.D
 Acq: 21 Jul 2021 13:39




Instrument :
 BNA_N
 ClientSampleId :

Tgt Ion:	128	Resp:	64143
Ion	Ratio	Lower	Upper
128	100		
129	10.9	8.7	13.1
127	13.0	10.4	15.6




Hexachlorobutadiene
 Concen: 0.385 ng
 RT: 10.863 min Scan# 769
 Delta R.T. 0.000 min
 Lab File: BN015643.D
 Acq: 21 Jul 2021 13:39

Tgt Ion:	225	Resp:	11452
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.9	51.1	76.7






Abundance Scan 957 (12.118 min): BN015643.D\data.ms (-946) (-)




2-Methylnaphthalene-d10
Concen: 0.386 ng
RT: 12.118 min Scan# 957
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39

Tgt Ion:152 Resp: 34851
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.4

Instrument :
BNA_N
ClientSampleId :






Abundance Scan 973 (12.189 min): BN015643.D\data.ms (-963) (-)



2-Methylnaphthalene
Concen: 0.398 ng
RT: 12.189 min Scan# 973
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39


Tgt Ion:142 Resp: 40254
Ion Ratio Lower Upper
142 100
141 88.9 71.1 106.7
115 33.5 26.8 40.2





Abundance Scan 1191 (14.370 min): BN015643.D\data.ms (-1191) (-)

Acenaphthene-d10
Concen: 0.400 ng
RT: 14.370 min Scan# 1191
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39


Instrument :
BNA_N
ClientSampleId :



Tgt Ion:	164	Resp:	29432
Ion Ratio	Lower	Upper	
164	100		
162	103.0	82.4	123.6
160	46.6	37.3	55.9





Abundance Scan 1191 (14.370 min): BN015643.D\data.ms (-1175) (-)




Abundance Scan 1309 (15.867 min): BN015643.D\data.ms (-1309) (-)

2,4,6-Tribromophenol
Concen: 0.438 ng
RT: 15.867 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39

Tgt Ion:	330	Resp:	4671
Ion Ratio	Lower	Upper	
330	100		
332	0.0	0.0	0.0
141	36.3	29.0	43.6




Abundance Scan 1309 (15.867 min): BN015643.D\data.ms (-1301) (-)





Abundance Scan 1083 (13.000 min): BN015643.D\data.ms (-1075) (-)

2-Fluorobiphenyl
Concen: 0.383 ng
RT: 13.000 min Scan# 1083
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39


Instrument :
BNA_N
ClientSampleId :



Tgt Ion:	172	Resp:	45463
Ion Ratio	Lower	Upper	
172	100		
171	34.0	27.2	40.8
170	22.2	17.8	26.6





Abundance Scan 1083 (13.000 min): BN015643.D\data.ms (-1075) (-)




Abundance Scan 1169 (14.091 min): BN015643.D\data.ms (-1153) (-)

Acenaphthylene
Concen: 0.370 ng
RT: 14.091 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39

Tgt Ion:	152	Resp:	50633
Ion Ratio	Lower	Upper	
152	100		
151	19.5	15.6	23.4
153	12.9	10.3	15.5




Abundance Scan 1169 (14.091 min): BN015643.D\data.ms (-1153) (-)






Abundance Scan 1196 (14.434 min): BN015643.D\data.ms (-1190) (-)

Acenaphthene
Concen: 0.383 ng
RT: 14.434 min Scan# 1196
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39

Instrument :
BNA_N
ClientSampled :







Tgt Ion:	154	Resp:	35052
Ion Ratio	Lower	Upper	
154	100		
153	112.7	90.2	135.2
152	54.6	43.7	65.5



Abundance Scan 1274 (15.423 min): BN015643.D\data.ms (-1269) (-)

Fluorene
Concen: 0.391 ng
RT: 15.423 min Scan# 1274
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39


Tgt Ion:	166	Resp:	42893
Ion Ratio	Lower	Upper	
166	100		
165	98.3	78.6	118.0
167	13.4	10.7	16.1





Abundance Scan 1412 (17.127 min): BN015643.D\data.ms (-1396) (-)

Phenanthrene-d10
Concen: 0.400 ng
RT: 17.127 min Scan# 1412
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39


Instrument :
BNA_N
ClientSampled :



Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	6.6	5.3	7.9





Abundance Scan 1412 (17.127 min): BN015643.D\data.ms (-1396) (-)




Abundance Scan 1281 (15.512 min): BN015643.D\data.ms (-1273) (-)

4,6-Dinitro-2-methylphenol
Concen: 0.474 ng
RT: 15.512 min Scan# 1281
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39


Tgt Ion	Ratio	Lower	Upper
198	100		
182	17.5	14.0	21.0
77	0.0	0.0	0.0



Abundance Scan 1281 (15.512 min): BN015643.D\data.ms (-1273) (-)





Abundance Scan 1346 (16.324 min): BN015643.D\data.ms (-1321) (-)




4-Bromophenyl-phenylether
Concen: 0.368 ng
RT: 16.324 min Scan# 1346
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39

Instrument :
BNA_N
ClientSampleId :


Tgt Ion:	248	Resp:	13159
Ion	Ratio	Lower	Upper
248	100		
250	102.6	82.1	123.1
141	35.6	28.5	42.7



Abundance Scan 1346 (16.324 min): BN015643.D\data.ms (-1338) (-)





Abundance Scan 1355 (16.433 min): BN015643.D\data.ms (-1322) (-)




Hexachlorobenzene
Concen: 0.369 ng
RT: 16.433 min Scan# 1355
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39


Tgt Ion:	284	Resp:	14852
Ion	Ratio	Lower	Upper
284	100		
142	33.9	27.1	40.7
249	30.9	24.7	37.1



Abundance Scan 1355 (16.433 min): BN015643.D\data.ms (-1347) (-)






Abundance Scan 1368 (16.592 min): BN015643.D\data.ms (-1323) (-)




Atrazine
Concen: 0.365 ng
RT: 16.592 min Scan# 1368
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39

Instrument :
BNA_N
ClientSampleId :

Tgt Ion:	200	Resp:	9434
Ion Ratio	Lower	Upper	
200	100		
173	26.5	21.2	31.8
215	48.5	38.8	58.2






Abundance Scan 1384 (16.786 min): BN015643.D\data.ms (-1324) (-)




Pentachlorophenol
Concen: 0.231 ng
RT: 16.786 min Scan# 1384
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39

Tgt Ion:	266	Resp:	1987
Ion Ratio	Lower	Upper	
266	100		
264	62.8	50.2	75.4
268	64.5	51.6	77.4






Abundance Scan 1415 (17.164 min): BN015643.D\data.ms (-1425) (-)




Phenanthrene
Concen: 0.382 ng
RT: 17.164 min Scan# 1415
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39

Instrument :
BNA_N
ClientSampled :

Tgt Ion:	178	Resp:	69605
Ion Ratio	Lower	Upper	
178	100		
176	18.8	15.0	22.6
179	15.4	12.3	18.5






Abundance Scan 1423 (17.261 min): BN015643.D\data.ms (-1425) (-)



Anthracene
Concen: 0.374 ng
RT: 17.261 min Scan# 1423
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39


Tgt Ion:	178	Resp:	58464
Ion Ratio	Lower	Upper	
178	100		
176	18.1	14.5	21.7
179	15.3	12.2	18.4






Abundance Scan 1716 (19.162 min): BN015643.D\data.ms (-1727) (-)

Fluoranthene-d10
Concen: 0.382 ng
RT: 19.162 min Scan# 1716
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39

Instrument :
BNA_N
ClientSampleId :




Tgt Ion:	212	Resp:	64973
Ion Ratio	Lower	Upper	
212	100		
106	14.1	11.3	16.9
104	7.6	6.1	9.1






Abundance Scan 1722 (19.192 min): BN015643.D\data.ms (-1728) (-)

Fluoranthene
Concen: 0.401 ng
RT: 19.192 min Scan# 1722
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39




Tgt Ion:	202	Resp:	79040
Ion Ratio	Lower	Upper	
202	100		
101	11.6	9.3	13.9
203	17.2	13.8	20.6






Abundance Scan 2010 (21.323 min): BN015643.D\data.ms (-2024) (-)

Chrysene-d12
Concen: 0.400 ng
RT: 21.323 min Scan# 2010
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39

Instrument :
BNA_N
ClientSampleId :




Tgt Ion:	240	Resp:	55203
Ion	Ratio	Lower	Upper
240	100		
120	8.9	7.1	10.7
236	25.9	20.7	31.1






Abundance Scan 1795 (19.554 min): BN015643.D\data.ms (-1736) (-)

Pyrene
Concen: 0.368 ng
RT: 19.554 min Scan# 1795
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39




Tgt Ion:	202	Resp:	78016
Ion	Ratio	Lower	Upper
202	100		
200	20.5	16.4	24.6
203	17.4	13.9	20.9





Abundance Scan 1838 (19.768 min): BN015643.D\data.ms (-1824) (-)

Terphenyl-d14
Concen: 0.374 ng
RT: 19.768 min Scan# 1838
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39


Instrument :
BNA_N
ClientSampled :



Tgt Ion	Ratio	Lower	Upper
244	100		
212	7.6	6.1	9.1
122	11.8	9.4	14.2






Abundance Scan 1838 (19.768 min): BN015643.D\data.ms (-1808) (-)




Abundance Scan 2008 (21.302 min): BN015643.D\data.ms (-2032) (-)

Benzo(a)anthracene
Concen: 0.390 ng
RT: 21.302 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39


Tgt Ion	Ratio	Lower	Upper
228	100		
226	27.6	22.1	33.1
229	19.1	15.3	22.9



Abundance Scan 2008 (21.302 min): BN015643.D\data.ms (-1989) (-)






Abundance Scan 2013 (21.355 min): BN015643.D\data.ms (-2039) (-)




Chrysene
Concen: 0.383 ng
RT: 21.355 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39

Instrument :
BNA_N
ClientSampleId :

Tgt Ion:	228	Resp:	78271
Ion	Ratio	Lower	Upper
228	100		
226	30.4	24.3	36.5
229	19.4	15.5	23.3






Abundance Scan 2003 (21.249 min): BN015643.D\data.ms (-1994) (-)



Bis(2-ethylhexyl)phthalate
Concen: 0.444 ng
RT: 21.249 min Scan# 2003
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39


Tgt Ion:	149	Resp:	36682
Ion	Ratio	Lower	Upper
149	100		
167	28.7	23.0	34.4
279	3.6	2.9	4.3






Abundance Scan 2471 (23.635 min): BN015643.D\data.ms (-2427) (-)

Perylene-d12
Concen: 0.400 ng
RT: 23.635 min Scan# 2471
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39

Instrument :
BNA_N
ClientSampleId :







Tgt Ion:	264	Resp:	51288
Ion	Ratio	Lower	Upper
264	100		
260	24.6	19.7	29.5
265	25.0	20.0	30.0




Abundance Scan 3164 (26.007 min): BN015643.D\data.ms (-3135) (-)

Indeno(1,2,3-cd)pyrene
Concen: 0.438 ng
RT: 26.007 min Scan# 3164
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39

Tgt Ion:	276	Resp:	80180
Ion	Ratio	Lower	Upper
276	100		
138	27.3	21.8	32.6
277	25.0	20.0	30.0






Abundance Scan 2267 (22.937 min): BN015643.D\data.ms (-2267) (-)




Benzo(b)fluoranthene
Concen: 0.394 ng
RT: 22.937 min Scan# 2267
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39

Tgt Ion:252 Resp: 79346
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.8
125 11.1 8.9 13.3

Instrument :
BNA_N
ClientSampleId :






Abundance Scan 2280 (22.981 min): BN015643.D\data.ms (-2280) (-)



Benzo(k)fluoranthene
Concen: 0.377 ng
RT: 22.981 min Scan# 2280
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39


Tgt Ion:252 Resp: 76522
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.6
125 11.1 8.9 13.3






Abundance Scan 2441 (23.532 min): BN015643.D\data.ms (-2383) (-)

Benzo(a)pyrene
Concen: 0.384 ng
RT: 23.532 min Scan# 2441
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39

Instrument :
BNA_N
ClientSampleId :




Tgt Ion:	252	Resp:	66514
Ion Ratio	Lower	Upper	
252	100		
253	22.0	17.6	26.4
125	12.6	10.1	15.1






Abundance Scan 3169 (26.024 min): BN015643.D\data.ms (-3146) (-)


Dibenzo(a,h)anthracene
Concen: 0.453 ng
RT: 26.024 min Scan# 3169
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39



Tgt Ion:	278	Resp:	65217
Ion Ratio	Lower	Upper	
278	100		
139	18.4	14.7	22.1
279	23.8	19.0	28.6




Abundance Scan 3376 (26.733 min): BN015643.D\data.ms (-3340) (-)



Benzo(g,h,i)perylene
Concen: 0.451 ng
RT: 26.733 min Scan# 3376
Delta R.T. 0.000 min
Lab File: BN015643.D
Acq: 21 Jul 2021 13:39

Instrument :
BNA_N
ClientSampleId :



Tgt Ion:	276	Resp:	70015
Ion Ratio	Lower	Upper	
276	100		
277	23.8	19.0	28.6
138	23.1	18.5	27.7

Abundance Scan 3376 (26.733 min): BN015643.D\data.ms (-3318) (-)

