

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN072420\
 Data File : BN011235.D
 Acq On : 24 Jul 2020 10:54
 Operator : CG/JU
 Sample : PB130415BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK15

Quant Time: Jul 24 11:25:22 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN071520MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jul 24 11:01:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.48	152	209476	20.00	ng/ul	0.00
18) Naphthalene-d8	10.23	136	837641	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.11	164	533647	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.87	188	1215984	20.00	ng/ul	0.00
78) Chrysene-d12	21.08	240	1138113	20.00	ng/ul	0.00
86) Perylene-d12	23.22	264	978596	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.14	96	34976	6.38	ng/uL	0.00
5) Phenol-d5	6.68	99	594854	35.72	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.84	67	306949	34.48	ng/ul	0.00
9) 2-Chlorophenol-d4	7.03	132	485246	33.35	ng/ul	0.00
13) 4-Methylphenol-d8	8.19	113	475999	34.65	ng/ul	0.00
19) Nitrobenzene-d5	8.62	128	218991	32.19	ng/ul	0.00
22) 2-Nitrophenol-d4	9.33	143	238227	32.97	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.86	165	462247	32.98	ng/ul	0.00
29) 4-Chloroaniline-d4	10.38	131	629956	42.66	ng/ul	0.00
44) Dimethylphthalate-d6	13.54	166	1419580	32.80	ng/ul	0.00
47) Acenaphthylene-d8	13.80	160	1646824	32.06	ng/ul	0.00
52) 4-Nitrophenol-d4	14.33	143	199627	25.82	ng/ul	0.00
58) Fluorene-d10	15.12	176	1242998	32.48	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.25	200	203912	27.48	ng/ul	0.00
71) Anthracene-d10	16.96	188	1980590	33.46	ng/ul	0.00
79) Pyrene-d10	19.27	212	2190824	38.61	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.08	264	1768150	33.35	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN072420\
Data File : BN011235.D
Acq On : 24 Jul 2020 10:54
Operator : CG/JU
Sample : PB130415BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK15

Quant Time: Jul 24 11:25:22 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN071520MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jul 24 11:01:43 2020
Response via : Initial Calibration

