

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN072420\
 Data File : BN011250.D
 Acq On : 24 Jul 2020 19:53
 Operator : CG/JU
 Sample : PB130437BL
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK37

Quant Time: Jul 25 00:43:24 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN071520MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jul 24 11:01:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.48	152	264791	20.00	ng/ul	0.00
18) Naphthalene-d8	10.23	136	1049332	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.11	164	667115	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.87	188	1406961	20.00	ng/ul	0.00
78) Chrysene-d12	21.08	240	1060156	20.00	ng/ul	0.00
86) Perylene-d12	23.21	264	957917	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.15	96	41810	6.03	ng/uL	0.00
5) Phenol-d5	6.68	99	737584	35.04	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.83	67	388344	34.52	ng/ul	0.00
9) 2-Chlorophenol-d4	7.02	132	618844	33.64	ng/ul	0.00
13) 4-Methylphenol-d8	8.19	113	609173	35.08	ng/ul	0.00
19) Nitrobenzene-d5	8.62	128	278535	32.68	ng/ul	0.00
22) 2-Nitrophenol-d4	9.33	143	316728	34.99	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.86	165	580213	33.05	ng/ul	0.00
29) 4-Chloroaniline-d4	10.38	131	776996	42.00	ng/ul	0.00
44) Dimethylphthalate-d6	13.53	166	1780295	32.91	ng/ul	0.00
47) Acenaphthylene-d8	13.80	160	2083908	32.46	ng/ul	0.00
52) 4-Nitrophenol-d4	14.33	143	239004	24.73	ng/ul	0.00
58) Fluorene-d10	15.12	176	1503938	31.43	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.25	200	248451	28.94	ng/ul	0.00
71) Anthracene-d10	16.96	188	2276460	33.24	ng/ul	0.00
79) Pyrene-d10	19.27	212	2270958	42.96	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.07	264	1729057	33.31	ng/ul	-0.01

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN072420\
Data File : BN011250.D
Acq On : 24 Jul 2020 19:53
Operator : CG/JU
Sample : PB130437BL
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK37

Quant Time: Jul 25 00:43:24 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN071520MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jul 24 11:01:43 2020
Response via : Initial Calibration

