

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072421\
 Data File : BN015666.D
 Acq On : 23 Jul 2021 12:37
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Jul 23 13:07:02 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 23 12:32:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.744	152	13004	0.400	ng	0.00
7) Naphthalene-d8	10.521	136	54881	0.400	ng	0.00
13) Acenaphthene-d10	14.370	164	28745	0.400	ng	0.00
19) Phenanthrene-d10	17.115	188	53213	0.400	ng	0.00
29) Chrysene-d12	21.323	240	51933	0.400	ng	# 0.01
35) Perylene-d12	23.628	264	43157	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.347	112	52311	1.396	ng	0.00
5) Phenol-d6	6.914	99	63330	1.420	ng	0.00
8) Nitrobenzene-d5	8.886	82	64351	1.632	ng	0.00
11) 2-Methylnaphthalene-d10	12.114	152	136112	1.491	ng	0.00
14) 2,4,6-Tribromophenol	15.868	330	18743	1.452	ng	0.00
15) 2-Fluorobiphenyl	13.000	172	183884	1.609	ng	0.00
27) Fluoranthene-d10	19.157	212	245908	1.505	ng	0.00
31) Terphenyl-d14	19.763	244	193453	1.603	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.281	88	6480	1.476	ng	98
3) n-Nitrosodimethylamine	3.641	42	16727	1.556	ng	100
6) bis(2-Chloroethyl)ether	7.172	93	61237	1.528	ng	100
9) Naphthalene	10.572	128	238603	1.520	ng	100
10) Hexachlorobutadiene	10.863	225	45025	1.530	ng	# 99
12) 2-Methylnaphthalene	12.190	142	156389	1.534	ng	100
16) Acenaphthylene	14.091	152	214028	1.614	ng	100
17) Acenaphthene	14.434	154	138221	1.535	ng	98
18) Fluorene	15.423	166	170129	1.560	ng	99
20) 4,6-Dinitro-2-methylph...	15.512	198	7055	1.611	ng	# 76
21) 4-Bromophenyl-phenylether	16.312	248	53064	1.593	ng	97
22) Hexachlorobenzene	16.434	284	57956	1.573	ng	100
23) Atrazine	16.592	200	37534	1.540	ng	99
24) Pentachlorophenol	16.774	266	14436	1.560	ng	99
25) Phenanthrene	17.164	178	267899	1.556	ng	100
26) Anthracene	17.249	178	238164	1.617	ng	100
28) Fluoranthene	19.187	202	304174	1.607	ng	100
30) Pyrene	19.550	202	300736	1.535	ng	100
32) Benzo(a)anthracene	21.302	228	281904	1.511	ng	100
33) Chrysene	21.355	228	296791	1.554	ng	100
34) Bis(2-ethylhexyl)phtha...	21.249	149	108550	1.225	ng	100
36) Indeno(1,2,3-cd)pyrene	25.997	276	271897	1.530	ng	100
37) Benzo(b)fluoranthene	22.930	252	282564	1.699	ng	100
38) Benzo(k)fluoranthene	22.975	252	274946	1.612	ng	99
39) Benzo(a)pyrene	23.526	252	234091	1.605	ng	98
40) Dibenzo(a,h)anthracene	26.014	278	217770	1.539	ng	99
41) Benzo(g,h,i)perylene	26.723	276	242922	1.590	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072421\
Data File : BN015666.D
Acq On : 23 Jul 2021 12:37
Operator : CG/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Quant Time: Jul 23 13:07:02 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 23 12:32:24 2021
Response via : Initial Calibration






Abundance Scan 506 (7.744 min): BN015665.D\data.ms (-499) (#1)




1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.744 min Scan# 506
Delta R.T. -0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:152 Resp: 13004
Ion Ratio Lower Upper
152 100
150 155.3 123.4 185.0
115 58.5 47.2 70.8





Abundance Scan 22 (3.281 min): BN015665.D\data.ms (-19) (#2)





1,4-Dioxane
Concen: 1.476 ng
RT: 3.281 min Scan# 22
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Tgt Ion: 88 Resp: 6480
Ion Ratio Lower Upper
88 100
43 28.3 23.5 35.3
58 78.0 64.2 96.4



Abundance Scan 22 (3.281 min): BN015666.D\data.ms (-11) (-)









n-Nitrosodimethylamine
 Concen: 1.556 ng
 RT: 3.641 min Scan# 61
 Delta R.T. -0.009 min
 Lab File: BN015666.D
 Acq: 23 Jul 2021 12:37




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:	42	Resp:	16727
Ion	Ratio	Lower	Upper
42	100		
74	151.2	121.0	181.6
44	7.5	6.2	9.4



2-Fluorophenol
 Concen: 1.396 ng
 RT: 5.347 min Scan# 246
 Delta R.T. 0.000 min
 Lab File: BN015666.D
 Acq: 23 Jul 2021 12:37

Tgt Ion:	112	Resp:	52311
Ion	Ratio	Lower	Upper
112	100		
64	56.2	44.9	67.3
63	28.0	22.6	33.8









Phenol-d6
 Concen: 1.420 ng
 RT: 6.914 min Scan# 416
 Delta R.T. 0.000 min
 Lab File: BN015666.D
 Acq: 23 Jul 2021 12:37




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Tgt Ion:	99	Resp:	63330
Ion	Ratio	Lower	Upper
99	100		
42	16.8	13.8	20.6
71	30.0	24.2	36.2



bis(2-Chloroethyl)ether
 Concen: 1.528 ng
 RT: 7.172 min Scan# 444
 Delta R.T. 0.000 min
 Lab File: BN015666.D
 Acq: 23 Jul 2021 12:37


Tgt Ion:	93	Resp:	61237
Ion	Ratio	Lower	Upper
93	100		
63	74.7	59.9	89.9
95	32.8	26.3	39.5






Abundance Scan 742 (10.521 min): BN015665.D\data.ms (-737) (-)

Naphthalene-d8
Concen: 0.400 ng
RT: 10.521 min Scan# 742
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Instrument :
BNA_N
ClientSampled :
SSTDICC1.6






Tgt Ion:	136	Resp:	54881
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.7	13.1
54	6.6	5.2	7.8
68	5.6	4.4	6.6




Abundance Scan 613 (8.886 min): BN015665.D\data.ms (-610) (-)

Nitrobenzene-d5
Concen: 1.632 ng
RT: 8.886 min Scan# 613
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Tgt Ion:	82	Resp:	64351
Ion	Ratio	Lower	Upper
82	100		
128	41.3	31.2	46.8
54	47.2	40.2	60.4






Abundance Scan 746 (10.571 min): BN015665.D\data.ms (-74#9-)




Naphthalene
Concen: 1.520 ng
RT: 10.572 min Scan# 746
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:	128	Resp:	238603
Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.8	13.2
127	12.7	10.2	15.4






Abundance Scan 769 (10.863 min): BN015665.D\data.ms (-76#10-)




Hexachlorobutadiene
Concen: 1.530 ng
RT: 10.863 min Scan# 769
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Tgt Ion:	225	Resp:	45025
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.9	50.6	76.0






Abundance Scan 956 (12.114 min): BN015665.D\data.ms (-943) #11




2-Methylnaphthalene-d10
Concen: 1.491 ng
RT: 12.114 min Scan# 956
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:152 Resp: 136112
Ion Ratio Lower Upper
152 100
151 20.4 16.4 24.6






Abundance Scan 973 (12.190 min): BN015665.D\data.ms (-963) #12




2-Methylnaphthalene
Concen: 1.534 ng
RT: 12.190 min Scan# 973
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Tgt Ion:142 Resp: 156389
Ion Ratio Lower Upper
142 100
141 86.7 69.5 104.3
115 32.6 26.1 39.1





Abundance Scan 1191 (14.370 min): BN015665.D\data.ms (-1191) (-)




Acenaphthene-d10
Concen: 0.400 ng
RT: 14.370 min Scan# 1191
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	164	Resp:	28745
Ion	Ratio	Lower	Upper
164	100		
162	102.6	80.6	120.8
160	46.1	37.0	55.4



Abundance Scan 1191 (14.370 min): BN015666.D\data.ms (-1175) (-)





Abundance Scan 1309 (15.867 min): BN015665.D\data.ms (-1309) (-)




2,4,6-Tribromophenol
Concen: 1.452 ng
RT: 15.868 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37


Tgt Ion:	330	Resp:	18743
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	35.8	28.7	43.1



Abundance Scan 1309 (15.868 min): BN015666.D\data.ms (-1301) (-)






Abundance Scan 1083 (13.000 min): BN015665.D\data.ms (-1075) (-)




2-Fluorobiphenyl
Concen: 1.609 ng
RT: 13.000 min Scan# 1083
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:	172	Resp:	183884
Ion	Ratio	Lower	Upper
172	100		
171	33.5	26.6	40.0
170	21.6	17.2	25.8






Abundance Scan 1169 (14.091 min): BN015665.D\data.ms (-1153) (-)




Acenaphthylene
Concen: 1.614 ng
RT: 14.091 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Tgt Ion:	152	Resp:	214028
Ion	Ratio	Lower	Upper
152	100		
151	19.2	15.4	23.2
153	13.1	10.3	15.5






Abundance Scan 1196 (14.434 min): BN015665.D\data.ms (-1197) (-)




Acenaphthene
Concen: 1.535 ng
RT: 14.434 min Scan# 1196
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:154 Resp: 138221
Ion Ratio Lower Upper
154 100
153 112.0 90.6 136.0
152 54.3 44.6 67.0






Abundance Scan 1274 (15.423 min): BN015665.D\data.ms (-1275) (-)



Fluorene
Concen: 1.560 ng
RT: 15.423 min Scan# 1274
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37


Tgt Ion:166 Resp: 170129
Ion Ratio Lower Upper
166 100
165 97.1 78.3 117.5
167 13.4 10.8 16.2






Abundance Scan 1411 (17.115 min): BN015665.D\data.ms (-1419) (-)

Phenanthrene-d10
Concen: 0.400 ng
RT: 17.115 min Scan# 1411
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6




Tgt Ion:	188	Resp:	53213
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	12.1	9.2	13.8






Abundance Scan 1281 (15.512 min): BN015665.D\data.ms (-1279) (-)


4,6-Dinitro-2-methylphenol
Concen: 1.611 ng
RT: 15.512 min Scan# 1281
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37



Tgt Ion:	198	Resp:	7055
Ion	Ratio	Lower	Upper
198	100		
182	10.0	17.2	25.8#
77	0.0	0.0	0.0





Abundance Scan 1345 (16.312 min): BN015665.D\data.ms (-1320) (-)




4-Bromophenyl-phenylether
Concen: 1.593 ng
RT: 16.312 min Scan# 1345
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	248	Resp:	53064
Ion	Ratio	Lower	Upper
248	100		
250	91.1	72.9	109.3
141	68.8	59.1	88.7



Abundance Scan 1345 (16.312 min): BN015666.D\data.ms (-1337) (-)





Abundance Scan 1355 (16.434 min): BN015665.D\data.ms (-1320) (-)




Hexachlorobenzene
Concen: 1.573 ng
RT: 16.434 min Scan# 1355
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Tgt Ion:	284	Resp:	57956
Ion	Ratio	Lower	Upper
284	100		
142	34.4	27.5	41.3
249	31.3	25.0	37.6




Abundance Scan 1355 (16.434 min): BN015666.D\data.ms (-1347) (-)






Abundance Scan 1368 (16.592 min): BN015665.D\data.ms (-1327) (-)

Atrazine
Concen: 1.540 ng
RT: 16.592 min Scan# 1368
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6




Tgt Ion:	200	Resp:	37534
Ion	Ratio	Lower	Upper
200	100		
173	22.4	17.8	26.8
215	48.9	39.8	59.8






Abundance Scan 1383 (16.774 min): BN015665.D\data.ms (-1327) (-)

Pentachlorophenol
Concen: 1.560 ng
RT: 16.774 min Scan# 1383
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37




Tgt Ion:	266	Resp:	14436
Ion	Ratio	Lower	Upper
266	100		
264	63.1	51.0	76.4
268	63.9	51.8	77.8






Abundance Scan 1415 (17.164 min): BN015665.D\data.ms (-1425) (-)

Phenanthrene
Concen: 1.556 ng
RT: 17.164 min Scan# 1415
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6




Tgt Ion:	178	Resp:	267899
Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.0	22.6
179	15.3	12.4	18.6






Abundance Scan 1422 (17.249 min): BN015665.D\data.ms (-1425) (-)


Anthracene
Concen: 1.617 ng
RT: 17.249 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37



Tgt Ion:	178	Resp:	238164
Ion	Ratio	Lower	Upper
178	100		
176	18.1	14.6	21.8
179	15.3	12.2	18.2





Abundance Scan 1715 (19.157 min): BN015665.D\data.ms (-1#25) (-)




Fluoranthene-d10
Concen: 1.505 ng
RT: 19.157 min Scan# 1715
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:212 Resp: 245908
Ion Ratio Lower Upper
212 100
106 14.5 11.5 17.3
104 7.9 6.2 9.4



Abundance Scan 1715 (19.157 min): BN015666.D\data.ms (-1695) (-)





Abundance Scan 1721 (19.187 min): BN015665.D\data.ms (-1#23) (-)




Fluoranthene
Concen: 1.607 ng
RT: 19.187 min Scan# 1721
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Tgt Ion:202 Resp: 304174
Ion Ratio Lower Upper
202 100
101 11.9 9.5 14.3
203 17.3 13.8 20.6




Abundance Scan 1721 (19.187 min): BN015666.D\data.ms (-1681) (-)






Abundance Scan 2009 (21.312 min): BN015665.D\data.ms (-1990) (-)

Chrysene-d12
Concen: 0.400 ng
RT: 21.323 min Scan# 2010
Delta R.T. 0.011 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6







Tgt Ion:	240	Resp:	51933
Ion	Ratio	Lower	Upper
240	100		
120	5.8	12.2	18.4#
236	25.2	21.5	32.3




Abundance Scan 1794 (19.549 min): BN015665.D\data.ms (-1734) (-)

Pyrene
Concen: 1.535 ng
RT: 19.550 min Scan# 1794
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Tgt Ion:	202	Resp:	300736
Ion	Ratio	Lower	Upper
202	100		
200	20.6	16.4	24.6
203	17.4	13.9	20.9






Abundance Scan 1837 (19.763 min): BN015665.D\data.ms (-1807149)




Terphenyl-d14
Concen: 1.603 ng
RT: 19.763 min Scan# 1837
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:244 Resp: 193453
Ion Ratio Lower Upper
244 100
212 7.6 6.1 9.1
122 12.7 10.1 15.1






Abundance Scan 2008 (21.302 min): BN015665.D\data.ms (-1989159)




Benzo(a)anthracene
Concen: 1.511 ng
RT: 21.302 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Tgt Ion:228 Resp: 281904
Ion Ratio Lower Upper
228 100
226 26.4 21.4 32.0
229 19.6 15.5 23.3






Abundance Scan 2013 (21.355 min): BN015665.D\data.ms (-2013) (-)




Chrysene
Concen: 1.554 ng
RT: 21.355 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:	228	Resp:	296791
Ion	Ratio	Lower	Upper
228	100		
226	29.1	23.5	35.3
229	19.6	15.6	23.4






Abundance Scan 2002 (21.238 min): BN015665.D\data.ms (-1936) (-)



Bis(2-ethylhexyl)phthalate
Concen: 1.225 ng
RT: 21.249 min Scan# 2003
Delta R.T. 0.011 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37


Tgt Ion:	149	Resp:	108550
Ion	Ratio	Lower	Upper
149	100		
167	28.0	22.2	33.4
279	3.6	2.8	4.2





Abundance Scan 2469 (23.628 min): BN015665.D\data.ms (-2425) (-)

Perylene-d12
Concen: 0.400 ng
RT: 23.628 min Scan# 2469
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37


Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6



Tgt Ion:	264	Resp:	43157
Ion	Ratio	Lower	Upper
264	100		
260	24.8	19.8	29.8
265	24.3	20.0	30.0






Abundance Scan 2469 (23.628 min): BN015666.D\data.ms (-2425) (-)




Abundance Scan 3161 (25.997 min): BN015665.D\data.ms (-3132) (-)

Indeno(1,2,3-cd)pyrene
Concen: 1.530 ng
RT: 25.997 min Scan# 3161
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37


Tgt Ion:	276	Resp:	271897
Ion	Ratio	Lower	Upper
276	100		
138	27.4	21.8	32.6
277	25.0	19.9	29.9



Abundance Scan 3161 (25.997 min): BN015666.D\data.ms (-3132) (-)






Abundance Scan 2265 (22.930 min): BN015666.D\data.ms (-22071.6)




Benzo(b)fluoranthene
Concen: 1.699 ng
RT: 22.930 min Scan# 2265
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:252 Resp: 282564
Ion Ratio Lower Upper
252 100
253 22.1 17.6 26.4
125 10.8 8.8 13.2






Abundance Scan 2278 (22.974 min): BN015666.D\data.ms (-21901.6)



Benzo(k)fluoranthene
Concen: 1.612 ng
RT: 22.975 min Scan# 2278
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37


Tgt Ion:252 Resp: 274946
Ion Ratio Lower Upper
252 100
253 21.5 17.5 26.3
125 10.8 8.8 13.2





Abundance Scan 2439 (23.525 min): BN015665.D\data.ms (-2381) (-)

Benzo(a)pyrene
Concen: 1.605 ng
RT: 23.526 min Scan# 2439
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37


Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6



Tgt Ion:	252	Resp:	234091
Ion	Ratio	Lower	Upper
252	100		
253	21.6	18.0	27.0
125	12.0	10.4	15.6




Abundance Scan 2439 (23.526 min): BN015666.D\data.ms (-2381) (-)





Abundance Scan 3166 (26.014 min): BN015665.D\data.ms (-3143) (-)


Dibenzo(a,h)anthracene
Concen: 1.539 ng
RT: 26.014 min Scan# 3166
Delta R.T. 0.000 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37




Tgt Ion:	278	Resp:	217770
Ion	Ratio	Lower	Upper
278	100		
139	18.2	14.6	22.0
279	23.6	19.3	28.9



Abundance Scan 3166 (26.014 min): BN015666.D\data.ms (-3143) (-)




Abundance Scan 3372 (26.719 min): BN015665.D\data.ms (-3341) (-)




Benzo(g,h,i)perylene
Concen: 1.590 ng
RT: 26.723 min Scan# 3373
Delta R.T. 0.004 min
Lab File: BN015666.D
Acq: 23 Jul 2021 12:37

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Abundance Scan 3373 (26.723 min): BN015666.D\data.ms



Tgt Ion:	276	Resp:	242922
Ion	Ratio	Lower	Upper
276	100		
277	23.8	19.1	28.7
138	23.5	19.1	28.7



Abundance Scan 3373 (26.723 min): BN015666.D\data.ms (-3314) (-)

