

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072421\
 Data File : BN015668.D
 Acq On : 23 Jul 2021 13:50
 Operator : CG/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Quant Time: Jul 23 14:38:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 23 14:37:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.744	152	13991	0.400	ng	0.00
7) Naphthalene-d8	10.521	136	55320	0.400	ng	0.00
13) Acenaphthene-d10	14.370	164	29612	0.400	ng	0.00
19) Phenanthrene-d10	17.115	188	55719	0.400	ng	0.00
29) Chrysene-d12	21.323	240	53602	0.400	ng	# 0.01
35) Perylene-d12	23.628	264	48653	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.347	112	168735	5.011	ng	0.00
5) Phenol-d6	6.914	99	205195	5.194	ng	0.00
8) Nitrobenzene-d5	8.886	82	215931	5.571	ng	0.00
11) 2-Methylnaphthalene-d10	12.114	152	400429	4.856	ng	0.00
14) 2,4,6-Tribromophenol	15.867	330	70317	7.632	ng	0.00
15) 2-Fluorobiphenyl	13.000	172	543655	4.504	ng	0.00
27) Fluoranthene-d10	19.157	212	772010	5.193	ng	0.00
31) Terphenyl-d14	19.763	244	584635	4.583	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.281	88	20131	4.823	ng	# 94
3) n-Nitrosodimethylamine	3.632	42	54387	5.275	ng	97
6) bis(2-Chloroethyl)ether	7.172	93	179914	4.534	ng	99
9) Naphthalene	10.572	128	693018	4.720	ng	100
10) Hexachlorobutadiene	10.863	225	131575	4.715	ng	# 99
12) 2-Methylnaphthalene	12.190	142	445632	4.706	ng	100
16) Acenaphthylene	14.091	152	675811	5.464	ng	100
17) Acenaphthene	14.434	154	417807	4.860	ng	96
18) Fluorene	15.423	166	503545	4.910	ng	99
20) 4,6-Dinitro-2-methylph...	15.512	198	37148	9.819	ng	# 66
21) 4-Bromophenyl-phenylether	16.312	248	166049	5.121	ng	98
22) Hexachlorobenzene	16.434	284	173499	4.705	ng	100
23) Atrazine	16.592	200	140512	7.277	ng	98
24) Pentachlorophenol	16.774	266	71013	8.501	ng	99
25) Phenanthrene	17.164	178	801415	4.766	ng	100
26) Anthracene	17.249	178	756741	5.416	ng	100
28) Fluoranthene	19.187	202	887492	4.890	ng	100
30) Pyrene	19.550	202	920771	4.783	ng	100
32) Benzo(a)anthracene	21.302	228	924873	5.506	ng	100
33) Chrysene	21.355	228	862282	4.633	ng	100
34) Bis(2-ethylhexyl)phtha...	21.249	149	542625	8.659	ng	99
36) Indeno(1,2,3-cd)pyrene	26.000	276	924614	5.245	ng	100
37) Benzo(b)fluoranthene	22.930	252	903434	4.938	ng	100
38) Benzo(k)fluoranthene	22.975	252	863577	4.714	ng	100
39) Benzo(a)pyrene	23.526	252	802101	5.410	ng	98
40) Dibenzo(a,h)anthracene	26.018	278	739845	5.337	ng	99
41) Benzo(g,h,i)perylene	26.726	276	823892	5.104	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072421\
Data File : BN015668.D
Acq On : 23 Jul 2021 13:50
Operator : CG/JU
Sample : SSTDIC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Quant Time: Jul 23 14:38:29 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 23 14:37:10 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.744 min Scan# 506
 Delta R.T. -0.000 min
 Lab File: BN015668.D
 Acq: 23 Jul 2021 13:50




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:152 Resp: 13991
 Ion Ratio Lower Upper
 152 100
 150 160.3 123.4 185.0
 115 59.0 47.2 70.8



1,4-Dioxane
 Concen: 4.823 ng
 RT: 3.281 min Scan# 22
 Delta R.T. 0.000 min
 Lab File: BN015668.D
 Acq: 23 Jul 2021 13:50

Tgt Ion: 88 Resp: 20131
 Ion Ratio Lower Upper
 88 100
 43 22.6 23.5 35.3#
 58 77.6 64.2 96.4









n-Nitrosodimethylamine
Concen: 5.275 ng
RT: 3.632 min Scan# 60
Delta R.T. -0.018 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	42	Resp:	54387
Ion	Ratio	Lower	Upper
42	100		
74	154.8	121.0	181.6
44	6.8	6.2	9.4



2-Fluorophenol
Concen: 5.011 ng
RT: 5.347 min Scan# 246
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion:	112	Resp:	168735
Ion	Ratio	Lower	Upper
112	100		
64	56.1	44.9	67.3
63	27.8	22.6	33.8









Phenol-d6
 Concen: 5.194 ng
 RT: 6.914 min Scan# 416
 Delta R.T. 0.000 min
 Lab File: BN015668.D
 Acq: 23 Jul 2021 13:50




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Tgt Ion:	99	Resp:	205195
Ion	Ratio	Lower	Upper
99	100		
42	16.6	13.8	20.6
71	30.1	24.2	36.2




bis(2-Chloroethyl)ether
 Concen: 4.534 ng
 RT: 7.172 min Scan# 444
 Delta R.T. 0.000 min
 Lab File: BN015668.D
 Acq: 23 Jul 2021 13:50

Tgt Ion:	93	Resp:	179914
Ion	Ratio	Lower	Upper
93	100		
63	74.2	59.9	89.9
95	32.8	26.3	39.5






Abundance Scan 742 (10.521 min): BN015665.D\data.ms (-737) (-)




Naphthalene-d8
Concen: 0.400 ng
RT: 10.521 min Scan# 742
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:	136	Resp:	55320
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.7	13.1
54	6.6	5.2	7.8
68	5.5	4.4	6.6





Abundance Scan 613 (8.886 min): BN015665.D\data.ms (-610) (-)




Nitrobenzene-d5
Concen: 5.571 ng
RT: 8.886 min Scan# 613
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50


Tgt Ion:	82	Resp:	215931
Ion	Ratio	Lower	Upper
82	100		
128	44.1	31.2	46.8
54	45.8	40.2	60.4



Abundance Scan 613 (8.886 min): BN015668.D\data.ms (-597) (-)






Abundance Scan 746 (10.571 min): BN015665.D\data.ms (-746) (-)




Naphthalene
Concen: 4.720 ng
RT: 10.572 min Scan# 746
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:128 Resp: 693018
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 12.7 10.2 15.4






Abundance Scan 769 (10.863 min): BN015665.D\data.ms (-769) (-)




Hexachlorobutadiene
Concen: 4.715 ng
RT: 10.863 min Scan# 769
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion:225 Resp: 131575
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.8 50.6 76.0






Abundance Scan 956 (12.114 min): BN015665.D\data.ms (-943) #11




2-Methylnaphthalene-d10
Concen: 4.856 ng
RT: 12.114 min Scan# 956
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:152 Resp: 400429
Ion Ratio Lower Upper
152 100
151 20.3 16.4 24.6






Abundance Scan 973 (12.190 min): BN015665.D\data.ms (-963) #12




2-Methylnaphthalene
Concen: 4.706 ng
RT: 12.190 min Scan# 973
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion:142 Resp: 445632
Ion Ratio Lower Upper
142 100
141 87.3 69.5 104.3
115 32.5 26.1 39.1





Abundance Scan 1191 (14.370 min): BN015665.D\data.ms (-1191) (-)




Acenaphthene-d10
Concen: 0.400 ng
RT: 14.370 min Scan# 1191
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:164 Resp: 29612
Ion Ratio Lower Upper
164 100
162 102.3 80.6 120.8
160 46.1 37.0 55.4



Abundance Scan 1191 (14.370 min): BN015668.D\data.ms (-1175) (-)





Abundance Scan 1309 (15.867 min): BN015665.D\data.ms (-1309) (-)




2,4,6-Tribromophenol
Concen: 7.632 ng
RT: 15.867 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50


Tgt Ion:330 Resp: 70317
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 35.6 28.7 43.1



Abundance Scan 1309 (15.867 min): BN015668.D\data.ms (-1301) (-)






Abundance Scan 1083 (13.000 min): BN015665.D\data.ms (-1715) (-)




2-Fluorobiphenyl
Concen: 4.504 ng
RT: 13.000 min Scan# 1083
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:	172	Resp:	543655
Ion	Ratio	Lower	Upper
172	100		
171	33.3	26.6	40.0
170	21.4	17.2	25.8






Abundance Scan 1169 (14.091 min): BN015665.D\data.ms (-1115) (-)




Acenaphthylene
Concen: 5.464 ng
RT: 14.091 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion:	152	Resp:	675811
Ion	Ratio	Lower	Upper
152	100		
151	19.3	15.4	23.2
153	13.1	10.3	15.5






Abundance Scan 1196 (14.434 min): BN015665.D\data.ms (-1196) (-)




Acenaphthene
Concen: 4.860 ng
RT: 14.434 min Scan# 1196
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:154 Resp: 417807
Ion Ratio Lower Upper
154 100
153 109.8 90.6 136.0
152 52.2 44.6 67.0






Abundance Scan 1274 (15.423 min): BN015665.D\data.ms (-1274) (-)



Fluorene
Concen: 4.910 ng
RT: 15.423 min Scan# 1274
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50


Tgt Ion:166 Resp: 503545
Ion Ratio Lower Upper
166 100
165 96.8 78.3 117.5
167 13.6 10.8 16.2





Abundance Scan 1411 (17.115 min): BN015665.D\data.ms (-1419) (-)

Phenanthrene-d10
Concen: 0.400 ng
RT: 17.115 min Scan# 1411
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50


Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0



Tgt Ion:	188	Resp:	55719
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	12.5	9.2	13.8





Abundance Scan 1411 (17.115 min): BN015668.D\data.ms (-1395) (-)




Abundance Scan 1281 (15.512 min): BN015665.D\data.ms (-1279) (-)

4,6-Dinitro-2-methylphenol
Concen: 9.819 ng
RT: 15.512 min Scan# 1281
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50


Tgt Ion:	198	Resp:	37148
Ion	Ratio	Lower	Upper
198	100		
182	5.6	17.2	25.8#
77	0.0	0.0	0.0



Abundance Scan 1281 (15.512 min): BN015668.D\data.ms (-1273) (-)






Abundance Scan 1345 (16.312 min): BN015665.D\data.ms (-1320) (-)




4-Bromophenyl-phenylether
Concen: 5.121 ng
RT: 16.312 min Scan# 1345
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:	248	Resp:	166049
Ion	Ratio	Lower	Upper
248	100		
250	92.7	72.9	109.3
141	72.4	59.1	88.7






Abundance Scan 1355 (16.434 min): BN015665.D\data.ms (-1320) (-)



Hexachlorobenzene
Concen: 4.705 ng
RT: 16.434 min Scan# 1355
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50


Tgt Ion:	284	Resp:	173499
Ion	Ratio	Lower	Upper
284	100		
142	34.2	27.5	41.3
249	31.1	25.0	37.6





Abundance Scan 1368 (16.592 min): BN015665.D\data.ms (-1325) (-)

Atrazine
Concen: 7.277 ng
RT: 16.592 min Scan# 1368
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50


Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0



Tgt Ion:	200	Resp:	140512
Ion	Ratio	Lower	Upper
200	100		
173	23.7	17.8	26.8
215	48.8	39.8	59.8




Abundance Scan 1368 (16.592 min): BN015668.D\data.ms (-1327) (-)





Abundance Scan 1383 (16.774 min): BN015665.D\data.ms (-1327) (-)

Pentachlorophenol
Concen: 8.501 ng
RT: 16.774 min Scan# 1383
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50


Tgt Ion:	266	Resp:	71013
Ion	Ratio	Lower	Upper
266	100		
264	63.0	51.0	76.4
268	63.7	51.8	77.8



Abundance Scan 1383 (16.774 min): BN015668.D\data.ms (-1375) (-)






Abundance Scan 1415 (17.164 min): BN015665.D\data.ms (-1425) (-)




Phenanthrene
Concen: 4.766 ng
RT: 17.164 min Scan# 1415
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion:	178	Resp:	801415
Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.0	22.6
179	15.4	12.4	18.6

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0






Abundance Scan 1422 (17.249 min): BN015665.D\data.ms (-1425) (-)




Anthracene
Concen: 5.416 ng
RT: 17.249 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion:	178	Resp:	756741
Ion	Ratio	Lower	Upper
178	100		
176	18.3	14.6	21.8
179	15.5	12.2	18.2





Abundance Scan 1715 (19.157 min): BN015665.D\data.ms (-1725) (-)




Fluoranthene-d10
Concen: 5.193 ng
RT: 19.157 min Scan# 1715
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:212 Resp: 772010
Ion Ratio Lower Upper
212 100
106 14.6 11.5 17.3
104 8.0 6.2 9.4



Abundance Scan 1715 (19.157 min): BN015668.D\data.ms (-1695) (-)





Abundance Scan 1721 (19.187 min): BN015665.D\data.ms (-1723) (-)




Fluoranthene
Concen: 4.890 ng
RT: 19.187 min Scan# 1721
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50


Tgt Ion:202 Resp: 887492
Ion Ratio Lower Upper
202 100
101 11.9 9.5 14.3
203 17.5 13.8 20.6



Abundance Scan 1721 (19.187 min): BN015668.D\data.ms (-1681) (-)






Abundance Scan 2009 (21.312 min): BN015665.D\data.ms (-1799) (-)




Chrysene-d12
Concen: 0.400 ng
RT: 21.323 min Scan# 2010
Delta R.T. 0.011 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:	240	Resp:	53602
Ion	Ratio	Lower	Upper
240	100		
120	6.0	12.2	18.4#
236	25.6	21.5	32.3






Abundance Scan 1794 (19.549 min): BN015665.D\data.ms (-1784) (-)



Pyrene
Concen: 4.783 ng
RT: 19.550 min Scan# 1794
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50


Tgt Ion:	202	Resp:	920771
Ion	Ratio	Lower	Upper
202	100		
200	20.6	16.4	24.6
203	17.6	13.9	20.9






Abundance Scan 1837 (19.763 min): BN015665.D\data.ms (-1879) (-)

Terphenyl-d14
Concen: 4.583 ng
RT: 19.763 min Scan# 1837
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0







Tgt Ion:	244	Resp:	584635
Ion	Ratio	Lower	Upper
244	100		
212	7.6	6.1	9.1
122	12.6	10.1	15.1




Abundance Scan 2008 (21.302 min): BN015665.D\data.ms (-1879) (-)

Benzo(a)anthracene
Concen: 5.506 ng
RT: 21.302 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion:	228	Resp:	924873
Ion	Ratio	Lower	Upper
228	100		
226	26.6	21.4	32.0
229	19.7	15.5	23.3









Scan 2013 (21.355 min): BN015665.D\data.ms (-2030) (-)
Chrysene
 Concen: 4.633 ng
 RT: 21.355 min Scan# 2013
 Delta R.T. 0.000 min
 Lab File: BN015668.D
 Acq: 23 Jul 2021 13:50




Instrument :
 BNA_N
ClientSampleId :
 SSTDICC5.0

Tgt Ion:	228	Resp:	862282
Ion	Ratio	Lower	Upper
228	100		
226	29.5	23.5	35.3
229	19.7	15.6	23.4



Scan 2002 (21.238 min): BN015665.D\data.ms (-1936) (-)
Bis(2-ethylhexyl)phthalate
 Concen: 8.659 ng
 RT: 21.249 min Scan# 2003
 Delta R.T. 0.011 min
 Lab File: BN015668.D
 Acq: 23 Jul 2021 13:50


Tgt Ion:	149	Resp:	542625
Ion	Ratio	Lower	Upper
149	100		
167	28.3	22.2	33.4
279	3.6	2.8	4.2






Abundance Scan 2469 (23.628 min): BN015665.D\data.ms (-2425) (-)

Perylene-d12
Concen: 0.400 ng
RT: 23.628 min Scan# 2469
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Instrument :
BNA_N
ClientSampled :
SSTDICC5.0




Tgt Ion:	264	Resp:	48653
Ion	Ratio	Lower	Upper
264	100		
260	24.8	19.8	29.8
265	24.0	20.0	30.0






Abundance Scan 3161 (25.997 min): BN015665.D\data.ms (-3134) (-)


Indeno(1,2,3-cd)pyrene
Concen: 5.245 ng
RT: 26.000 min Scan# 3162
Delta R.T. 0.004 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50



Tgt Ion:	276	Resp:	924614
Ion	Ratio	Lower	Upper
276	100		
138	27.5	21.8	32.6
277	25.0	19.9	29.9





Abundance Scan 2265 (22.930 min): BN015665.D\data.ms (-2230) (-)




Benzo(b)fluoranthene
Concen: 4.938 ng
RT: 22.930 min Scan# 2265
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	252	Resp:	903434
Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.6	26.4
125	10.8	8.8	13.2



Abundance Scan 2265 (22.930 min): BN015668.D\data.ms (-2207) (-)





Abundance Scan 2278 (22.974 min): BN015665.D\data.ms (-2230) (-)




Benzo(k)fluoranthene
Concen: 4.714 ng
RT: 22.975 min Scan# 2278
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion:	252	Resp:	863577
Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.5	26.3
125	10.8	8.8	13.2




Abundance Scan 2278 (22.975 min): BN015668.D\data.ms (-2190) (-)





Abundance Scan 2439 (23.525 min): BN015665.D\data.ms (-2381) (-)

Benzo(a)pyrene
Concen: 5.410 ng
RT: 23.526 min Scan# 2439
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50


Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0



Tgt Ion:	252	Resp:	802101
Ion	Ratio	Lower	Upper
252	100		
253	21.7	18.0	27.0
125	11.9	10.4	15.6





Abundance Scan 2439 (23.526 min): BN015668.D\data.ms (-2381) (-)




Abundance Scan 3166 (26.014 min): BN015665.D\data.ms (-3143) (-)

Dibenzo(a,h)anthracene
Concen: 5.337 ng
RT: 26.018 min Scan# 3167
Delta R.T. 0.004 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50


Tgt Ion:	278	Resp:	739845
Ion	Ratio	Lower	Upper
278	100		
139	18.4	14.6	22.0
279	23.3	19.3	28.9



Abundance Scan 3167 (26.018 min): BN015668.D\data.ms (-3143) (-)




Abundance Scan 3372 (26.719 min): BN015665.D\data.ms (-3372) (-)




Benzo(g,h,i)perylene
Concen: 5.104 ng
RT: 26.726 min Scan# 3374
Delta R.T. 0.007 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Abundance Scan 3374 (26.726 min): BN015668.D\data.ms



Tgt Ion:	276	Resp:	823892
Ion	Ratio	Lower	Upper
276	100		
277	23.5	19.1	28.7
138	23.6	19.1	28.7



Abundance Scan 3374 (26.726 min): BN015668.D\data.ms (-3314) (-)

