

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072421\
 Data File : BN015669.D
 Acq On : 23 Jul 2021 15:22
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN072421

Quant Time: Jul 23 15:54:47 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 23 14:45:01 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.119	0.118	0.8	102	0.00
3	n-Nitrosodimethylamine	0.295	0.278	5.8	100	0.00
4 S	2-Fluorophenol	0.963	0.924	4.0	103	0.00
5 S	Phenol-d6	1.130	1.065	5.8	103	0.00
6	bis(2-Chloroethyl)ether	1.134	1.129	0.4	99	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
8 S	Nitrobenzene-d5	0.280	0.244	12.9	98	0.00
9	Naphthalene	1.062	1.043	1.8	100	0.00
10	Hexachlorobutadiene	0.202	0.199	1.5	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.596	0.577	3.2	100	0.00
12	2-Methylnaphthalene	0.685	0.671	2.0	99	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.124	0.110	11.3	99	0.00
15 S	2-Fluorobiphenyl	1.631	1.562	4.2	100	0.00
16	Acenaphthylene	1.671	1.500	10.2	98	0.00
17	Acenaphthene	1.161	1.102	5.1	100	0.00
18	Fluorene	1.385	1.330	4.0	101	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
20	4,6-Dinitro-2-methylphenol	0.027	0.013	51.9#	97	0.00
21	4-Bromophenyl-phenylether	0.233	0.220	5.6	99	0.00
22	Hexachlorobenzene	0.265	0.263	0.8	100	0.00
23	Atrazine	0.139	0.127	8.6	102	0.00
24	Pentachlorophenol	0.060	0.036	40.0#	101	0.00
25	Phenanthrene	1.207	1.181	2.2	100	0.00
26	Anthracene	1.003	0.918	8.5	101	0.00
27 SURR	Fluoranthene-d10	1.067	0.993	6.9	100	0.00
28	Fluoranthene	1.303	1.261	3.2	100	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	101	0.01
30	Pyrene	1.437	1.386	3.5	101	0.00
31 S	Terphenyl-d14	0.952	0.913	4.1	100	0.00
32	Benzo(a)anthracene	1.254	1.132	9.7	104	0.00
33	Chrysene	1.389	1.363	1.9	99	0.00
34	Bis(2-ethylhexyl)phthalate	0.525	0.343	34.7#	103	0.01
35 I	Perylene-d12	1.000	1.000	0.0	103	0.00
36	Indeno(1,2,3-cd)pyrene	1.449	1.349	6.9	103	0.00
37	Benzo(b)fluoranthene	1.504	1.404	6.6	101	0.00
38	Benzo(k)fluoranthene	1.506	1.454	3.5	99	0.00
39 C	Benzo(a)pyrene	1.219	1.097	10.0	103	0.00
40	Dibenzo(a,h)anthracene	1.140	1.060	7.0	103	0.00
41	Benzo(g,h,i)perylene	1.327	1.229	7.4	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN072421