

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072421\
 Data File : BN015669.D
 Acq On : 23 Jul 2021 15:22
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN072421

Quant Time: Jul 23 15:54:47 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 23 14:45:01 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	99	0.00
2	1,4-Dioxane	0.400	0.394	1.5	102	0.00
3	n-Nitrosodimethylamine	0.400	0.378	5.5	100	0.00
4 S	2-Fluorophenol	0.400	0.384	4.0	103	0.00
5 S	Phenol-d6	0.400	0.377	5.8	103	0.00
6	bis(2-Chloroethyl)ether	0.400	0.398	0.5	99	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	99	0.00
8 S	Nitrobenzene-d5	0.400	0.348	13.0	98	0.00
9	Naphthalene	0.400	0.393	1.8	100	0.00
10	Hexachlorobutadiene	0.400	0.394	1.5	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.387	3.3	100	0.00
12	2-Methylnaphthalene	0.400	0.392	2.0	99	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.400	0.354	11.5	99	0.00
15 S	2-Fluorobiphenyl	0.400	0.383	4.3	100	0.00
16	Acenaphthylene	0.400	0.359	10.3	98	0.00
17	Acenaphthene	0.400	0.379	5.3	100	0.00
18	Fluorene	0.400	0.384	4.0	101	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	99	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.345	13.8	97	0.00
21	4-Bromophenyl-phenylether	0.400	0.377	5.8	99	0.00
22	Hexachlorobenzene	0.400	0.398	0.5	100	0.00
23	Atrazine	0.400	0.367	8.3	102	0.00
24	Pentachlorophenol	0.400	0.352	12.0	101	0.00
25	Phenanthrene	0.400	0.391	2.3	100	0.00
26	Anthracene	0.400	0.366	8.5	101	0.00
27 SURR	Fluoranthene-d10	0.400	0.372	7.0	100	0.00
28	Fluoranthene	0.400	0.387	3.3	100	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	101	0.01
30	Pyrene	0.400	0.386	3.5	101	0.00
31 S	Terphenyl-d14	0.400	0.384	4.0	100	0.00
32	Benzo(a)anthracene	0.400	0.361	9.8	104	0.00
33	Chrysene	0.400	0.393	1.8	99	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.407	-1.7	103	0.01
35 I	Perylene-d12	0.400	0.400	0.0	103	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.372	7.0	103	0.00
37	Benzo(b)fluoranthene	0.400	0.373	6.8	101	0.00
38	Benzo(k)fluoranthene	0.400	0.386	3.5	99	0.00
39 C	Benzo(a)pyrene	0.400	0.360	10.0	103	0.00
40	Dibenzo(a,h)anthracene	0.400	0.372	7.0	103	0.00
41	Benzo(g,h,i)perylene	0.400	0.370	7.5	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN072421