

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072421\
 Data File : BN015672.D
 Acq On : 23 Jul 2021 17:51
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 23 18:21:30 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 23 14:45:01 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	105	0.00
2	1,4-Dioxane	0.400	0.387	3.3	107	0.00
3	n-Nitrosodimethylamine	0.400	0.376	6.0	106	0.00
4 S	2-Fluorophenol	0.400	0.367	8.3	105	0.00
5 S	Phenol-d6	0.400	0.366	8.5	107	0.00
6	bis(2-Chloroethyl)ether	0.400	0.397	0.8	105	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	106	0.00
8 S	Nitrobenzene-d5	0.400	0.342	14.5	103	0.00
9	Naphthalene	0.400	0.391	2.3	106	0.00
10	Hexachlorobutadiene	0.400	0.393	1.8	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.384	4.0	105	0.00
12	2-Methylnaphthalene	0.400	0.392	2.0	106	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	106	0.00
14 S	2,4,6-Tribromophenol	0.400	0.379	5.3	112	0.00
15 S	2-Fluorobiphenyl	0.400	0.394	1.5	108	0.00
16	Acenaphthylene	0.400	0.368	8.0	106	0.00
17	Acenaphthene	0.400	0.391	2.3	108	0.00
18	Fluorene	0.400	0.391	2.3	108	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	106	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.375	6.3	125	0.00
21	4-Bromophenyl-phenylether	0.400	0.386	3.5	109	0.00
22	Hexachlorobenzene	0.400	0.409	-2.2	110	0.00
23	Atrazine	0.400	0.370	7.5	110	0.00
24	Pentachlorophenol	0.400	0.385	3.8	126	0.00
25	Phenanthrene	0.400	0.399	0.3	109	0.00
26	Anthracene	0.400	0.372	7.0	110	0.00
27 SURR	Fluoranthene-d10	0.400	0.368	8.0	106	0.00
28	Fluoranthene	0.400	0.396	1.0	110	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	107	0.00
30	Pyrene	0.400	0.396	1.0	110	0.00
31 S	Terphenyl-d14	0.400	0.395	1.3	110	0.00
32	Benzo(a)anthracene	0.400	0.364	9.0	111	0.00
33	Chrysene	0.400	0.410	-2.5	109	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.416	-4.0	114	0.00
35 I	Perylene-d12	0.400	0.400	0.0	120	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.410	-2.5	133	0.00
37	Benzo(b)fluoranthene	0.400	0.370	7.5	117	0.00
38	Benzo(k)fluoranthene	0.400	0.399	0.3	119	0.00
39 C	Benzo(a)pyrene	0.400	0.380	5.0	128	0.00
40	Dibenzo(a,h)anthracene	0.400	0.404	-1.0	131	0.00
41	Benzo(g,h,i)perylene	0.400	0.374	6.5	119	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4