

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072421\
 Data File : BN015668.D
 Acq On : 23 Jul 2021 13:50
 Operator : CG/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Quant Time: Jul 23 14:38:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 23 14:37:10 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.744	152	13991	0.400	ng	0.00
7) Naphthalene-d8	10.521	136	55320	0.400	ng	0.00
13) Acenaphthene-d10	14.370	164	29612	0.400	ng	0.00
19) Phenanthrene-d10	17.115	188	55719	0.400	ng	0.00
29) Chrysene-d12	21.323	240	53602	0.400	ng	# 0.01
35) Perylene-d12	23.628	264	48653	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.347	112	168735	5.011	ng	0.00
5) Phenol-d6	6.914	99	205195	5.194	ng	0.00
8) Nitrobenzene-d5	8.886	82	215931	5.571	ng	0.00
11) 2-Methylnaphthalene-d10	12.114	152	400429	4.856	ng	0.00
14) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
15) 2-Fluorobiphenyl	13.000	172	543655	4.504	ng	0.00
27) Fluoranthene-d10	19.157	212	772010	5.193	ng	0.00
31) Terphenyl-d14	19.763	244	584635	4.583	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.281	88	20131	4.823	ng	# 94
3) n-Nitrosodimethylamine	3.632	42	54387	5.275	ng	97
6) bis(2-Chloroethyl)ether	7.172	93	179914	4.534	ng	99
9) Naphthalene	10.572	128	693018	4.720	ng	100
10) Hexachlorobutadiene	10.863	225	131575	4.715	ng	# 99
12) 2-Methylnaphthalene	12.190	142	445632	4.706	ng	100
16) Acenaphthylene	14.091	152	675811	5.464	ng	100
17) Acenaphthene	14.434	154	417807	4.860	ng	96
18) Fluorene	15.423	166	503545	4.910	ng	99
20) 4,6-Dinitro-2-methylph...	15.512	198	37148	9.819	ng	# 66
21) 4-Bromophenyl-phenylether	16.312	248	166049	5.121	ng	98
22) Hexachlorobenzene	16.434	284	173499	4.705	ng	100
24) Pentachlorophenol	16.774	266	71013	8.501	ng	99
25) Phenanthrene	17.164	178	801415	4.766	ng	100
26) Anthracene	17.249	178	756741	5.416	ng	100
28) Fluoranthene	19.187	202	887492	4.890	ng	100
30) Pyrene	19.550	202	920771	4.783	ng	100
32) Benzo(a)anthracene	21.302	228	924873	5.506	ng	100
33) Chrysene	21.355	228	862282	4.633	ng	100
34) Bis(2-ethylhexyl)phtha...	21.249	149	542625	8.659	ng	99
36) Indeno(1,2,3-cd)pyrene	26.000	276	924614	5.245	ng	100
37) Benzo(b)fluoranthene	22.930	252	903434	4.938	ng	100
38) Benzo(k)fluoranthene	22.975	252	863577	4.714	ng	100
39) Benzo(a)pyrene	23.526	252	802101	5.410	ng	98
40) Dibenzo(a,h)anthracene	26.018	278	739845	5.337	ng	99
41) Benzo(g,h,i)perylene	26.726	276	823892	5.104	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072421\
Data File : BN015668.D
Acq On : 23 Jul 2021 13:50
Operator : CG/JU
Sample : SSTDIC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Quant Time: Jul 23 14:38:29 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 23 14:37:10 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.744 min Scan# 506
 Delta R.T. -0.000 min
 Lab File: BN015668.D
 Acq: 23 Jul 2021 13:50

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Tgt Ion:152 Resp: 13991




Ion	Ratio	Lower	Upper
152	100		
150	160.3	123.4	185.0
115	59.0	47.2	70.8




#2
 1,4-Dioxane
 Concen: 4.823 ng
 RT: 3.281 min Scan# 22
 Delta R.T. 0.000 min
 Lab File: BN015668.D
 Acq: 23 Jul 2021 13:50

Tgt Ion: 88 Resp: 20131

Ion	Ratio	Lower	Upper
88	100		
43	22.6	23.5	35.3#
58	77.6	64.2	96.4









#3
n-Nitrosodimethylamine
Concen: 5.275 ng
RT: 3.632 min Scan# 60
Delta R.T. -0.018 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion: 42 Resp: 54387




Ion	Ratio	Lower	Upper
42	100		
74	154.8	121.0	181.6
44	6.8	6.2	9.4




#4
2-Fluorophenol
Concen: 5.011 ng
RT: 5.347 min Scan# 246
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion: 112 Resp: 168735

Ion	Ratio	Lower	Upper
112	100		
64	56.1	44.9	67.3
63	27.8	22.6	33.8









#5
Phenol-d6
Concen: 5.194 ng
RT: 6.914 min Scan# 416
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 99 Resp: 205195
Ion Ratio Lower Upper
99 100
42 16.6 13.8 20.6
71 30.1 24.2 36.2



#6
bis(2-Chloroethyl)ether
Concen: 4.534 ng
RT: 7.172 min Scan# 444
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion: 93 Resp: 179914
Ion Ratio Lower Upper
93 100
63 74.2 59.9 89.9
95 32.8 26.3 39.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.521 min Scan# 742
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:136 Resp: 55320




Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.7	13.1
54	6.6	5.2	7.8
68	5.5	4.4	6.6




#8
Nitrobenzene-d5
Concen: 5.571 ng
RT: 8.886 min Scan# 613
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion: 82 Resp: 215931

Ion	Ratio	Lower	Upper
82	100		
128	44.1	31.2	46.8
54	45.8	40.2	60.4









#9
Naphthalene
Concen: 4.720 ng
RT: 10.572 min Scan# 746
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:128 Resp: 693018
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 12.7 10.2 15.4



#10
Hexachlorobutadiene
Concen: 4.715 ng
RT: 10.863 min Scan# 769
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion:225 Resp: 131575
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.8 50.6 76.0









#11
2-Methylnaphthalene-d10
Concen: 4.856 ng
RT: 12.114 min Scan# 956
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 400429
Ion Ratio Lower Upper
152 100
151 20.3 16.4 24.6



#12
2-Methylnaphthalene
Concen: 4.706 ng
RT: 12.190 min Scan# 973
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion:142 Resp: 445632
Ion Ratio Lower Upper
142 100
141 87.3 69.5 104.3
115 32.5 26.1 39.1









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.370 min Scan# 1191
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:164 Resp: 29612
Ion Ratio Lower Upper
164 100
162 102.3 80.6 120.8
160 46.1 37.0 55.4




#14
2,4,6-Tribromophenol
Concen: 0.000 ng
Expected RT: 15.87 min

Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion: 330
Sig Exp Ratio
330 100
332 0.0
141 35.9









#15
2-Fluorobiphenyl
Concen: 4.504 ng
RT: 13.000 min Scan# 1083
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:172 Resp: 543655
Ion Ratio Lower Upper
172 100
171 33.3 26.6 40.0
170 21.4 17.2 25.8



#16
Acenaphthylene
Concen: 5.464 ng
RT: 14.091 min Scan# 1169
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion:152 Resp: 675811
Ion Ratio Lower Upper
152 100
151 19.3 15.4 23.2
153 13.1 10.3 15.5









#17
Acenaphthene
Concen: 4.860 ng
RT: 14.434 min Scan# 1196
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:154 Resp: 417807
Ion Ratio Lower Upper
154 100
153 109.8 90.6 136.0
152 52.2 44.6 67.0



#18
Fluorene
Concen: 4.910 ng
RT: 15.423 min Scan# 1274
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion:166 Resp: 503545
Ion Ratio Lower Upper
166 100
165 96.8 78.3 117.5
167 13.6 10.8 16.2









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.115 min Scan# 1411
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:188 Resp: 55719
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 12.5 9.2 13.8



#20
4,6-Dinitro-2-methylphenol
Concen: 9.819 ng
RT: 15.512 min Scan# 1281
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion:198 Resp: 37148
Ion Ratio Lower Upper
198 100
182 5.6 17.2 25.8#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 5.121 ng
RT: 16.312 min Scan# 1345
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:248 Resp: 166049




Ion	Ratio	Lower	Upper
248	100		
250	92.7	72.9	109.3
141	72.4	59.1	88.7




#22
Hexachlorobenzene
Concen: 4.705 ng
RT: 16.434 min Scan# 1355
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion:284 Resp: 173499

Ion	Ratio	Lower	Upper
284	100		
142	34.2	27.5	41.3
249	31.1	25.0	37.6









#24
Pentachlorophenol
Concen: 8.501 ng
RT: 16.774 min Scan# 1383
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:266 Resp: 71013
Ion Ratio Lower Upper
266 100
264 63.0 51.0 76.4
268 63.7 51.8 77.8



#25
Phenanthrene
Concen: 4.766 ng
RT: 17.164 min Scan# 1415
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion:178 Resp: 801415
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.6
179 15.4 12.4 18.6









#26
 Anthracene
 Concen: 5.416 ng
 RT: 17.249 min Scan# 1422
 Delta R.T. 0.000 min
 Lab File: BN015668.D
 Acq: 23 Jul 2021 13:50

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Tgt Ion:178 Resp: 756741




Ion	Ratio	Lower	Upper
178	100		
176	18.3	14.6	21.8
179	15.5	12.2	18.2




#27
 Fluoranthene-d10
 Concen: 5.193 ng
 RT: 19.157 min Scan# 1715
 Delta R.T. 0.000 min
 Lab File: BN015668.D
 Acq: 23 Jul 2021 13:50

Tgt Ion:212 Resp: 772010

Ion	Ratio	Lower	Upper
212	100		
106	14.6	11.5	17.3
104	8.0	6.2	9.4









#28
Fluoranthene
Concen: 4.890 ng
RT: 19.187 min Scan# 1721
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:202 Resp: 887492
Ion Ratio Lower Upper
202 100
101 11.9 9.5 14.3
203 17.5 13.8 20.6



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.323 min Scan# 2010
Delta R.T. 0.011 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion:240 Resp: 53602
Ion Ratio Lower Upper
240 100
120 6.0 12.2 18.4#
236 25.6 21.5 32.3









#30
Pyrene
Concen: 4.783 ng
RT: 19.550 min Scan# 1794
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:202 Resp: 920771




Ion	Ratio	Lower	Upper
202	100		
200	20.6	16.4	24.6
203	17.6	13.9	20.9




#31
Terphenyl-d14
Concen: 4.583 ng
RT: 19.763 min Scan# 1837
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion:244 Resp: 584635

Ion	Ratio	Lower	Upper
244	100		
212	7.6	6.1	9.1
122	12.6	10.1	15.1









#32
Benzo(a)anthracene
Concen: 5.506 ng
RT: 21.302 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:228 Resp: 924873
Ion Ratio Lower Upper
228 100
226 26.6 21.4 32.0
229 19.7 15.5 23.3



#33
Chrysene
Concen: 4.633 ng
RT: 21.355 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion:228 Resp: 862282
Ion Ratio Lower Upper
228 100
226 29.5 23.5 35.3
229 19.7 15.6 23.4









#34
Bis(2-ethylhexyl)phthalate
Concen: 8.659 ng
RT: 21.249 min Scan# 2003
Delta R.T. 0.011 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:149 Resp: 542625
Ion Ratio Lower Upper
149 100
167 28.3 22.2 33.4
279 3.6 2.8 4.2



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.628 min Scan# 2469
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion:264 Resp: 48653
Ion Ratio Lower Upper
264 100
260 24.8 19.8 29.8
265 24.0 20.0 30.0









#36
Indeno(1,2,3-cd)pyrene
Concen: 5.245 ng
RT: 26.000 min Scan# 3162
Delta R.T. 0.004 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:276 Resp: 924614
Ion Ratio Lower Upper
276 100
138 27.5 21.8 32.6
277 25.0 19.9 29.9



#37
Benzo(b)fluoranthene
Concen: 4.938 ng
RT: 22.930 min Scan# 2265
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50


Tgt Ion:252 Resp: 903434
Ion Ratio Lower Upper
252 100
253 21.9 17.6 26.4
125 10.8 8.8 13.2








#38
Benzo(k)fluoranthene
Concen: 4.714 ng
RT: 22.975 min Scan# 2278
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50


Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0





Tgt Ion:252 Resp: 863577
Ion Ratio Lower Upper
252 100
253 21.7 17.5 26.3
125 10.8 8.8 13.2




#39
Benzo(a)pyrene
Concen: 5.410 ng
RT: 23.526 min Scan# 2439
Delta R.T. 0.000 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50



Tgt Ion:252 Resp: 802101
Ion Ratio Lower Upper
252 100
253 21.7 18.0 27.0
125 11.9 10.4 15.6









#40
Dibenzo(a,h)anthracene
Concen: 5.337 ng
RT: 26.018 min Scan# 3167
Delta R.T. 0.004 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:278 Resp: 739845
Ion Ratio Lower Upper
278 100
139 18.4 14.6 22.0
279 23.3 19.3 28.9



#41
Benzo(g,h,i)perylene
Concen: 5.104 ng
RT: 26.726 min Scan# 3374
Delta R.T. 0.007 min
Lab File: BN015668.D
Acq: 23 Jul 2021 13:50

Tgt Ion:276 Resp: 823892
Ion Ratio Lower Upper
276 100
277 23.5 19.1 28.7
138 23.6 19.1 28.7

