

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072421\
 Data File : BN015669.D
 Acq On : 23 Jul 2021 15:22
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN072421

Quant Time: Jul 23 15:54:47 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 23 14:45:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.744	152	12448	0.400	ng	0.00
7) Naphthalene-d8	10.521	136	54254	0.400	ng	0.00
13) Acenaphthene-d10	14.370	164	27705	0.400	ng	0.00
19) Phenanthrene-d10	17.115	188	50336	0.400	ng	0.00
29) Chrysene-d12	21.323	240	44699	0.400	ng	# 0.01
35) Perylene-d12	23.632	264	38282	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.346	112	11503	0.384	ng	0.00
5) Phenol-d6	6.914	99	13257	0.377	ng	0.00
8) Nitrobenzene-d5	8.886	82	13242	0.348	ng	0.00
11) 2-Methylnaphthalene-d10	12.114	152	31301	0.387	ng	0.00
14) 2,4,6-Tribromophenol	15.867	330	3054	0.354	ng	0.00
15) 2-Fluorobiphenyl	13.000	172	43284	0.383	ng	0.00
27) Fluoranthene-d10	19.157	212	49968	0.372	ng	0.00
31) Terphenyl-d14	19.763	244	40804	0.384	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.281	88	1463	0.394	ng	97
3) n-Nitrosodimethylamine	3.650	42	3466	0.378	ng	100
6) bis(2-Chloroethyl)ether	7.172	93	14056	0.398	ng	100
9) Naphthalene	10.571	128	56574	0.393	ng	100
10) Hexachlorobutadiene	10.863	225	10788	0.394	ng	# 99
12) 2-Methylnaphthalene	12.189	142	36388	0.392	ng	99
16) Acenaphthylene	14.091	152	41563	0.359	ng	100
17) Acenaphthene	14.434	154	30518	0.379	ng	100
18) Fluorene	15.423	166	36836	0.384	ng	100
20) 4,6-Dinitro-2-methylph...	15.512	198	649	0.345	ng	100
21) 4-Bromophenyl-phenylether	16.312	248	11053	0.377	ng	98
22) Hexachlorobenzene	16.433	284	13251	0.398	ng	99
23) Atrazine	16.592	200	6403	0.367	ng	98
24) Pentachlorophenol	16.774	266	1801	0.352	ng	98
25) Phenanthrene	17.163	178	59454	0.391	ng	100
26) Anthracene	17.249	178	46184	0.366	ng	100
28) Fluoranthene	19.187	202	63487	0.387	ng	100
30) Pyrene	19.554	202	61955	0.386	ng	100
32) Benzo(a)anthracene	21.302	228	50620	0.361	ng	99
33) Chrysene	21.355	228	60919	0.393	ng	99
34) Bis(2-ethylhexyl)phtha...	21.249	149	15344	0.407	ng	100
36) Indeno(1,2,3-cd)pyrene	26.000	276	51650	0.372	ng	100
37) Benzo(b)fluoranthene	22.933	252	53735	0.373	ng	100
38) Benzo(k)fluoranthene	22.978	252	55675	0.386	ng	100
39) Benzo(a)pyrene	23.532	252	41977	0.360	ng	100
40) Dibenzo(a,h)anthracene	26.021	278	40568	0.372	ng	99
41) Benzo(g,h,i)perylene	26.726	276	47030	0.370	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072421\
Data File : BN015669.D
Acq On : 23 Jul 2021 15:22
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN072421

Quant Time: Jul 23 15:54:47 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 23 14:45:01 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.744 min Scan# 506
 Delta R.T. -0.000 min
 Lab File: BN015669.D
 Acq: 23 Jul 2021 15:22

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN072421

Tgt Ion: 152 Resp: 12448




Ion	Ratio	Lower	Upper
152	100		
150	155.2	123.4	185.0
115	59.7	47.2	70.8




#2
 1,4-Dioxane
 Concen: 0.394 ng
 RT: 3.281 min Scan# 22
 Delta R.T. -0.000 min
 Lab File: BN015669.D
 Acq: 23 Jul 2021 15:22

Tgt Ion: 88 Resp: 1463

Ion	Ratio	Lower	Upper
88	100		
43	30.1	23.5	35.3
58	83.5	64.2	96.4









#3
n-Nitrosodimethylamine
Concen: 0.378 ng
RT: 3.650 min Scan# 62
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22




Instrument :
BNA_N
ClientSampleId :
ICVBN072421


Tgt Ion: 42 Resp: 3466
Ion Ratio Lower Upper
42 100
74 151.8 121.0 181.6
44 7.5 6.2 9.4



#4
2-Fluorophenol
Concen: 0.384 ng
RT: 5.346 min Scan# 246
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22

Tgt Ion: 112 Resp: 11503
Ion Ratio Lower Upper
112 100
64 56.1 44.9 67.3
63 28.1 22.6 33.8









#5
Phenol-d6
Concen: 0.377 ng
RT: 6.914 min Scan# 416
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22

Instrument :
BNA_N
ClientSampleId :
ICVBN072421

Tgt Ion: 99 Resp: 13257




Ion	Ratio	Lower	Upper
99	100		
42	17.0	13.8	20.6
71	30.4	24.2	36.2




#6
bis(2-Chloroethyl)ether
Concen: 0.398 ng
RT: 7.172 min Scan# 444
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22

Tgt Ion: 93 Resp: 14056

Ion	Ratio	Lower	Upper
93	100		
63	75.4	59.9	89.9
95	33.1	26.3	39.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.521 min Scan# 742
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22




Instrument :
BNA_N
ClientSampleId :
ICVBN072421


Tgt Ion:	136	Resp:	54254
Ion Ratio	Lower	Upper	
136	100		
137	11.1	8.7	13.1
54	6.6	5.2	7.8
68	5.7	4.4	6.6



#8
Nitrobenzene-d5
Concen: 0.348 ng
RT: 8.886 min Scan# 613
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22

Tgt Ion:	82	Resp:	13242
Ion Ratio	Lower	Upper	
82	100		
128	39.7	31.2	46.8
54	48.9	40.2	60.4









#9
Naphthalene
Concen: 0.393 ng
RT: 10.571 min Scan# 746
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22




Instrument :
BNA_N
ClientSampleId :
ICVBN072421


Tgt Ion:128 Resp: 56574
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 12.8 10.2 15.4



#10
Hexachlorobutadiene
Concen: 0.394 ng
RT: 10.863 min Scan# 769
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22

Tgt Ion:225 Resp: 10788
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.9 50.6 76.0









#11
2-Methylnaphthalene-d10
Concen: 0.387 ng
RT: 12.114 min Scan# 956
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22




Instrument :
BNA_N
ClientSampleId :
ICVBN072421


Tgt Ion:152 Resp: 31301
Ion Ratio Lower Upper
152 100
151 20.5 16.4 24.6



#12
2-Methylnaphthalene
Concen: 0.392 ng
RT: 12.189 min Scan# 973
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22

Tgt Ion:142 Resp: 36388
Ion Ratio Lower Upper
142 100
141 88.0 69.5 104.3
115 32.4 26.1 39.1









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.370 min Scan# 1191
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22




Instrument :
BNA_N
ClientSampleId :
ICVBN072421


Tgt Ion:164 Resp: 27705
Ion Ratio Lower Upper
164 100
162 101.2 80.6 120.8
160 45.6 37.0 55.4



#14
2,4,6-Tribromophenol
Concen: 0.354 ng
RT: 15.867 min Scan# 1309
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22

Tgt Ion:330 Resp: 3054
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 37.0 28.7 43.1









#15
2-Fluorobiphenyl
Concen: 0.383 ng
RT: 13.000 min Scan# 1083
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22




Instrument :
BNA_N
ClientSampleId :
ICVBN072421


Tgt Ion:	172	Resp:	43284
Ion Ratio	Lower	Upper	
172	100		
171	33.3	26.6	40.0
170	21.7	17.2	25.8



#16
Acenaphthylene
Concen: 0.359 ng
RT: 14.091 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22

Tgt Ion:	152	Resp:	41563
Ion Ratio	Lower	Upper	
152	100		
151	19.4	15.4	23.2
153	13.0	10.3	15.5









#17
Acenaphthene
Concen: 0.379 ng
RT: 14.434 min Scan# 1196
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22




Instrument :
BNA_N
ClientSampleId :
ICVBN072421


Tgt Ion:154 Resp: 30518
Ion Ratio Lower Upper
154 100
153 113.2 90.6 136.0
152 55.9 44.6 67.0



#18
Fluorene
Concen: 0.384 ng
RT: 15.423 min Scan# 1274
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22

Tgt Ion:166 Resp: 36836
Ion Ratio Lower Upper
166 100
165 97.4 78.3 117.5
167 13.4 10.8 16.2









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.115 min Scan# 1411
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22




Instrument :
BNA_N
ClientSampleId :
ICVBN072421


Tgt Ion:188	Resp:	50336
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	12.4	9.2 13.8



#20
4,6-Dinitro-2-methylphenol
Concen: 0.345 ng
RT: 15.512 min Scan# 1281
Delta R.T. 0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22

Tgt Ion:198	Resp:	649
Ion Ratio	Lower	Upper
198	100	
182	21.6	17.2 25.8
77	0.0	0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.377 ng
RT: 16.312 min Scan# 1345
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22




Instrument :
BNA_N
ClientSampleId :
ICVBN072421


Tgt Ion:	248	Resp:	11053
Ion	Ratio	Lower	Upper
248	100		
250	92.0	72.9	109.3
141	76.1	59.1	88.7



#22
Hexachlorobenzene
Concen: 0.398 ng
RT: 16.433 min Scan# 1355
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22

Tgt Ion:	284	Resp:	13251
Ion	Ratio	Lower	Upper
284	100		
142	34.1	27.5	41.3
249	30.9	25.0	37.6









#23
Atrazine
Concen: 0.367 ng
RT: 16.592 min Scan# 1368
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22




Instrument :
BNA_N
ClientSampleId :
ICVBN072421


Tgt Ion:	200	Resp:	6403
Ion	Ratio	Lower	Upper
200	100		
173	23.1	17.8	26.8
215	50.9	39.8	59.8



#24
Pentachlorophenol
Concen: 0.352 ng
RT: 16.774 min Scan# 1383
Delta R.T. 0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22

Tgt Ion:	266	Resp:	1801
Ion	Ratio	Lower	Upper
266	100		
264	62.7	51.0	76.4
268	63.2	51.8	77.8









#25
Phenanthrene
Concen: 0.391 ng
RT: 17.163 min Scan# 1415
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22




Instrument :
BNA_N
ClientSampleId :
ICVBN072421


Tgt Ion:178 Resp: 59454
Ion Ratio Lower Upper
178 100
176 18.7 15.0 22.6
179 15.3 12.4 18.6



#26
Anthracene
Concen: 0.366 ng
RT: 17.249 min Scan# 1422
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22

Tgt Ion:178 Resp: 46184
Ion Ratio Lower Upper
178 100
176 18.2 14.6 21.8
179 15.2 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.372 ng
RT: 19.157 min Scan# 1715
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22




Instrument :
BNA_N
ClientSampleId :
ICVBN072421


Tgt Ion:	212	Resp:	49968
Ion Ratio	Lower	Upper	
212	100		
106	14.5	11.5	17.3
104	7.9	6.2	9.4



#28
Fluoranthene
Concen: 0.387 ng
RT: 19.187 min Scan# 1721
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22

Tgt Ion:	202	Resp:	63487
Ion Ratio	Lower	Upper	
202	100		
101	12.0	9.5	14.3
203	17.2	13.8	20.6









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.323 min Scan# 2010
Delta R.T. 0.011 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22




Instrument :
BNA_N
ClientSampleId :
ICVBN072421


Tgt Ion:240 Resp: 44699
Ion Ratio Lower Upper
240 100
120 7.5 12.2 18.4#
236 25.8 21.5 32.3



#30
Pyrene
Concen: 0.386 ng
RT: 19.554 min Scan# 1795
Delta R.T. 0.005 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22

Tgt Ion:202 Resp: 61955
Ion Ratio Lower Upper
202 100
200 20.5 16.4 24.6
203 17.4 13.9 20.9









#31
Terphenyl-d14
Concen: 0.384 ng
RT: 19.763 min Scan# 1837
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22




Instrument :
BNA_N
ClientSampleId :
ICVBN072421


Tgt Ion:244 Resp: 40804
Ion Ratio Lower Upper
244 100
212 7.8 6.1 9.1
122 13.7 10.1 15.1



#32
Benzo(a)anthracene
Concen: 0.361 ng
RT: 21.302 min Scan# 2008
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22

Tgt Ion:228 Resp: 50620
Ion Ratio Lower Upper
228 100
226 27.5 21.4 32.0
229 19.1 15.5 23.3









#33
Chrysene
Concen: 0.393 ng
RT: 21.355 min Scan# 2013
Delta R.T. -0.000 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22




Instrument :
BNA_N
ClientSampleId :
ICVBN072421


Tgt Ion:	228	Resp:	60919
Ion Ratio	Lower	Upper	
228	100		
226	30.1	23.5	35.3
229	19.4	15.6	23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.407 ng
RT: 21.249 min Scan# 2003
Delta R.T. 0.011 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22


Tgt Ion:	149	Resp:	15344
Ion Ratio	Lower	Upper	
149	100		
167	27.8	22.2	33.4
279	3.5	2.8	4.2








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.632 min Scan# 2470
Delta R.T. 0.003 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22

Instrument :
BNA_N
ClientSampleId :
ICVBN072421







Tgt Ion:264 Resp: 38282
Ion Ratio Lower Upper
264 100
260 25.1 19.8 29.8
265 24.6 20.0 30.0



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.372 ng
RT: 26.000 min Scan# 3162
Delta R.T. 0.003 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22

Tgt Ion:276 Resp: 51650
Ion Ratio Lower Upper
276 100
138 27.4 21.8 32.6
277 24.9 19.9 29.9









#37
Benzo(b)fluoranthene
Concen: 0.373 ng
RT: 22.933 min Scan# 2266
Delta R.T. 0.003 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22




Instrument :
BNA_N
ClientSampleId :
ICVBN072421


Tgt Ion:252 Resp: 53735
Ion Ratio Lower Upper
252 100
253 22.1 17.6 26.4
125 11.2 8.8 13.2



#38
Benzo(k)fluoranthene
Concen: 0.386 ng
RT: 22.978 min Scan# 2279
Delta R.T. 0.003 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22


Tgt Ion:252 Resp: 55675
Ion Ratio Lower Upper
252 100
253 22.1 17.5 26.3
125 11.1 8.8 13.2








#39
Benzo(a)pyrene
Concen: 0.360 ng
RT: 23.532 min Scan# 2441
Delta R.T. 0.007 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22

Instrument :
BNA_N
ClientSampleId :
ICVBN072421







Tgt Ion:252 Resp: 41977
Ion Ratio Lower Upper
252 100
253 22.6 18.0 27.0
125 12.8 10.4 15.6



#40
Dibenzo(a,h)anthracene
Concen: 0.372 ng
RT: 26.021 min Scan# 3168
Delta R.T. 0.007 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22

Tgt Ion:278 Resp: 40568
Ion Ratio Lower Upper
278 100
139 18.3 14.6 22.0
279 23.6 19.3 28.9





#41
Benzo(g,h,i)perylene
Concen: 0.370 ng
RT: 26.726 min Scan# 3374
Delta R.T. 0.007 min
Lab File: BN015669.D
Acq: 23 Jul 2021 15:22

Instrument :
BNA_N
ClientSampleId :
ICVBN072421

Tgt Ion:276 Resp: 47030
Ion Ratio Lower Upper
276 100
277 24.1 19.1 28.7
138 23.8 19.1 28.7

