

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072521\
 Data File : BN015697.D
 Acq On : 24 Jul 2021 21:15
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN072521


Quant Time: Jul 26 02:26:28 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 26 02:17:00 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.688	152	14548	0.40	ng	0.00
7) Naphthalene-d8	10.470	136	65724	0.40	ng	0.00
13) Acenaphthene-d10	14.319	164	36311	0.40	ng	0.00
19) Phenanthrene-d10	17.078	188	71280	0.40	ng	0.00
29) Chrysene-d12	21.270	240	69675	0.40	ng	#-0.01
35) Perylene-d12	23.560	264	68255	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.291	112	15074	0.39	ng	0.00
5) Phenol-d6	6.868	99	18119	0.39	ng	0.00
8) Nitrobenzene-d5	8.835	82	17869	0.36	ng	0.00
11) 2-Methylnaphthalene-d10	12.065	152	39444	0.39	ng	0.00
14) 2,4,6-Tribromophenol	15.817	330	5442	0.35	ng	0.00
15) 2-Fluorobiphenyl	12.949	172	55222	0.38	ng	0.00
27) Fluoranthene-d10	19.113	212	74359	0.38	ng	0.00
31) Terphenyl-d14	19.718	244	63212	0.39	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.226	88	1968	0.45	ng	# 95
3) n-Nitrosodimethylamine	3.585	42	4223	0.38	ng	97
6) bis(2-Chloroethyl)ether	7.126	93	16719	0.40	ng	100
9) Naphthalene	10.521	128	68671	0.39	ng	99
10) Hexachlorobutadiene	10.812	225	12805	0.39	ng	# 99
12) 2-Methylnaphthalene	12.136	142	45886	0.40	ng	98
16) Acenaphthylene	14.040	152	59823	0.38	ng	100
17) Acenaphthene	14.383	154	40932	0.39	ng	99
18) Fluorene	15.373	166	50511	0.39	ng	100
20) 4,6-Dinitro-2-methylph...	15.461	198	987	0.40	ng	93
21) 4-Bromophenyl-phenylether	16.275	248	15590	0.38	ng	98
22) Hexachlorobenzene	16.385	284	17681	0.39	ng	100
23) Atrazine	16.543	200	11025	0.35	ng	100
24) Pentachlorophenol	16.725	266	4607	0.29	ng	99
25) Phenanthrene	17.115	178	82752	0.39	ng	100
26) Anthracene	17.200	178	70572	0.38	ng	100
28) Fluoranthene	19.142	202	92750	0.40	ng	100
30) Pyrene	19.505	202	96412	0.40	ng	100
32) Benzo(a)anthracene	21.259	228	89378	0.38	ng	99
33) Chrysene	21.312	228	94421	0.40	ng	100
34) Bis(2-ethylhexyl)phtha...	21.206	149	39487	0.43	ng	100
36) Indeno(1,2,3-cd)pyrene	25.894	276	102338	0.39	ng	100
37) Benzo(b)fluoranthene	22.872	252	92722	0.39	ng	100
38) Benzo(k)fluoranthene	22.916	252	98745	0.41	ng	99
39) Benzo(a)pyrene	23.460	252	81930	0.38	ng	99
40) Dibenzo(a,h)anthracene	25.915	278	83850	0.39	ng	100
41) Benzo(g,h,i)perylene	26.610	276	83988	0.39	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 26 02:26:28 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 26 02:17:00 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.688 min Scan# 500
 Delta R.T. 0.000 min
 Lab File: BN015697.D
 Acq: 24 Jul 2021 21:15




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN072521


Tgt Ion:152 Resp: 14548
 Ion Ratio Lower Upper
 152 100
 150 156.9 125.2 187.8
 115 60.2 48.6 73.0



#2
 1,4-Dioxane
 Concen: 0.45 ng
 RT: 3.226 min Scan# 16
 Delta R.T. 0.000 min
 Lab File: BN015697.D
 Acq: 24 Jul 2021 21:15

Tgt Ion: 88 Resp: 1968
 Ion Ratio Lower Upper
 88 100
 43 24.2 0.0 0.0#
 58 63.4 53.8 80.6









#3
n-Nitrosodimethylamine
Concen: 0.38 ng
RT: 3.585 min Scan# 55
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15




Instrument :
BNA_N
ClientSampleId :
ICVBN072521


Tgt Ion: 42 Resp: 4223
Ion Ratio Lower Upper
42 100
74 154.8 126.7 190.1
44 7.9 5.8 8.8



#4
2-Fluorophenol
Concen: 0.39 ng
RT: 5.291 min Scan# 240
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15

Tgt Ion: 112 Resp: 15074
Ion Ratio Lower Upper
112 100
64 55.9 44.8 67.2
63 28.0 22.2 33.4









#5
Phenol-d6
Concen: 0.39 ng
RT: 6.868 min Scan# 411
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15




Instrument :
BNA_N
ClientSampleId :
ICVBN072521


Tgt Ion: 99 Resp: 18119
Ion Ratio Lower Upper
99 100
42 16.6 13.4 20.0
71 30.4 24.5 36.7



#6
bis(2-Chloroethyl)ether
Concen: 0.40 ng
RT: 7.126 min Scan# 439
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15

Tgt Ion: 93 Resp: 16719
Ion Ratio Lower Upper
93 100
63 74.4 59.8 89.8
95 32.8 26.4 39.6









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.470 min Scan# 738
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15




Instrument :
BNA_N
ClientSampleId :
ICVBN072521


Tgt Ion:	136	Resp:	65724
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.9	13.3
54	6.2	5.0	7.4
68	5.2	4.3	6.5



#8
Nitrobenzene-d5
Concen: 0.36 ng
RT: 8.835 min Scan# 609
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15

Tgt Ion:	82	Resp:	17869
Ion	Ratio	Lower	Upper
82	100		
128	41.3	32.0	48.0
54	47.8	38.9	58.3









#9
Naphthalene
Concen: 0.39 ng
RT: 10.521 min Scan# 742
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15




Instrument :
BNA_N
ClientSampleId :
ICVBN072521


Tgt Ion:128 Resp: 68671
Ion Ratio Lower Upper
128 100
129 10.9 8.9 13.3
127 12.8 10.4 15.6



#10
Hexachlorobutadiene
Concen: 0.39 ng
RT: 10.812 min Scan# 765
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15

Tgt Ion:225 Resp: 12805
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.3 51.1 76.7









#11
2-Methylnaphthalene-d10
Concen: 0.39 ng
RT: 12.065 min Scan# 945
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15




Instrument :
BNA_N
ClientSampleId :
ICVBN072521


Tgt Ion:152 Resp: 39444
Ion Ratio Lower Upper
152 100
151 20.5 16.3 24.5



#12
2-Methylnaphthalene
Concen: 0.40 ng
RT: 12.136 min Scan# 961
Delta R.T. -0.004 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15

Tgt Ion:142 Resp: 45886
Ion Ratio Lower Upper
142 100
141 88.9 69.4 104.2
115 33.3 25.9 38.9









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.319 min Scan# 1187
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15




Instrument :
BNA_N
ClientSampleId :
ICVBN072521


Tgt Ion:164 Resp: 36311
Ion Ratio Lower Upper
164 100
162 103.0 83.6 125.4
160 47.5 38.6 58.0



#14
2,4,6-Tribromophenol
Concen: 0.35 ng
RT: 15.817 min Scan# 1305
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15

Tgt Ion:330 Resp: 5442
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 35.7 28.5 42.7









#15
2-Fluorobiphenyl
Concen: 0.38 ng
RT: 12.949 min Scan# 1079
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15




Instrument :
BNA_N
ClientSampleId :
ICVBN072521


Tgt Ion:	172	Resp:	55222
Ion Ratio	Lower	Upper	
172	100		
171	33.9	27.4	41.0
170	22.4	17.9	26.9



#16
Acenaphthylene
Concen: 0.38 ng
RT: 14.040 min Scan# 1165
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15

Tgt Ion:	152	Resp:	59823
Ion Ratio	Lower	Upper	
152	100		
151	19.2	15.4	23.2
153	13.0	10.4	15.6









#17
Acenaphthene
Concen: 0.39 ng
RT: 14.383 min Scan# 1192
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15




Instrument :
BNA_N
ClientSampleId :
ICVBN072521


Tgt Ion:154 Resp: 40932
Ion Ratio Lower Upper
154 100
153 112.2 89.0 133.6
152 54.4 43.3 64.9



#18
Fluorene
Concen: 0.39 ng
RT: 15.373 min Scan# 1270
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15

Tgt Ion:166 Resp: 50511
Ion Ratio Lower Upper
166 100
165 97.8 77.9 116.9
167 13.4 10.8 16.2









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.078 min Scan# 1408
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15




Instrument :
BNA_N
ClientSampleId :
ICVBN072521


Tgt Ion:188 Resp: 71280
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.1 5.3 7.9



#20
4,6-Dinitro-2-methylphenol
Concen: 0.40 ng
RT: 15.461 min Scan# 1277
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15

Tgt Ion:198 Resp: 987
Ion Ratio Lower Upper
198 100
182 17.1 16.4 24.6
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.38 ng
RT: 16.275 min Scan# 1342
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15




Instrument :
BNA_N
ClientSampleId :
ICVBN072521


Tgt Ion:	248	Resp:	15590
Ion	Ratio	Lower	Upper
248	100		
250	104.1	81.6	122.4
141	35.2	29.4	44.0



#22
Hexachlorobenzene
Concen: 0.39 ng
RT: 16.385 min Scan# 1351
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15

Tgt Ion:	284	Resp:	17681
Ion	Ratio	Lower	Upper
284	100		
142	34.3	27.6	41.4
249	31.1	24.6	37.0









#23
Atrazine
Concen: 0.35 ng
RT: 16.543 min Scan# 1364
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15




Instrument :
BNA_N
ClientSampleId :
ICVBN072521


Tgt Ion:200 Resp: 11025
Ion Ratio Lower Upper
200 100
173 26.8 21.8 32.6
215 47.7 38.2 57.2



#24
Pentachlorophenol
Concen: 0.29 ng
RT: 16.725 min Scan# 1379
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15

Tgt Ion:266 Resp: 4607
Ion Ratio Lower Upper
266 100
264 62.9 49.9 74.9
268 63.8 50.8 76.2









#25
Phenanthrene
Concen: 0.39 ng
RT: 17.115 min Scan# 1411
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15




Instrument :
BNA_N
ClientSampleId :
ICVBN072521


Tgt Ion:178 Resp: 82752
Ion Ratio Lower Upper
178 100
176 18.8 15.1 22.7
179 15.4 12.4 18.6



#26
Anthracene
Concen: 0.38 ng
RT: 17.200 min Scan# 1418
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15

Tgt Ion:178 Resp: 70572
Ion Ratio Lower Upper
178 100
176 18.2 14.6 21.8
179 15.2 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.38 ng
RT: 19.113 min Scan# 1706
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15




Instrument :
BNA_N
ClientSampleId :
ICVBN072521


Tgt Ion:212 Resp: 74359
Ion Ratio Lower Upper
212 100
106 14.4 11.5 17.3
104 7.8 6.3 9.5



#28
Fluoranthene
Concen: 0.40 ng
RT: 19.142 min Scan# 1712
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15

Tgt Ion:202 Resp: 92750
Ion Ratio Lower Upper
202 100
101 11.8 9.4 14.2
203 17.3 13.8 20.8









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.270 min Scan# 2005
Delta R.T. -0.011 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15




Instrument :
BNA_N
ClientSampleId :
ICVBN072521


Tgt Ion:240 Resp: 69675
Ion Ratio Lower Upper
240 100
120 15.4 4.9 7.3#
236 27.2 20.4 30.6



#30
Pyrene
Concen: 0.40 ng
RT: 19.505 min Scan# 1785
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15

Tgt Ion:202 Resp: 96412
Ion Ratio Lower Upper
202 100
200 20.6 16.4 24.6
203 17.4 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.39 ng
 RT: 19.718 min Scan# 1828
 Delta R.T. 0.000 min
 Lab File: BN015697.D
 Acq: 24 Jul 2021 21:15




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN072521


Tgt Ion:244 Resp: 63212
 Ion Ratio Lower Upper
 244 100
 212 7.7 6.2 9.2
 122 13.3 10.9 16.3



#32
 Benzo(a)anthracene
 Concen: 0.38 ng
 RT: 21.259 min Scan# 2004
 Delta R.T. 0.000 min
 Lab File: BN015697.D
 Acq: 24 Jul 2021 21:15

Tgt Ion:228 Resp: 89378
 Ion Ratio Lower Upper
 228 100
 226 26.3 21.2 31.8
 229 19.8 15.5 23.3





#33

Chrysene

Concen: 0.40 ng

RT: 21.312 min Scan# 2009

Delta R.T. 0.000 min

Lab File: BN015697.D

Acq: 24 Jul 2021 21:15





Tgt Ion:228 Resp: 94421

Ion Ratio Lower Upper

228 100

226 29.0 23.4 35.0

229 19.7 15.5 23.3



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.43 ng

RT: 21.206 min Scan# 1999

Delta R.T. 0.000 min

Lab File: BN015697.D

Acq: 24 Jul 2021 21:15




Tgt Ion:149 Resp: 39487


Ion Ratio Lower Upper

149 100

167 28.5 22.6 34.0


279 3.7 2.8 4.2








#35
 Perylene-d12
 Concen: 0.40 ng
 RT: 23.560 min Scan# 2449
 Delta R.T. -0.003 min
 Lab File: BN015697.D
 Acq: 24 Jul 2021 21:15

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN072521







Tgt Ion:264	Resp:	68255
Ion Ratio	Lower	Upper
264	100	
260	25.1	19.8 29.8
265	23.3	19.0 28.6



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.39 ng
 RT: 25.894 min Scan# 3131
 Delta R.T. -0.003 min
 Lab File: BN015697.D
 Acq: 24 Jul 2021 21:15

Tgt Ion:276	Resp:	102338
Ion Ratio	Lower	Upper
276	100	
138	29.1	23.0 34.4
277	25.0	20.0 30.0









#37
Benzo(b)fluoranthene
Concen: 0.39 ng
RT: 22.872 min Scan# 2248
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15




Instrument :
BNA_N
ClientSampleId :
ICVBN072521


Tgt Ion:252 Resp: 92722
Ion Ratio Lower Upper
252 100
253 22.2 17.5 26.3
125 11.1 8.9 13.3



#38
Benzo(k)fluoranthene
Concen: 0.41 ng
RT: 22.916 min Scan# 2261
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15

Tgt Ion:252 Resp: 98745
Ion Ratio Lower Upper
252 100
253 22.2 17.6 26.4
125 11.5 9.4 14.2









#39
Benzo(a)pyrene
Concen: 0.38 ng
RT: 23.460 min Scan# 2420
Delta R.T. 0.000 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15




Instrument :
BNA_N
ClientSampleId :
ICVBN072521


Tgt Ion:252	Resp:	81930
Ion Ratio	Lower	Upper
252	100	
253	22.3	17.7 26.5
125	13.1	10.2 15.2



#40
Dibenzo(a,h)anthracene
Concen: 0.39 ng
RT: 25.915 min Scan# 3137
Delta R.T. -0.003 min
Lab File: BN015697.D
Acq: 24 Jul 2021 21:15

Tgt Ion:278	Resp:	83850
Ion Ratio	Lower	Upper
278	100	
139	19.4	15.1 22.7
279	23.7	19.0 28.4





#41
 Benzo(g,h,i)perylene
 Concen: 0.39 ng
 RT: 26.610 min Scan# 3340
 Delta R.T. -0.003 min
 Lab File: BN015697.D
 Acq: 24 Jul 2021 21:15

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN072521

Tgt Ion:	276	Resp:	83988
Ion Ratio	Lower	Upper	
276	100		
277	24.0	19.4	29.0
138	23.9	19.0	28.6

