

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072522\
 Data File : BN020832.D
 Acq On : 25 Jul 2022 19:20
 Operator : CG/JU
 Sample : PB146369BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK369

Quant Time: Jul 26 00:31:19 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 25 18:06:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.817	152	2510	0.400	ng/ul	0.00
4) Naphthalene-d8	10.612	136	8887	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.469	164	5632	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.217	188	11775	0.400	ng/ul	0.00
17) Chrysene-d12	21.433	240	11503	0.400	ng/ul	0.00
23) Perylene-d12	23.772	264	13147	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.214	96	16755	5.555	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.223	152	4499	0.333	ng/ul	0.00
18) Fluoranthene-d10	19.257	212	9538	0.260	ng/ul	0.00
Target Compounds						
					Qvalue	
12) Fluorene	15.464	166	1391	0.056	ng/ul#	14
20) Pyrene	19.620	202	1580	0.028	ng/ul#	1
26) Benzo(a)pyrene	23.617	252	3907	0.068	ng/ul#	73


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072522\
Data File : BN020832.D
Acq On : 25 Jul 2022 19:20
Operator : CG/JU
Sample : PB146369BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK369

Quant Time: Jul 26 00:31:19 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN072522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 25 18:06:54 2022
Response via : Initial Calibration









#12
Fluorene
Concen: 0.056 ng/ul
RT: 15.464 min Scan# 21
Delta R.T. -0.060 min
Lab File: BN020832.D
Acq: 25 Jul 2022 19:20




Instrument :
BNA_N
ClientSampleId :
SBLK369


Tgt Ion:166 Resp: 1391
Ion Ratio Lower Upper
166 100
165 8.8 78.2 117.2#
167 36.2 11.3 16.9#



#20
Pyrene
Concen: 0.028 ng/ul
RT: 19.620 min Scan# 3656
Delta R.T. -0.033 min
Lab File: BN020832.D
Acq: 25 Jul 2022 19:20

Tgt Ion:202 Resp: 1580
Ion Ratio Lower Upper
202 100
101 161.7 11.6 17.4#
100 1161.6 9.2 13.8#





#26

Benzo(a)pyrene

Concen: 0.068 ng/ul

RT: 23.617 min Scan# 4911

Delta R.T. -0.059 min

Lab File: BN020832.D


Acq: 25 Jul 2022 19:20

Instrument :

BNA_N

ClientSampleId :

SBLK369



Tgt Ion:252 Resp: 3907

Ion Ratio Lower Upper

252 100

253 42.1 21.0 31.4#

125 34.1 18.8 28.2#

