

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072522\
 Data File : BN020834.D
 Acq On : 25 Jul 2022 21:13
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4252

Quant Time: Jul 26 00:31:44 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 25 18:06:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.821	152	2171	0.400	ng/ul	0.00
4) Naphthalene-d8	10.617	136	7659	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.470	164	4529	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.221	188	9343	0.400	ng/ul	0.00
17) Chrysene-d12	21.426	240	8122	0.400	ng/ul	0.00
23) Perylene-d12	23.770	264	6950	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.222	96	896	0.343	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.218	152	4338	0.373	ng/ul	0.00
18) Fluoranthene-d10	19.258	212	9770	0.377	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.256	88	879	0.328	ng/ul	95
5) Naphthalene	10.667	128	9353	0.390	ng/ul	99
7) 2-Methylnaphthalene	12.295	142	5854	0.386	ng/ul	100
8) 1-Methylnaphthalene	12.509	142	5308	0.361	ng/ul	99
10) Acenaphthylene	14.187	152	8788	0.390	ng/ul	99
11) Acenaphthene	14.530	153	6980	0.383	ng/ul	99
12) Fluorene	15.524	166	7770	0.389	ng/ul	99
14) Pentachlorophenol	16.888	266	583	0.414	ng/ul	100
15) Phenanthrene	17.264	178	12942	0.378	ng/ul	100
16) Anthracene	17.361	178	10540	0.384	ng/ul	99
19) Fluoranthene	19.285	202	14640	0.385	ng/ul	99
20) Pyrene	19.653	202	15188	0.383	ng/ul	99
21) Benzo(a)anthracene	21.411	228	10365	0.386	ng/ul	99
22) Chrysene	21.461	228	13091	0.379	ng/ul	99
24) Benzo(b)fluoranthene	23.060	252	12049	0.366	ng/ul	99
25) Benzo(k)fluoranthene	23.107	252	12803	0.376	ng/ul	98
26) Benzo(a)pyrene	23.668	252	11319	0.373	ng/ul	99
27) Indeno(1,2,3-cd)pyrene	26.205	276	11975	0.375	ng/ul#	99
28) Dibenzo(a,h)anthracene	26.225	278	9207	0.378	ng/ul	98
29) Benzo(g,h,i)perylene	26.943	276	10625	0.369	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072522\
Data File : BN020834.D
Acq On : 25 Jul 2022 21:13
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4252

Quant Time: Jul 26 00:31:44 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN072522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 25 18:06:54 2022
Response via : Initial Calibration

