



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072522\
 Data File : BN020824.D
 Acq On : 25 Jul 2022 11:13
 Operator : CG/JU
 Sample : DFTPP268
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP268

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN072522.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jul 25 18:06:54 2022



AutoFind: Scans 2480, 2481, 2482; Background Corrected with Scan 2472

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.7	223296	PASS
68	69	0.00	2	1.9	3963	PASS
69	198	0.00	100	42.4	212160	PASS
70	69	0.00	2	0.5	1137	PASS
127	198	10	80	52.8	263957	PASS
197	198	0.00	2	0.2	900	PASS
198	198	100	100	100.0	499968	PASS
199	198	5	9	6.8	33811	PASS
275	198	10	60	25.6	128035	PASS
365	198	1	100	3.5	17278	PASS
441	198	0.01	100	16.1	80301	PASS
442	198	50	100	99.6	497963	PASS
443	442	15	24	20.8	103787	PASS

m/z	Abundance
40.00	809.0
40.90	219.0
42.90	264.0
43.90	2252.0
46.90	163.0
49.00	832.0
50.80	235.0
72.90	407.0
83.90	446.0
85.90	210.0
207.00	564.0

Instrument :
BNA_N
ClientSampleId :
DFTPP268

m/z	Abundance
35.90	201.0
36.90	980.0
38.00	3144.0
39.00	18856.0
40.00	1289.0
41.00	696.0
43.00	203.0
44.00	1522.0
44.90	484.0
46.90	186.0
48.00	207.0
49.10	2419.0
50.00	67800.0
51.00	257088.0
52.00	14055.0
53.00	551.0
55.00	1603.0
56.00	7321.0
57.00	17592.0
58.00	708.0
60.00	230.0
61.00	3274.0
62.00	3776.0
63.00	11376.0
64.00	1457.0
65.00	4660.0
66.00	306.0
67.10	409.0
68.00	4442.0
69.00	243072.0
70.00	1270.0
70.90	170.0
73.10	2199.0
74.00	24680.0
75.00	40272.0
76.00	14358.0
77.00	273920.0
78.00	18968.0
79.00	18952.0
80.00	14277.0
81.00	20168.0
82.00	4828.0
83.00	4082.0
84.00	769.0
85.00	3552.0
85.90	5990.0
87.00	2875.0
88.00	1012.0
88.90	595.0
91.00	4602.0
91.90	5038.0
93.00	32224.0
94.00	2016.0
95.10	521.0
96.00	1416.0
98.00	26040.0
99.00	19736.0
99.90	1677.0
101.00	11663.0
101.90	582.0
102.90	3792.0
104.00	6997.0
104.90	6778.0
106.10	2558.0
107.00	89792.0
108.00	13461.0
109.10	2656.0
110.00	155584.0
111.00	23984.0
111.90	2753.0
113.00	1029.0
114.10	205.0
114.80	468.0
115.90	5122.0
117.00	71184.0
118.00	4737.0
118.90	573.0
119.90	1149.0
121.00	398.0
122.00	5908.0
123.00	8832.0
124.00	4366.0
125.00	3229.0
127.00	287872.0
128.00	22744.0
129.00	116240.0
129.90	9528.0
130.90	1733.0
132.00	1100.0
132.90	444.0
133.90	3517.0
135.00	9947.0

Instrument :
BNA_N
ClientSampleId :
DFTPP268

135.90	3584.0
137.00	5271.0
137.90	1052.0
138.90	652.0
140.00	1331.0
140.90	15063.0
141.90	4779.0
142.90	3449.0
144.00	939.0
145.00	715.0
146.00	2621.0
147.00	7362.0
148.00	17944.0
149.00	3572.0
150.00	955.0
151.00	1948.0
151.90	1344.0
152.90	5153.0
153.90	3339.0
155.00	7836.0
156.00	11742.0
157.00	2500.0
157.90	2855.0
159.00	2382.0
160.00	5447.0
161.00	7317.0
162.00	1991.0
163.00	525.0
164.00	710.0
165.00	5711.0
166.00	4651.0
167.00	28440.0
168.00	14646.0
169.00	2682.0
170.00	929.0
171.00	1334.0
171.90	2534.0
173.00	3392.0
174.00	6175.0
175.00	11915.0
176.00	3308.0
176.90	5537.0
178.00	1828.0
179.00	21688.0
180.00	14766.0
181.00	6692.0
182.00	1187.0
182.90	496.0
183.90	1756.0
185.00	10656.0
186.00	78120.0
187.00	23104.0
188.00	2596.0
189.00	4910.0
190.00	1068.0
191.00	2449.0
192.00	6788.0
193.00	6929.0
194.00	1537.0
194.90	826.0
196.00	15245.0
197.90	507904.0
198.90	36096.0
200.00	2950.0
201.50	2076.0
202.90	4123.0
204.00	19280.0
205.00	32616.0
206.00	142464.0
207.00	19504.0
208.00	4698.0
208.90	1572.0
210.00	1929.0
211.00	5396.0
211.80	1245.0
212.90	461.0
214.10	206.0
215.00	1598.0
216.00	3156.0
216.90	36840.0
218.00	5395.0
218.90	421.0
221.00	17536.0
221.80	4627.0
223.00	7875.0
224.00	75672.0
225.00	20280.0
226.00	2452.0
226.90	33248.0
227.90	4842.0
229.00	6444.0
229.90	847.0
231.00	2723.0
231.90	505.0
232.90	501.0
234.00	2083.0

Instrument :
BNA_N
ClientSampleId :
DFTPP268

235.00	2112.0
236.00	1564.0
237.00	3186.0
238.00	519.0
238.90	1242.0
240.00	1081.0
241.00	1705.0
242.00	3899.0
243.00	3749.0
244.00	56216.0
245.00	7589.0
246.00	11444.0
247.00	2520.0
248.00	526.0
248.90	2033.0
249.80	311.0
250.90	493.0
252.00	492.0
253.00	1329.0
255.00	293504.0
256.00	45024.0
257.00	3755.0
257.90	16648.0
258.90	2614.0
260.00	457.0
261.00	445.0
263.00	193.0
264.10	467.0
265.00	7111.0
266.00	1145.0
266.80	212.0
270.00	254.0
271.00	665.0
272.00	878.0
273.00	8558.0
273.90	22520.0
275.00	119536.0
276.00	15676.0
277.00	11687.0
278.00	2231.0
278.80	351.0
281.10	190.0
282.00	249.0
282.90	1110.0
284.10	648.0
285.00	1995.0
285.90	334.0
288.90	526.0
289.90	385.0
291.00	315.0
292.00	525.0
293.00	2307.0
294.00	580.0
296.00	36104.0
297.00	5314.0
298.00	543.0
300.90	426.0
301.90	541.0
303.00	4333.0
304.00	1157.0
307.90	330.0
309.00	395.0
310.00	660.0
312.90	289.0
314.00	1575.0
314.90	4390.0
316.00	2191.0
316.90	455.0
319.90	173.0
321.00	1185.0
322.00	784.0
323.00	12265.0
323.90	2238.0
325.00	160.0
325.90	240.0
326.90	2263.0
327.90	1240.0
328.90	201.0
331.90	969.0
333.00	1094.0
334.00	6823.0
335.00	2222.0
336.00	210.0
338.90	153.0
340.00	170.0
341.00	1306.0
342.00	338.0
345.90	2490.0
346.80	588.0
351.00	236.0
351.90	3278.0
352.90	2696.0
354.00	3587.0
354.90	644.0
358.90	316.0
364.90	15723.0

Instrument :
BNA_N
ClientSampleId :
DFTPP268

365.90	2366.0
366.90	168.0
370.00	348.0
370.90	962.0
372.00	5819.0
373.00	1872.0
373.90	189.0
382.90	1791.0
383.90	574.0
384.80	169.0
389.90	869.0
390.80	554.0
392.00	338.0
401.10	395.0
401.90	2073.0
402.90	3407.0
403.90	1485.0
420.90	2431.0
422.00	3333.0
422.90	21344.0
424.00	6152.0
424.90	777.0
438.90	208.0
440.90	59096.0
442.00	442048.0
443.00	84640.0
444.00	7776.0
444.90	549.0

Instrument :

BNA_N

ClientSampleId :

DFTPP268

m/z	Abundance
37.00	907.0
38.00	3267.0
39.00	17968.0
40.00	1066.0
41.00	527.0
43.00	237.0
44.00	1651.0
45.00	506.0
48.00	185.0
49.10	2340.0
50.00	64584.0
51.00	252352.0
52.00	12491.0
53.00	618.0
55.00	1387.0
56.00	7589.0
57.00	17672.0
57.90	742.0
59.00	167.0
59.90	166.0
61.00	3392.0
62.00	3828.0
63.00	10108.0
64.00	1308.0
65.00	4294.0
65.90	307.0
67.00	444.0
68.10	4646.0
69.00	245248.0
70.00	1181.0
73.00	1922.0
74.00	26096.0
75.00	38864.0
76.00	13485.0
77.00	271104.0
78.00	18344.0
79.00	19288.0
80.00	15099.0
81.00	21128.0
82.00	4537.0
83.00	4611.0
83.90	767.0
84.90	3774.0
86.00	5711.0
86.90	2707.0
87.90	1131.0
89.00	382.0
90.90	4275.0
91.90	4827.0
93.00	34024.0
94.00	2332.0
95.00	603.0
96.00	1387.0
98.00	26032.0
99.00	19496.0
100.00	1785.0
101.00	11287.0
102.00	546.0
102.90	3797.0
104.00	7821.0
105.00	6997.0
106.10	2427.0
107.00	90888.0
108.00	13847.0
109.10	2809.0
110.00	161856.0
111.00	23960.0
111.90	2695.0
113.00	1147.0
114.10	157.0
114.90	309.0
116.00	5108.0
117.00	74744.0
118.00	5274.0
119.00	639.0
120.00	1143.0
121.00	444.0
122.00	5994.0
123.00	9076.0
123.90	4209.0
124.90	4016.0
127.00	299968.0
128.00	23624.0
129.00	122240.0
130.00	10499.0
131.00	1892.0
132.00	979.0
132.90	484.0
134.00	3820.0
135.00	9839.0
136.00	3827.0
137.00	5297.0

Instrument :
BNA_N
ClientSampleId :
DFTPP268

138.00	1255.0
138.90	699.0
140.00	1773.0
141.00	16198.0
142.00	5048.0
143.00	3282.0
143.90	910.0
145.00	1019.0
146.00	3015.0
147.00	7749.0
147.90	17456.0
149.00	4008.0
150.10	981.0
151.10	2212.0
151.80	1244.0
153.00	5018.0
154.00	3594.0
155.00	8976.0
156.00	13354.0
157.00	2394.0
157.90	3310.0
158.90	2211.0
160.00	5509.0
161.00	6343.0
162.00	1851.0
162.90	487.0
163.90	719.0
165.00	6182.0
166.00	4475.0
167.00	30048.0
168.00	14560.0
169.00	2588.0
170.00	1067.0
171.00	1350.0
171.90	2674.0
173.00	3949.0
174.00	6686.0
175.00	12111.0
176.00	3242.0
176.90	5559.0
177.90	1865.0
178.90	23456.0
180.00	16023.0
181.00	7333.0
182.00	1470.0
183.00	400.0
184.00	2037.0
185.00	10213.0
186.00	86928.0
187.00	24640.0
188.00	2593.0
189.00	5513.0
189.90	907.0
191.00	2753.0
191.90	7030.0
193.00	7880.0
194.00	1758.0
195.10	1033.0
196.00	15814.0
198.00	586944.0
198.90	38104.0
200.00	2809.0
201.40	2279.0
203.00	4575.0
204.00	22488.0
205.00	35512.0
206.00	161152.0
207.00	19776.0
208.00	5566.0
208.90	1571.0
210.20	2427.0
211.00	6394.0
213.00	340.0
214.00	255.0
214.90	1891.0
216.00	3343.0
217.00	40272.0
217.90	5009.0
219.10	421.0
221.00	18400.0
222.90	9591.0
224.00	83200.0
225.00	23656.0
226.00	2154.0
227.00	36808.0
228.00	4864.0
228.90	7609.0
229.90	878.0
231.00	2670.0
232.00	546.0
232.90	936.0
233.90	2121.0
235.00	2754.0
236.00	1809.0
237.00	2864.0
237.90	594.0

Instrument :
BNA_N
ClientSampleId :
DFTPP268

239.00	1554.0
239.80	1247.0
240.90	2398.0
242.00	3916.0
243.10	4739.0
244.00	62680.0
245.00	8235.0
246.00	12491.0
246.90	2652.0
247.90	492.0
249.00	2052.0
250.00	504.0
251.00	562.0
252.00	547.0
253.00	1571.0
255.00	343104.0
256.00	50088.0
257.00	3792.0
258.00	19640.0
259.00	2802.0
260.00	569.0
261.00	641.0
262.90	286.0
263.90	726.0
265.00	8041.0
265.90	1160.0
269.90	266.0
270.90	720.0
272.00	986.0
272.90	10044.0
274.00	28552.0
275.00	157824.0
275.90	19472.0
277.00	13827.0
278.00	2399.0
279.00	571.0
282.00	337.0
282.90	1437.0
284.00	953.0
285.00	2342.0
285.90	466.0
286.90	161.0
288.00	192.0
288.90	440.0
289.90	411.0
291.00	357.0
292.00	805.0
293.00	2843.0
293.90	818.0
296.00	44008.0
297.00	6167.0
297.90	617.0
301.00	627.0
301.90	770.0
303.00	5095.0
304.00	1537.0
305.00	313.0
307.90	516.0
308.90	415.0
309.90	530.0
310.80	165.0
311.90	215.0
312.80	358.0
314.00	2503.0
314.90	4969.0
315.90	2848.0
316.80	388.0
321.00	1311.0
322.10	763.0
323.00	13676.0
324.00	2864.0
325.00	282.0
326.00	424.0
326.90	2777.0
328.00	1474.0
328.80	165.0
329.00	164.0
332.00	904.0
332.90	1419.0
334.00	8808.0
335.00	2413.0
336.00	244.0
339.00	211.0
340.00	172.0
341.00	1574.0
342.00	455.0
345.90	3490.0
346.80	623.0
349.90	187.0
350.90	320.0
352.00	3625.0
353.00	3054.0
354.00	5092.0
354.90	865.0
359.00	329.0
364.90	19400.0

Instrument :
BNA_N
ClientSampleId :
DFTPP268

365.90	2685.0
366.90	255.0
369.80	378.0
371.00	1372.0
372.00	6999.0
373.00	2045.0
373.90	255.0
376.90	239.0
383.00	2008.0
384.00	499.0
385.00	194.0
389.90	1197.0
390.90	749.0
392.00	569.0
400.80	481.0
402.00	3372.0
402.90	4991.0
404.00	1670.0
404.80	321.0
409.90	174.0
414.90	215.0
420.90	5074.0
422.00	4710.0
423.00	30736.0
423.90	7099.0
425.00	986.0
439.40	385.0
441.00	99640.0
442.00	549952.0
443.00	122440.0
444.00	11634.0
444.90	649.0

Instrument :
BNA_N
ClientSampleId :
DFTPP268

m/z	Abundance
37.00	651.0
38.00	1855.0
39.00	11572.0
40.00	1198.0
41.00	388.0
43.00	264.0
44.00	1766.0
44.90	328.0
49.10	1860.0
50.00	39552.0
51.00	161152.0
52.00	8539.0
53.00	424.0
54.90	873.0
56.00	5113.0
57.00	10234.0
57.90	417.0
61.00	1832.0
62.00	2317.0
63.00	5999.0
64.00	763.0
65.10	2822.0
65.90	230.0
67.20	253.0
68.10	2800.0
69.00	148160.0
70.00	959.0
73.00	1582.0
74.00	15956.0
75.00	24992.0
76.10	8330.0
77.00	169984.0
78.00	10712.0
79.00	11590.0
80.00	8656.0
81.00	12668.0
82.00	3355.0
83.00	2900.0
83.90	578.0
85.00	2335.0
85.90	4017.0
87.00	1940.0
88.00	678.0
89.00	346.0
90.90	3044.0
91.90	3090.0
93.00	21200.0
94.00	1466.0
94.90	426.0
96.00	981.0
98.00	15922.0
99.00	12580.0
100.00	1056.0
101.00	7797.0
101.90	230.0
103.00	2587.0
103.90	5272.0
105.00	4370.0
106.00	1471.0
107.00	54304.0
107.90	8221.0
109.00	1704.0
109.90	102312.0
111.00	15339.0
112.00	1951.0
113.00	740.0
114.00	165.0
114.80	240.0
116.00	3183.0
117.00	48040.0
118.00	3348.0
119.00	434.0
120.10	754.0
121.00	284.0
121.90	3921.0
123.00	6452.0
123.90	2595.0
124.90	2497.0
127.00	204032.0
128.00	15601.0
128.90	78320.0
130.00	7010.0
131.00	1419.0
131.90	890.0
133.00	421.0
134.00	2400.0
135.00	6298.0
136.00	2298.0
136.90	3458.0
137.90	832.0
138.80	366.0
140.00	1022.0

Instrument :
BNA_N
ClientSampleId :
DFTPP268

140.90	10202.0
141.90	3202.0
143.00	2116.0
143.90	692.0
145.00	568.0
146.00	1919.0
147.00	5545.0
147.90	11935.0
149.00	2427.0
150.00	735.0
151.10	1106.0
151.80	774.0
152.90	3471.0
153.90	2261.0
155.00	5703.0
156.00	8897.0
157.00	1655.0
157.90	2051.0
158.90	1304.0
160.00	3260.0
161.00	4807.0
161.90	1495.0
162.90	437.0
163.90	537.0
164.90	4293.0
166.00	3082.0
167.00	20152.0
168.00	10231.0
169.00	1795.0
169.90	649.0
170.90	982.0
172.00	2148.0
173.00	2345.0
174.00	4250.0
175.00	8502.0
176.00	2215.0
177.00	3981.0
177.90	1475.0
179.00	15585.0
180.00	9908.0
181.00	5260.0
182.00	868.0
182.90	402.0
183.90	1434.0
185.00	7330.0
186.00	58320.0
187.00	17400.0
188.00	1687.0
189.00	3711.0
189.80	640.0
191.00	1713.0
191.90	4366.0
193.00	5353.0
194.00	1182.0
194.90	572.0
196.00	10664.0
196.90	2700.0
197.90	405056.0
198.90	27232.0
199.90	1781.0
201.40	1909.0
202.90	3250.0
204.00	14911.0
205.00	25176.0
206.00	112008.0
207.00	14938.0
208.00	3869.0
208.90	1325.0
210.00	1781.0
211.00	4193.0
212.90	184.0
214.00	200.0
215.00	1328.0
216.00	2595.0
217.00	29472.0
217.90	3298.0
219.00	315.0
221.00	13057.0
221.90	3854.0
222.90	6188.0
224.00	59112.0
225.00	14207.0
226.10	1659.0
227.00	25728.0
228.00	3487.0
229.00	5144.0
229.90	846.0
231.00	2072.0
232.00	254.0
233.00	659.0
233.90	1774.0
235.00	2099.0
235.90	1286.0
237.00	2204.0
237.90	277.0
238.90	992.0

Instrument :
BNA_N
ClientSampleId :
DFTPP268

239.90	1060.0
240.90	1335.0
241.90	2997.0
243.00	3691.0
244.00	46024.0
245.00	5738.0
246.00	9359.0
246.90	1916.0
248.00	342.0
248.90	1773.0
250.00	278.0
251.00	458.0
251.90	547.0
252.90	931.0
255.00	251136.0
256.00	34336.0
257.00	2507.0
258.00	13895.0
259.00	2374.0
259.90	425.0
261.00	321.0
263.80	321.0
264.90	6262.0
265.90	732.0
270.00	181.0
270.90	594.0
272.00	608.0
273.00	7767.0
274.00	20472.0
275.00	106744.0
276.00	13727.0
277.00	9785.0
278.00	1736.0
278.90	297.0
282.90	949.0
284.00	677.0
285.00	1784.0
286.00	278.0
289.00	432.0
289.80	207.0
290.90	284.0
292.00	421.0
293.00	2407.0
294.00	600.0
296.00	31768.0
296.90	4603.0
297.90	314.0
301.00	358.0
301.90	565.0
303.00	3361.0
303.90	927.0
304.90	230.0
308.10	341.0
308.90	307.0
310.00	329.0
312.90	373.0
314.00	1355.0
315.00	3849.0
316.00	2085.0
316.80	356.0
320.90	942.0
322.10	600.0
323.00	11040.0
323.90	1687.0
325.90	207.0
326.90	1862.0
327.90	891.0
329.00	164.0
331.90	857.0
332.90	1036.0
334.00	6946.0
335.00	1869.0
335.90	249.0
338.80	177.0
340.00	157.0
341.00	1118.0
342.00	466.0
346.00	2083.0
346.90	538.0
350.80	267.0
352.00	3442.0
353.00	2380.0
354.00	3832.0
354.90	593.0
358.80	155.0
364.90	16712.0
365.90	2153.0
369.90	353.0
371.00	771.0
372.00	5925.0
373.00	1330.0
373.90	212.0
382.90	1560.0
383.80	470.0
389.90	779.0
390.90	589.0

Instrument :
BNA_N
ClientSampleId :
DFTPP268

392.00	401.0
400.90	364.0
402.00	2941.0
402.90	3956.0
403.90	1309.0
404.90	204.0
420.90	2957.0
422.00	3673.0
423.00	23728.0
424.00	5768.0
425.00	615.0
438.70	180.0
441.00	82168.0
442.00	501888.0
443.00	104280.0
444.00	7858.0
444.90	531.0

Instrument :
BNA_N
ClientSampleId :
DFTPP268