

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072721\
 Data File : BN015706.D
 Acq On : 27 Jul 2021 16:19
 Operator : CG/JU
 Sample : M3186-03
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 AREA-2-(0-7)

Quant Time: Jul 27 16:52:00 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 27 13:51:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.689	152	15626	0.400	ng	0.00
7) Naphthalene-d8	10.470	136	73987	0.400	ng	0.00
13) Acenaphthene-d10	14.332	164	43042	0.400	ng	0.01
19) Phenanthrene-d10	17.078	188	85119	0.400	ng	0.00
29) Chrysene-d12	21.291	240	71810	0.400	ng	# 0.01
35) Perylene-d12	23.587	264	77970	0.400	ng	# 0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.310	112	14541	0.352	ng	0.02
5) Phenol-d6	6.877	99	17825	0.358	ng	0.00
8) Nitrobenzene-d5	8.835	82	20195	0.364	ng	0.00
11) 2-Methylnaphthalene-d10	12.065	152	41104	0.364	ng	0.00
14) 2,4,6-Tribromophenol	15.817	330	7241	0.389	ng	0.00
15) 2-Fluorobiphenyl	12.949	172	56206	0.330	ng	0.00
27) Fluoranthene-d10	19.118	212	86550	0.371	ng	0.00
31) Terphenyl-d14	19.723	244	67512	0.402	ng	0.00
Target Compounds						
						Qvalue
3) n-Nitrosodimethylamine	3.512	42	1266	0.106	ng	# 1
9) Naphthalene	10.521	128	55775	0.283	ng	100
12) 2-Methylnaphthalene	12.141	142	31541	0.244	ng	98
16) Acenaphthylene	14.040	152	245916	1.322	ng	99
17) Acenaphthene	14.383	154	24519	0.197	ng	94
18) Fluorene	15.385	166	51522	0.339	ng	# 80
20) 4,6-Dinitro-2-methylph...	15.461	198	37	0.135	ng	# 1
25) Phenanthrene	17.115	178	698717	2.780	ng	100
26) Anthracene	17.212	178	282234	1.279	ng	99
28) Fluoranthene	19.147	202	1252698	4.476	ng	99
30) Pyrene	19.515	202	1354950	5.413	ng	# 96
32) Benzo(a)anthracene	21.323	228	764210	3.144	ng	94
33) Chrysene	21.323	228	760712	3.138	ng	97
34) Bis(2-ethylhexyl)phtha...	21.217	149	15585	0.278	ng	# 99
36) Indeno(1,2,3-cd)pyrene	25.928	276	488203	1.638	ng	94
37) Benzo(b)fluoranthene	22.899	252	937978	3.457	ng	98
38) Benzo(k)fluoranthene	22.899	252	937978	3.379	ng	99
39) Benzo(a)pyrene	23.484	252	796878	3.266	ng	99
40) Dibenzo(a,h)anthracene	25.935	278	140297	0.575	ng	95
41) Benzo(g,h,i)perylene	26.647	276	507770	2.040	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072721\
Data File : BN015706.D
Acq On : 27 Jul 2021 16:19
Operator : CG/JU
Sample : M3186-03
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
AREA-2-(0-7)


Quant Time: Jul 27 16:52:00 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 27 13:51:38 2021
Response via : Initial Calibration








1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.689 min Scan# 500
 Delta R.T. -0.000 min
 Lab File: BN015706.D
 Acq: 27 Jul 2021 16:19

Instrument :
 BNA_N
 ClientSampleId :
 AREA-2-(0-7)







Tgt Ion:	152	Resp:	15626
Ion	Ratio	Lower	Upper
152	100		
150	159.6	125.2	187.8
115	63.3	48.6	73.0



n-Nitrosodimethylamine
 Concen: 0.106 ng
 RT: 3.512 min Scan# 47
 Delta R.T. -0.074 min
 Lab File: BN015706.D
 Acq: 27 Jul 2021 16:19

Tgt Ion:	42	Resp:	1266
Ion	Ratio	Lower	Upper
42	100		
74	1.2	126.7	190.1#
44	179.2	5.8	8.8#









2-Fluorophenol
 Concen: 0.352 ng
 RT: 5.310 min Scan# 242
 Delta R.T. 0.018 min
 Lab File: BN015706.D
 Acq: 27 Jul 2021 16:19




Instrument :
 BNA_N
 ClientSampleId :
 AREA-2-(0-7)

Tgt Ion:	112	Resp:	14541
Ion	Ratio	Lower	Upper
112	100		
64	56.5	44.8	67.2
63	28.9	22.2	33.4




Phenol-d6
 Concen: 0.358 ng
 RT: 6.877 min Scan# 412
 Delta R.T. 0.009 min
 Lab File: BN015706.D
 Acq: 27 Jul 2021 16:19

Tgt Ion:	99	Resp:	17825
Ion	Ratio	Lower	Upper
99	100		
42	16.4	13.4	20.0
71	32.2	24.5	36.7





Abundance Scan 738 (10.470 min): BN015692.D\data.ms (-738) (-)




Naphthalene-d8
Concen: 0.400 ng
RT: 10.470 min Scan# 738
Delta R.T. -0.000 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19

Instrument :
BNA_N
ClientSampleId :
AREA-2-(0-7)


Tgt Ion:	136	Resp:	73987
Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.9	13.3
54	7.3	5.0	7.4
68	6.1	4.3	6.5



Abundance Scan 738 (10.470 min): BN015706.D\data.ms (-722) (-)





Abundance Scan 609 (8.835 min): BN015692.D\data.ms (-606) (-)




Nitrobenzene-d5
Concen: 0.364 ng
RT: 8.835 min Scan# 609
Delta R.T. -0.000 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19


Tgt Ion:	82	Resp:	20195
Ion	Ratio	Lower	Upper
82	100		
128	39.0	32.0	48.0
54	49.6	38.9	58.3



Abundance Scan 609 (8.835 min): BN015706.D\data.ms (-593) (-)






Abundance Scan 742 (10.521 min): BN015692.D\data.ms (-734) (-)




Naphthalene
Concen: 0.283 ng
RT: 10.521 min Scan# 742
Delta R.T. -0.000 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19

Instrument :
BNA_N
ClientSampleId :
AREA-2-(0-7)

Tgt Ion:	128	Resp:	55775
Ion	Ratio	Lower	Upper
128	100		
129	11.3	8.9	13.3
127	13.1	10.4	15.6






Abundance Scan 945 (12.065 min): BN015692.D\data.ms (-934) (-)




2-Methylnaphthalene-d10
Concen: 0.364 ng
RT: 12.065 min Scan# 945
Delta R.T. -0.000 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19

Tgt Ion:	152	Resp:	41104
Ion	Ratio	Lower	Upper
152	100		
151	20.6	16.3	24.5





Abundance Scan 962 (12.141 min): BN015692.D\data.ms (-950) (-)




2-Methylnaphthalene
Concen: 0.244 ng
RT: 12.141 min Scan# 962
Delta R.T. -0.000 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19

Instrument :
BNA_N
ClientSampleId :
AREA-2-(0-7)


Tgt Ion:	142	Resp:	31541
Ion	Ratio	Lower	Upper
142	100		
141	87.7	69.4	104.2
115	34.4	25.9	38.9



Abundance Scan 962 (12.141 min): BN015706.D\data.ms (-940) (-)






Abundance Scan 1187 (14.319 min): BN015692.D\data.ms (-1113) (-)



Acenaphthene-d10
Concen: 0.400 ng
RT: 14.332 min Scan# 1188
Delta R.T. 0.013 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19


Tgt Ion:	164	Resp:	43042
Ion	Ratio	Lower	Upper
164	100		
162	98.0	83.6	125.4
160	42.6	38.6	58.0





Abundance Scan 1305 (15.817 min): BN015692.D\data.ms (-1305) (-)

2,4,6-Tribromophenol
Concen: 0.389 ng
RT: 15.817 min Scan# 1305
Delta R.T. -0.000 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19


Instrument :
BNA_N
ClientSampleId :
AREA-2-(0-7)



Tgt Ion:330 Resp: 7241
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 47.9 28.5 42.7#






Abundance Scan 1305 (15.817 min): BN015706.D\data.ms (-1297) (-)



Abundance Scan 1079 (12.949 min): BN015692.D\data.ms (-1079) (-)

2-Fluorobiphenyl
Concen: 0.330 ng
RT: 12.949 min Scan# 1079
Delta R.T. -0.000 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19

Tgt Ion:172 Resp: 56206
Ion Ratio Lower Upper
172 100
171 34.9 27.4 41.0
170 23.1 17.9 26.9



Abundance Scan 1079 (12.949 min): BN015706.D\data.ms (-1071) (-)

Sub

Abundance Scan 1165 (14.040 min): BN015692.D\data.ms (-1149) (-)

Acenaphthylene

Concen: 1.322 ng

RT: 14.040 min Scan# 1165

Delta R.T. -0.000 min

Lab File: BN015706.D

Acq: 27 Jul 2021 16:19

Tgt Ion:152 Resp: 245916

Ion Ratio Lower Upper

152 100

151 19.6 15.4 23.2

153 13.6 10.4 15.6

Instrument :

BNA_N

ClientSampleId :

AREA-2-(0-7)

Abundance Scan 1165 (14.040 min): BN015706.D\data.ms

Ref 50

63.0 105.0 152.0 204.0

m/z-->

Abundance Scan 1165 (14.040 min): BN015706.D\data.ms

Raw 50

63.0 105.0 152.0 204.0 330.0

m/z-->

Abundance Scan 1165 (14.040 min): BN015706.D\data.ms (-1149) (-)

Sub 50

63.0 105.0 152.0 204.0

m/z-->

Abundance

150000

100000

50000

0

14.00 14.20

Time-->

Abundance Scan 1192 (14.383 min): BN015692.D\data.ms (-1149) (-)

Acenaphthene

Concen: 0.197 ng

RT: 14.383 min Scan# 1192

Delta R.T. -0.000 min

Lab File: BN015706.D

Acq: 27 Jul 2021 16:19

Tgt Ion:154 Resp: 24519

Ion Ratio Lower Upper

154 100

153 117.5 89.0 133.6

152 59.5 43.3 64.9

Abundance Scan 1192 (14.383 min): BN015706.D\data.ms

Ref 50

63.0 105.0 153.0 198.0 330.0

m/z-->

Abundance Scan 1192 (14.383 min): BN015706.D\data.ms

Raw 50

63.0 105.0 153.0 198.0 330.0

m/z-->

Abundance Scan 1192 (14.383 min): BN015706.D\data.ms (-1184) (-)

Sub 50

63.0 105.0 153.0 206.0 331.0

m/z-->

Abundance

15000

10000


5000

0

14.40

Time-->




Abundance Scan 1270 (15.373 min): BN015692.D\data.ms (-1213) (-)




Fluorene
Concen: 0.339 ng
RT: 15.385 min Scan# 1271
Delta R.T. 0.013 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19

Instrument :
BNA_N
ClientSampleId :
AREA-2-(0-7)

Tgt Ion:166 Resp: 51522
Ion Ratio Lower Upper
166 100
165 118.0 77.9 116.9#
167 18.8 10.8 16.2#






Abundance Scan 1408 (17.078 min): BN015692.D\data.ms (-1412) (-)




Phenanthrene-d10
Concen: 0.400 ng
RT: 17.078 min Scan# 1408
Delta R.T. -0.000 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19

Tgt Ion:188 Resp: 85119
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.8 5.3 7.9



Abundance Scan 1277 (15.461 min): BN015692.D\data.ms (-1270) (-)





4,6-Dinitro-2-methylphenol
Concen: 0.135 ng
RT: 15.461 min Scan# 1277
Delta R.T. -0.000 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19


Instrument :
BNA_N
ClientSampleId :
AREA-2-(0-7)

Tgt Ion:	198	Resp:	37
Ion	Ratio	Lower	Upper
198	100		
182	160.0	16.4	24.6#
77	0.0	0.0	0.0


Abundance Scan 1277 (15.461 min): BN015706.D\data.ms



Abundance Scan 1277 (15.461 min): BN015706.D\data.ms (-1269) (-)





Abundance Scan 1411 (17.115 min): BN015692.D\data.ms (-1425) (-)




Phenanthrene
Concen: 2.780 ng
RT: 17.115 min Scan# 1411
Delta R.T. -0.000 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19

Tgt Ion:	178	Resp:	698717
Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.1	22.7
179	15.7	12.4	18.6


Abundance Scan 1411 (17.115 min): BN015706.D\data.ms



Abundance Scan 1411 (17.115 min): BN015706.D\data.ms (-1386) (-)






Abundance Scan 1418 (17.200 min): BN015692.D\data.ms (-125) (-)




Anthracene
Concen: 1.279 ng
RT: 17.212 min Scan# 1419
Delta R.T. 0.012 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19

Instrument :
BNA_N
ClientSampleId :
AREA-2-(0-7)

Tgt Ion:	178	Resp:	282234
Ion	Ratio	Lower	Upper
178	100		
176	18.1	14.6	21.8
179	15.6	12.2	18.2






Abundance Scan 1706 (19.113 min): BN015692.D\data.ms (-1627) (-)




Fluoranthene-d10
Concen: 0.371 ng
RT: 19.118 min Scan# 1707
Delta R.T. 0.005 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19

Tgt Ion:	212	Resp:	86550
Ion	Ratio	Lower	Upper
212	100		
106	14.0	11.5	17.3
104	8.2	6.3	9.5






Abundance Scan 1712 (19.142 min): BN015692.D\data.ms (-1728) (-)




Fluoranthene
Concen: 4.476 ng
RT: 19.147 min Scan# 1713
Delta R.T. 0.005 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19

Instrument :
BNA_N
ClientSampleId :
AREA-2-(0-7)

Tgt Ion:202 Resp: 1252698
Ion Ratio Lower Upper
202 100
101 11.5 9.4 14.2
203 17.8 13.8 20.8






Abundance Scan 2006 (21.280 min): BN015692.D\data.ms (-1929) (-)




Chrysene-d12
Concen: 0.400 ng
RT: 21.291 min Scan# 2007
Delta R.T. 0.011 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19

Tgt Ion:240 Resp: 71810
Ion Ratio Lower Upper
240 100
120 9.4 4.9 7.3#
236 27.2 20.4 30.6






Abundance Scan 1785 (19.505 min): BN015692.D\data.ms (-1785) (-)




Pyrene
Concen: 5.413 ng
RT: 19.515 min Scan# 1787
Delta R.T. 0.010 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19

Instrument :
BNA_N
ClientSampleId :
AREA-2-(0-7)

Tgt Ion:202 Resp: 1354950
Ion Ratio Lower Upper
202 100
200 20.8 16.4 24.6
203 21.1 13.9 20.9#






Abundance Scan 1828 (19.718 min): BN015692.D\data.ms (-1828) (-)




Terphenyl-d14
Concen: 0.402 ng
RT: 19.723 min Scan# 1829
Delta R.T. -0.000 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19

Tgt Ion:244 Resp: 67512
Ion Ratio Lower Upper
244 100
212 8.3 6.2 9.2
122 15.0 10.9 16.3






Abundance Scan 2004 (21.259 min): BN015692.D\data.ms (-1932)




Benzo(a)anthracene
Concen: 3.144 ng
RT: 21.323 min Scan# 2010
Delta R.T. 0.064 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19

Instrument :
BNA_N
ClientSampleId :
AREA-2-(0-7)

Tgt Ion:	228	Resp:	764210
Ion	Ratio	Lower	Upper
228	100		
226	30.3	21.2	31.8
229	21.4	15.5	23.3






Abundance Scan 2009 (21.312 min): BN015692.D\data.ms (-2032)



Chrysene
Concen: 3.138 ng
RT: 21.323 min Scan# 2010
Delta R.T. 0.011 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19


Tgt Ion:	228	Resp:	760712
Ion	Ratio	Lower	Upper
228	100		
226	30.3	23.4	35.0
229	21.4	15.5	23.3






Abundance Scan 1999 (21.206 min): BN015692.D\data.ms (-1999) (-)

Bis(2-ethylhexyl)phthalate
Concen: 0.278 ng
RT: 21.217 min Scan# 2000
Delta R.T. 0.011 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19

Instrument :
BNA_N
ClientSampleId :
AREA-2-(0-7)






Tgt Ion:	149	Resp:	15585
Ion	Ratio	Lower	Upper
149	100		
167	27.9	22.6	34.0
279	4.6	2.8	4.2#



Abundance Scan 2450 (23.563 min): BN015692.D\data.ms (-2450) (-)

Perylene-d12
Concen: 0.400 ng
RT: 23.587 min Scan# 2457
Delta R.T. 0.017 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19


Tgt Ion:	264	Resp:	77970
Ion	Ratio	Lower	Upper
264	100		
260	25.8	19.8	29.8
265	33.8	19.0	28.6#






Abundance Scan 3132 (25.898 min): BN015692.D\data.ms (-3106) (-)

Indeno(1,2,3-cd)pyrene
Concen: 1.638 ng
RT: 25.928 min Scan# 3141
Delta R.T. 0.020 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19

Instrument :
BNA_N
ClientSampleId :
AREA-2-(0-7)




Tgt Ion:	276	Resp:	488203
Ion	Ratio	Lower	Upper
276	100		
138	23.3	23.0	34.4
277	24.9	20.0	30.0






Abundance Scan 2248 (22.872 min): BN015692.D\data.ms (-2192) (-)


Benzo(b)fluoranthene
Concen: 3.457 ng
RT: 22.899 min Scan# 2256
Delta R.T. 0.021 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19



Tgt Ion:	252	Resp:	937978
Ion	Ratio	Lower	Upper
252	100		
253	22.6	17.5	26.3
125	11.8	8.9	13.3





Abundance Scan 2261 (22.916 min): BN015692.D\data.ms (-2175) (-)




Benzo(k)fluoranthene
Concen: 3.379 ng
RT: 22.899 min Scan# 2256
Delta R.T. -0.024 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19

Tgt Ion:252 Resp: 937978
Ion Ratio Lower Upper
252 100
253 22.6 17.6 26.4
125 11.8 9.4 14.2


Instrument :
BNA_N
ClientSampleId :
AREA-2-(0-7)



Abundance Scan 2256 (22.899 min): BN015706.D\data.ms (-2175) (-)





Abundance Scan 2420 (23.460 min): BN015692.D\data.ms (-2175) (-)




Benzo(a)pyrene
Concen: 3.266 ng
RT: 23.484 min Scan# 2427
Delta R.T. 0.014 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19


Tgt Ion:252 Resp: 796878
Ion Ratio Lower Upper
252 100
253 22.4 17.7 26.5
125 12.0 10.2 15.2



Abundance Scan 2427 (23.484 min): BN015706.D\data.ms (-2365) (-)





Abundance Scan 3138 (25.918 min): BN015692.D\data.ms (-3116) (-)




Dibenzo(a,h)anthracene
Concen: 0.575 ng
RT: 25.935 min Scan# 3143
Delta R.T. 0.014 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19

Instrument :
BNA_N
ClientSampleId :
AREA-2-(0-7)


Tgt Ion:278 Resp: 140297
Ion Ratio Lower Upper
278 100
139 21.4 15.1 22.7
279 26.0 19.0 28.4



Abundance Scan 3143 (25.935 min): BN015706.D\data.ms (-3116) (-)





Abundance Scan 3341 (26.613 min): BN015692.D\data.ms (-3341) (-)



Benzo(g,h,i)perylene
Concen: 2.040 ng
RT: 26.647 min Scan# 3351
Delta R.T. 0.027 min
Lab File: BN015706.D
Acq: 27 Jul 2021 16:19

Tgt Ion:276 Resp: 507770
Ion Ratio Lower Upper
276 100
277 24.3 19.4 29.0
138 23.5 19.0 28.6



Abundance Scan 3351 (26.647 min): BN015706.D\data.ms (-3285) (-)

