

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN072820\
 Data File : BN011284.D
 Acq On : 28 Jul 2020 17:31
 Operator : CG/JU
 Sample : L3403-06
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 W-1

Quant Time: Jul 30 13:43:19 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN072820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 28 14:38:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.48	152	1994	0.40	ng	0.00
7) Naphthalene-d8	10.22	136	7427	0.40	ng	0.00
13) Acenaphthene-d10	14.10	164	5330	0.40	ng	0.00
19) Phenanthrene-d10	16.86	188	10705	0.40	ng	0.00
29) Chrysene-d12	21.09	240	9422	0.40	ng	0.00
36) Perylene-d12	23.22	264	8827	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.15	112	325	0.06	ng	0.00
5) Phenol-d6	6.68	99	268	0.04	ng	0.00
8) Nitrobenzene-d5	8.63	82	1731	0.26	ng	0.00
11) 2-Methylnaphthalene-d10	11.82	152	2599	0.19	ng	0.00
14) 2,4,6-Tribromophenol	15.61	330	330	0.13	ng	0.00
15) 2-Fluorobiphenyl	12.72	172	8109	0.35	ng	0.00
27) Fluoranthene-d10	18.91	212	8902	0.27	ng	0.00
31) Terphenyl-d14	19.52	244	8616	0.30	ng	0.00
Target Compounds						
16) Acenaphthylene	13.82	152	719	0.026	ng	Qvalue # 81
28) Fluoranthene	18.94	202	1749	0.044	ng	# 95
30) Pyrene	19.30	202	4156	0.098	ng	98
34) Bis(2-ethylhexyl)phthalate	21.04	149	7465	0.441	ng	100
35) Indeno(1,2,3-cd)pyrene	25.30	276	1569	0.045	ng	98
41) Benzo(q,h,i)perylene	25.93	276	1966	0.056	ng	# 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN072820\
Data File : BN011284.D
Acq On : 28 Jul 2020 17:31
Operator : CG/JU
Sample : L3403-06
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
W-1

Quant Time: Jul 30 13:43:19 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN072820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 28 14:38:05 2020
Response via : Initial Calibration

