

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN072820\
 Data File : BN011280.D
 Acq On : 28 Jul 2020 14:50
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN072820

Quant Time: Jul 28 15:34:34 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN072820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 28 14:38:05 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	139	0.00
2	1,4-Dioxane	0.563	0.530	5.9	127	0.00
3	n-Nitrosodimethylamine	0.289	0.278	3.8	158#	0.00
4 S	2-Fluorophenol	1.051	0.962	8.5	142	0.00
5 S	Phenol-d6	1.278	1.181	7.6	126	0.00
6	bis(2-Chloroethyl)ether	1.064	0.965	9.3	125	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	135	0.00
8 S	Nitrobenzene-d5	0.357	0.335	6.2	137	-0.01
9	Naphthalene	1.186	1.113	6.2	132	0.00
10	Hexachlorobutadiene	0.287	0.265	7.7	128	0.00
11 SURR	2-Methylnaphthalene-d10	0.730	0.675	7.5	135	0.00
12	2-Methylnaphthalene	0.841	0.772	8.2	134	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	118	0.00
14 S	2,4,6-Tribromophenol	0.185	0.151	18.4	115	0.00
15 S	2-Fluorobiphenyl	1.727	1.683	2.5	133	0.00
16	Acenaphthylene	2.059	1.827	11.3	128	0.00
17	Acenaphthene	1.343	1.258	6.3	122	0.00
18	Fluorene	1.723	1.609	6.6	129	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
20	4,6-Dinitro-2-methylphenol	0.055	0.044	20.0	130	0.00
21	4-Bromophenyl-phenylether	0.265	0.258	2.6	128	0.00
22	Hexachlorobenzene	0.301	0.288	4.3	122	0.00
23	Atrazine	0.206	0.185	10.2	123	0.00
24	Pentachlorophenol	0.093	0.079	15.1	126	0.00
25	Phenanthrene	1.302	1.202	7.7	110	0.00
26	Anthracene	1.082	0.949	12.3	118	0.00
27 SURR	Fluoranthene-d10	1.221	1.053	13.8	118	0.00
28	Fluoranthene	1.480	1.262	14.7	118	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
30	Pyrene	1.803	1.807	-0.2	115	0.00
31 S	Terphenyl-d14	1.211	1.231	-1.7	117	0.00
32	Benzo(a)anthracene	1.341	1.369	-2.1	112	0.01
33	Chrysene	1.504	1.485	1.3	106	0.01
34	Bis(2-ethylhexyl)phthalate	0.719	0.646	10.2	97	0.01
35	Indeno(1,2,3-cd)pyrene	1.469	1.658	-12.9	129	0.01
36 I	Perylene-d12	1.000	1.000	0.0	118	0.01
37	Benzo(b)fluoranthene	1.595	1.439	9.8	122	0.01
38	Benzo(k)fluoranthene	1.756	1.723	1.9	124	0.01
39 C	Benzo(a)pyrene	1.384	1.298	6.2	123	0.01
40	Dibenzo(a,h)anthracene	1.432	1.437	-0.3	137	0.01
41	Benzo(g,h,i)perylene	1.588	1.498	5.7	122	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN072820\
Data File : BN011280.D
Acq On : 28 Jul 2020 14:50
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN072820

Quant Time: Jul 28 15:34:34 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN072820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 28 14:38:05 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0