

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN072820\
 Data File : BN011285.D
 Acq On : 28 Jul 2020 18:09
 Operator : CG/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDC00.4

Quant Time: Jul 28 18:45:13 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN072820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 28 14:38:05 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	129	0.00
2	1,4-Dioxane	0.563	0.584	-3.7	131	0.00
3	n-Nitrosodimethylamine	0.289	0.287	0.7	151#	0.00
4 S	2-Fluorophenol	1.051	1.027	2.3	141	0.00
5 S	Phenol-d6	1.278	1.353	-5.9	134	0.00
6	bis(2-Chloroethyl)ether	1.064	1.001	5.9	121	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	128	0.00
8 S	Nitrobenzene-d5	0.357	0.319	10.6	124	0.00
9	Naphthalene	1.186	1.142	3.7	128	0.00
10	Hexachlorobutadiene	0.287	0.264	8.0	122	0.00
11 SURR	2-Methylnaphthalene-d10	0.730	0.640	12.3	121	0.00
12	2-Methylnaphthalene	0.841	0.739	12.1	122	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.185	0.148	20.0	99	0.00
15 S	2-Fluorobiphenyl	1.727	1.729	-0.1	119	0.00
16	Acenaphthylene	2.059	1.960	4.8	120	0.00
17	Acenaphthene	1.343	1.261	6.1	107	0.00
18	Fluorene	1.723	1.556	9.7	108	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.00
20	4,6-Dinitro-2-methylphenol	0.055	0.056	-1.8	136	0.00
21	4-Bromophenyl-phenylether	0.265	0.246	7.2	101	0.00
22	Hexachlorobenzene	0.301	0.297	1.3	104	0.00
23	Atrazine	0.206	0.181	12.1	100	0.00
24	Pentachlorophenol	0.093	0.084	9.7	110	0.00
25	Phenanthrene	1.302	1.182	9.2	90	0.00
26	Anthracene	1.082	0.977	9.7	100	0.00
27 SURR	Fluoranthene-d10	1.221	1.086	11.1	101	0.00
28	Fluoranthene	1.480	1.285	13.2	99	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	105	-0.01
30	Pyrene	1.803	1.517	15.9	95	0.00
31 S	Terphenyl-d14	1.211	1.032	14.8	97	0.00
32	Benzo(a)anthracene	1.341	1.375	-2.5	111	0.00
33	Chrysene	1.504	1.413	6.1	100	0.00
34	Bis(2-ethylhexyl)phthalate	0.719	0.553	23.1	82	0.00
35	Indeno(1,2,3-cd)pyrene	1.469	1.824	-24.2	140	0.00
36 I	Perylene-d12	1.000	1.000	0.0	126	0.00
37	Benzo(b)fluoranthene	1.595	1.438	9.8	130	0.00
38	Benzo(k)fluoranthene	1.756	1.545	12.0	118	0.00
39 C	Benzo(a)pyrene	1.384	1.263	8.7	127	0.00
40	Dibenzo(a,h)anthracene	1.432	1.409	1.6	143	-0.01
41	Benzo(g,h,i)perylene	1.588	1.408	11.3	122	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN072820\
Data File : BN011285.D
Acq On : 28 Jul 2020 18:09
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 28 18:45:13 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN072820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 28 14:38:05 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0