

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072921\
 Data File : BN015716.D
 Acq On : 29 Jul 2021 20:15
 Operator : CG/JU
 Sample : SSTDIC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC3.2

Quant Time: Jul 30 02:15:03 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 29 19:08:51 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.790	152	38560	0.40	ng	0.00
7) Naphthalene-d8	10.571	136	159240	0.40	ng	0.00
13) Acenaphthene-d10	14.421	164	92236	0.40	ng	0.00
19) Phenanthrene-d10	17.164	188	186354	0.40	ng	#-0.01
29) Chrysene-d12	21.376	240	175254	0.40	ng	0.00
35) Perylene-d12	23.717	264	184145	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.393	112	305839	3.00	ng	0.00
5) Phenol-d6	6.960	99	382147	3.11	ng	0.00
8) Nitrobenzene-d5	8.937	82	346051	2.90	ng	0.00
11) 2-Methylnaphthalene-d10	12.167	152	790257	3.25	ng	0.00
14) 2,4,6-Tribromophenol	15.918	330	129466	3.25	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	1067792	2.93	ng	0.00
27) Fluoranthene-d10	19.207	212	1669088	3.27	ng	0.00
31) Terphenyl-d14	19.813	244	1353279	3.31	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.346	88	41013	3.50	ng	# 100
3) n-Nitrosodimethylamine	3.696	42	97374	3.31	ng	# 1
6) bis(2-Chloroethyl)ether	7.228	93	317208	2.89	ng	98
9) Naphthalene	10.622	128	1283705	3.03	ng	99
10) Hexachlorobutadiene	10.926	225	251364	3.17	ng	# 100
12) 2-Methylnaphthalene	12.238	142	847582	3.04	ng	99
16) Acenaphthylene	14.142	152	1325088	3.32	ng	100
17) Acenaphthene	14.484	154	838484	3.15	ng	97
18) Fluorene	15.474	166	1040247	3.20	ng	99
20) 4,6-Dinitro-2-methylph...	15.550	198	74033	7.46	ng	# 83
21) 4-Bromophenyl-phenylether	16.373	248	350026	3.25	ng	95
22) Hexachlorobenzene	16.482	284	343548	2.90	ng	99
23) Atrazine	16.640	200	325802	3.99	ng	99
24) Pentachlorophenol	16.823	266	151991	3.60	ng	99
25) Phenanthrene	17.212	178	1610090	2.93	ng	100
26) Anthracene	17.297	178	1542207	3.19	ng	100
28) Fluoranthene	19.237	202	1786555	2.92	ng	100
30) Pyrene	19.604	202	1887986	3.09	ng	100
32) Benzo(a)anthracene	21.355	228	1915109	3.23	ng	99
33) Chrysene	21.408	228	1787428	3.02	ng	99
34) Bis(2-ethylhexyl)phtha...	21.302	149	1194432	3.75	ng	99
36) Indeno(1,2,3-cd)pyrene	26.120	276	2157175	3.06	ng	98
37) Benzo(b)fluoranthene	23.009	252	1918528	2.99	ng	98
38) Benzo(k)fluoranthene	23.053	252	1924114	2.93	ng	# 92
39) Benzo(a)pyrene	23.611	252	1781838	3.09	ng	98
40) Dibenzo(a,h)anthracene	26.141	278	1811474	3.15	ng	# 94
41) Benzo(g,h,i)perylene	26.856	276	1838026	3.13	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072921\
Data File : BN015716.D
Acq On : 29 Jul 2021 20:15
Operator : CG/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Quant Time: Jul 30 02:15:03 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 29 19:08:51 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.790 min Scan# 511
 Delta R.T. -0.000 min
 Lab File: BN015716.D
 Acq: 29 Jul 2021 20:15




Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC3.2


Tgt Ion:152 Resp: 38560
 Ion Ratio Lower Upper
 152 100
 150 156.8 124.8 187.2
 115 61.0 49.3 73.9



#2
 1,4-Dioxane
 Concen: 3.50 ng
 RT: 3.346 min Scan# 29
 Delta R.T. -0.009 min
 Lab File: BN015716.D
 Acq: 29 Jul 2021 20:15

Tgt Ion: 88 Resp: 41013
 Ion Ratio Lower Upper
 88 100
 43 28.2 0.0 0.0#
 58 78.3 0.0 0.0#









#3
n-Nitrosodimethylamine
Concen: 3.31 ng
RT: 3.696 min Scan# 67
Delta R.T. -0.028 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 97374
Ion Ratio Lower Upper
42 100
74 148.0 34.8 52.2#
44 6.6 2.6 4.0#



#4
2-Fluorophenol
Concen: 3.00 ng
RT: 5.393 min Scan# 251
Delta R.T. 0.009 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion: 112 Resp: 305839
Ion Ratio Lower Upper
112 100
64 55.4 44.6 67.0
63 28.5 23.1 34.7









#5
Phenol-d6
Concen: 3.11 ng
RT: 6.960 min Scan# 421
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 382147
Ion Ratio Lower Upper
99 100
42 17.0 13.8 20.6
71 29.8 24.4 36.6



#6
bis(2-Chloroethyl)ether
Concen: 2.89 ng
RT: 7.228 min Scan# 450
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion: 93 Resp: 317208
Ion Ratio Lower Upper
93 100
63 74.9 61.4 92.2
95 32.8 27.0 40.4









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.571 min Scan# 746
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	159240
Ion Ratio	Lower	Upper	
136	100		
137	11.1	9.0	13.4
54	8.6	7.4	11.2
68	6.9	6.0	9.0



#8
Nitrobenzene-d5
Concen: 2.90 ng
RT: 8.937 min Scan# 617
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:	82	Resp:	346051
Ion Ratio	Lower	Upper	
82	100		
128	41.2	31.8	47.8
54	49.1	42.1	63.1









#9
Naphthalene
Concen: 3.03 ng
RT: 10.622 min Scan# 750
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:128 Resp: 1283705
Ion Ratio Lower Upper
128 100
129 11.1 9.0 13.6
127 12.9 10.6 15.8



#10
Hexachlorobutadiene
Concen: 3.17 ng
RT: 10.926 min Scan# 774
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:225 Resp: 251364
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.9 51.0 76.4









#11
2-Methylnaphthalene-d10
Concen: 3.25 ng
RT: 12.167 min Scan# 968
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 790257
Ion Ratio Lower Upper
152 100
151 19.9 16.0 24.0



#12
2-Methylnaphthalene
Concen: 3.04 ng
RT: 12.238 min Scan# 984
Delta R.T. -0.004 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:142 Resp: 847582
Ion Ratio Lower Upper
142 100
141 88.0 69.9 104.9
115 33.4 26.8 40.2









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.421 min Scan# 1195
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:164 Resp: 92236
Ion Ratio Lower Upper
164 100
162 99.9 80.1 120.1
160 43.7 35.2 52.8



#14
2,4,6-Tribromophenol
Concen: 3.25 ng
RT: 15.918 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:330 Resp: 129466
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 35.7 29.1 43.7









#15
2-Fluorobiphenyl
Concen: 2.93 ng
RT: 13.051 min Scan# 1087
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:172 Resp: 1067792
Ion Ratio Lower Upper
172 100
171 33.2 27.0 40.4
170 21.2 17.6 26.4



#16
Acenaphthylene
Concen: 3.32 ng
RT: 14.142 min Scan# 1173
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:152 Resp: 1325088
Ion Ratio Lower Upper
152 100
151 19.4 15.7 23.5
153 13.2 10.4 15.6









#17
Acenaphthene
Concen: 3.15 ng
RT: 14.484 min Scan# 1200
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 838484
Ion Ratio Lower Upper
154 100
153 108.4 89.6 134.4
152 51.0 42.6 63.8



#18
Fluorene
Concen: 3.20 ng
RT: 15.474 min Scan# 1278
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:166 Resp: 1040247
Ion Ratio Lower Upper
166 100
165 96.2 78.0 117.0
167 13.6 10.8 16.2









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.164 min Scan# 1415
Delta R.T. -0.012 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:188 Resp: 186354
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.9 5.0 7.4#



#20
4,6-Dinitro-2-methylphenol
Concen: 7.46 ng
RT: 15.550 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:198 Resp: 74033
Ion Ratio Lower Upper
198 100
182 5.9 10.1 15.1#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 3.25 ng
RT: 16.373 min Scan# 1350
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:248 Resp: 350026
Ion Ratio Lower Upper
248 100
250 106.5 82.2 123.4
141 28.3 26.9 40.3



#22
Hexachlorobenzene
Concen: 2.90 ng
RT: 16.482 min Scan# 1359
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:284 Resp: 343548
Ion Ratio Lower Upper
284 100
142 34.9 28.6 43.0
249 30.8 24.9 37.3





#23

Atrazine

Concen: 3.99 ng

RT: 16.640 min Scan# 1372

Delta R.T. 0.000 min

Lab File: BN015716.D

Acq: 29 Jul 2021 20:15

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:200 Resp: 325802

Ion Ratio Lower Upper

200 100

173 23.6 19.6 29.4

215 48.8 39.1 58.7



#24

Pentachlorophenol

Concen: 3.60 ng

RT: 16.823 min Scan# 1387

Delta R.T. 0.000 min

Lab File: BN015716.D

Acq: 29 Jul 2021 20:15




Tgt Ion:266 Resp: 151991


Ion Ratio Lower Upper

266 100

264 62.8 50.1 75.1

268 63.5 51.4 77.2





#25

Phenanthrene

Concen: 2.93 ng

RT: 17.212 min Scan# 1419

Delta R.T. 0.000 min

Lab File: BN015716.D

Acq: 29 Jul 2021 20:15

Tgt Ion:178 Resp: 1610090

Ion Ratio Lower Upper

178 100

176 18.8 15.2 22.8





179 15.5 12.3 18.5

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2



#26

Anthracene

Concen: 3.19 ng

RT: 17.297 min Scan# 1426

Delta R.T. 0.000 min

Lab File: BN015716.D

Acq: 29 Jul 2021 20:15




Tgt Ion:178 Resp: 1542207


Ion Ratio Lower Upper

178 100

176 18.3 14.6 22.0

179 15.5 12.2 18.4









#27
Fluoranthene-d10
Concen: 3.27 ng
RT: 19.207 min Scan# 1725
Delta R.T. -0.005 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:212 Resp: 1669088
Ion Ratio Lower Upper
212 100
106 12.0 9.6 14.4
104 6.6 5.4 8.0



#28
Fluoranthene
Concen: 2.92 ng
RT: 19.237 min Scan# 1731
Delta R.T. -0.005 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:202 Resp: 1786555
Ion Ratio Lower Upper
202 100
101 10.5 8.3 12.5
203 17.6 13.9 20.9









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.376 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:240 Resp: 175254
Ion Ratio Lower Upper
240 100
120 6.0 5.4 8.0
236 25.5 20.3 30.5



#30
Pyrene
Concen: 3.09 ng
RT: 19.604 min Scan# 1805
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15


Tgt Ion:202 Resp: 1887986
Ion Ratio Lower Upper
202 100
200 20.6 16.5 24.7
203 17.8 14.1 21.1








#31
Terphenyl-d14
Concen: 3.31 ng
RT: 19.813 min Scan# 1847
Delta R.T. -0.005 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15


Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2





Tgt Ion:244 Resp: 1353279
Ion Ratio Lower Upper
244 100
212 7.6 5.8 8.8
122 12.0 8.4 12.6




#32
Benzo(a)anthracene
Concen: 3.23 ng
RT: 21.355 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15



Tgt Ion:228 Resp: 1915109
Ion Ratio Lower Upper
228 100
226 26.8 22.2 33.2
229 20.0 15.8 23.6









#33
Chrysene
Concen: 3.02 ng
RT: 21.408 min Scan# 2018
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:228 Resp: 1787428
Ion Ratio Lower Upper
228 100
226 29.8 24.1 36.1
229 20.0 15.4 23.0



#34
Bis(2-ethylhexyl)phthalate
Concen: 3.75 ng
RT: 21.302 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:149 Resp: 1194432
Ion Ratio Lower Upper
149 100
167 28.7 22.5 33.7
279 4.3 3.5 5.3









#35
Perylene-d12
Concen: 0.40 ng
RT: 23.717 min Scan# 2495
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:264 Resp: 184145




Ion	Ratio	Lower	Upper
264	100		
260	24.3	19.7	29.5
265	26.4	21.1	31.7




#36
Indeno(1,2,3-cd)pyrene
Concen: 3.06 ng
RT: 26.120 min Scan# 3197
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:276 Resp: 2157175

Ion	Ratio	Lower	Upper
276	100		
138	23.5	19.8	29.6
277	25.1	19.5	29.3









#37
Benzo(b)fluoranthene
Concen: 2.99 ng
RT: 23.009 min Scan# 2288
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 1918528
Ion Ratio Lower Upper
252 100
253 22.1 18.2 27.4
125 9.1 7.8 11.8



#38
Benzo(k)fluoranthene
Concen: 2.93 ng
RT: 23.053 min Scan# 2301
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:252 Resp: 1924114
Ion Ratio Lower Upper
252 100
253 22.0 18.7 28.1
125 9.5 12.6 18.8#









#39
Benzo(a)pyrene
Concen: 3.09 ng
RT: 23.611 min Scan# 2464
Delta R.T. -0.003 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 1781838
Ion Ratio Lower Upper
252 100
253 22.3 18.3 27.5
125 10.2 8.7 13.1



#40
Dibenzo(a,h)anthracene
Concen: 3.15 ng
RT: 26.141 min Scan# 3203
Delta R.T. -0.003 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15


Tgt Ion:278 Resp: 1811474
Ion Ratio Lower Upper
278 100
139 15.8 17.4 26.0#
279 23.7 19.3 28.9





#41
Benzo(g,h,i)perylene
Concen: 3.13 ng
RT: 26.856 min Scan# 3412
Delta R.T. 0.004 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2



Tgt Ion: 276 Resp: 1838026

Ion	Ratio	Lower	Upper
276	100		
277	23.8	19.0	28.6
138	20.0	16.4	24.6

