

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072921\
 Data File : BN015717.D
 Acq On : 29 Jul 2021 20:52
 Operator : CG/JU
 Sample : SSTDIC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC5.0

Quant Time: Jul 30 02:15:12 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 29 19:08:51 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.790	152	42256	0.40	ng	0.00
7) Naphthalene-d8	10.571	136	166925	0.40	ng	0.00
13) Acenaphthene-d10	14.421	164	97192	0.40	ng	0.00
19) Phenanthrene-d10	17.176	188	193304	0.40	ng	0.00
29) Chrysene-d12	21.376	240	182548	0.40	ng	0.00
35) Perylene-d12	23.714	264	193440	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.383	112	484671	4.34	ng	0.00
5) Phenol-d6	6.960	99	601375	4.47	ng	0.00
8) Nitrobenzene-d5	8.936	82	568026	4.54	ng	0.00
11) 2-Methylnaphthalene-d10	12.167	152	1219215	4.78	ng	0.00
14) 2,4,6-Tribromophenol	15.918	330	214337	5.10	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	1663236	4.33	ng	0.00
27) Fluoranthene-d10	19.212	212	2560653	4.83	ng	0.00
31) Terphenyl-d14	19.817	244	2076547	4.87	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.345	88	71528	5.57	ng	Qvalue # 100
3) n-Nitrosodimethylamine	3.687	42	159689	4.96	ng	# 1
6) bis(2-Chloroethyl)ether	7.227	93	495863	4.12	ng	98
9) Naphthalene	10.622	128	1993426	4.49	ng	99
10) Hexachlorobutadiene	10.926	225	393546	4.73	ng	# 99
12) 2-Methylnaphthalene	12.238	142	1314634	4.50	ng	100
16) Acenaphthylene	14.142	152	2054815	4.89	ng	99
17) Acenaphthene	14.484	154	1305761	4.65	ng	96
18) Fluorene	15.474	166	1595991	4.65	ng	98
20) 4,6-Dinitro-2-methylph...	15.550	198	140318	13.64	ng	# 81
21) 4-Bromophenyl-phenylether	16.372	248	552449	4.94	ng	96
22) Hexachlorobenzene	16.482	284	536687	4.36	ng	99
23) Atrazine	16.640	200	500850	5.91	ng	99
24) Pentachlorophenol	16.823	266	255745	5.85	ng	100
25) Phenanthrene	17.212	178	2464108	4.32	ng	100
26) Anthracene	17.297	178	2384410	4.76	ng	100
28) Fluoranthene	19.242	202	2759803	4.34	ng	99
30) Pyrene	19.604	202	2894012	4.55	ng	99
32) Benzo(a)anthracene	21.355	228	2947505	4.77	ng	99
33) Chrysene	21.408	228	2789591	4.53	ng	99
34) Bis(2-ethylhexyl)phtha...	21.302	149	1904116	5.75	ng	99
36) Indeno(1,2,3-cd)pyrene	26.120	276	3383000	4.57	ng	98
37) Benzo(b)fluoranthene	23.009	252	3015890	4.48	ng	99
38) Benzo(k)fluoranthene	23.056	252	2950385	4.28	ng	# 93
39) Benzo(a)pyrene	23.611	252	2783611	4.60	ng	98
40) Dibenzo(a,h)anthracene	26.144	278	2878211	4.76	ng	# 93
41) Benzo(g,h,i)perylene	26.859	276	2881663	4.67	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072921\
Data File : BN015717.D
Acq On : 29 Jul 2021 20:52
Operator : CG/JU
Sample : SSTDIC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Quant Time: Jul 30 02:15:12 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 29 19:08:51 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.790 min Scan# 511
 Delta R.T. -0.000 min
 Lab File: BN015717.D
 Acq: 29 Jul 2021 20:52




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:152 Resp: 42256
 Ion Ratio Lower Upper
 152 100
 150 157.4 124.8 187.2
 115 59.9 49.3 73.9



#2
 1,4-Dioxane
 Concen: 5.57 ng
 RT: 3.345 min Scan# 29
 Delta R.T. -0.009 min
 Lab File: BN015717.D
 Acq: 29 Jul 2021 20:52

Tgt Ion: 88 Resp: 71528
 Ion Ratio Lower Upper
 88 100
 43 27.9 0.0 0.0#
 58 77.2 0.0 0.0#









#3
n-Nitrosodimethylamine
Concen: 4.96 ng
RT: 3.687 min Scan# 66
Delta R.T. -0.037 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 42 Resp: 159689
Ion Ratio Lower Upper
42 100
74 148.7 34.8 52.2#
44 6.4 2.6 4.0#



#4
2-Fluorophenol
Concen: 4.34 ng
RT: 5.383 min Scan# 250
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52

Tgt Ion: 112 Resp: 484671
Ion Ratio Lower Upper
112 100
64 55.2 44.6 67.0
63 28.4 23.1 34.7









#5
Phenol-d6
Concen: 4.47 ng
RT: 6.960 min Scan# 421
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 99 Resp: 601375
Ion Ratio Lower Upper
99 100
42 16.8 13.8 20.6
71 29.7 24.4 36.6



#6
bis(2-Chloroethyl)ether
Concen: 4.12 ng
RT: 7.227 min Scan# 450
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52

Tgt Ion: 93 Resp: 495863
Ion Ratio Lower Upper
93 100
63 74.6 61.4 92.2
95 32.8 27.0 40.4









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.571 min Scan# 746
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	136	Resp:	166925
Ion	Ratio	Lower	Upper
136	100		
137	11.1	9.0	13.4
54	8.2	7.4	11.2
68	6.6	6.0	9.0



#8
Nitrobenzene-d5
Concen: 4.54 ng
RT: 8.936 min Scan# 617
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52

Tgt Ion:	82	Resp:	568026
Ion	Ratio	Lower	Upper
82	100		
128	42.5	31.8	47.8
54	48.9	42.1	63.1









#9
Naphthalene
Concen: 4.49 ng
RT: 10.622 min Scan# 750
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:128 Resp: 1993426
Ion Ratio Lower Upper
128 100
129 11.2 9.0 13.6
127 12.8 10.6 15.8



#10
Hexachlorobutadiene
Concen: 4.73 ng
RT: 10.926 min Scan# 774
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52

Tgt Ion:225 Resp: 393546
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.2 51.0 76.4









#11
2-Methylnaphthalene-d10
Concen: 4.78 ng
RT: 12.167 min Scan# 968
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 1219215
Ion Ratio Lower Upper
152 100
151 20.0 16.0 24.0



#12
2-Methylnaphthalene
Concen: 4.50 ng
RT: 12.238 min Scan# 984
Delta R.T. -0.005 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52

Tgt Ion:142 Resp: 1314634
Ion Ratio Lower Upper
142 100
141 87.7 69.9 104.9
115 33.0 26.8 40.2









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.421 min Scan# 1195
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:164 Resp: 97192
Ion Ratio Lower Upper
164 100
162 99.1 80.1 120.1
160 43.8 35.2 52.8



#14
2,4,6-Tribromophenol
Concen: 5.10 ng
RT: 15.918 min Scan# 1313
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52

Tgt Ion:330 Resp: 214337
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 35.1 29.1 43.7









#15
2-Fluorobiphenyl
Concen: 4.33 ng
RT: 13.051 min Scan# 1087
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:172 Resp: 1663236
Ion Ratio Lower Upper
172 100
171 33.0 27.0 40.4
170 21.4 17.6 26.4



#16
Acenaphthylene
Concen: 4.89 ng
RT: 14.142 min Scan# 1173
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52

Tgt Ion:152 Resp: 2054815
Ion Ratio Lower Upper
152 100
151 19.4 15.7 23.5
153 13.3 10.4 15.6









#17
Acenaphthene
Concen: 4.65 ng
RT: 14.484 min Scan# 1200
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:154 Resp: 1305761
Ion Ratio Lower Upper
154 100
153 107.0 89.6 134.4
152 50.4 42.6 63.8



#18
Fluorene
Concen: 4.65 ng
RT: 15.474 min Scan# 1278
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52

Tgt Ion:166 Resp: 1595991
Ion Ratio Lower Upper
166 100
165 95.6 78.0 117.0
167 13.6 10.8 16.2









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.176 min Scan# 1416
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:188 Resp: 193304
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.1 5.0 7.4



#20
4,6-Dinitro-2-methylphenol
Concen: 13.64 ng
RT: 15.550 min Scan# 1284
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52

Tgt Ion:198 Resp: 140318
Ion Ratio Lower Upper
198 100
182 5.1 10.1 15.1#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 4.94 ng
RT: 16.372 min Scan# 1350
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:248 Resp: 552449
Ion Ratio Lower Upper
248 100
250 106.0 82.2 123.4
141 30.4 26.9 40.3



#22
Hexachlorobenzene
Concen: 4.36 ng
RT: 16.482 min Scan# 1359
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52

Tgt Ion:284 Resp: 536687
Ion Ratio Lower Upper
284 100
142 34.8 28.6 43.0
249 30.7 24.9 37.3









#23
Atrazine
Concen: 5.91 ng
RT: 16.640 min Scan# 1372
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:200 Resp: 500850
Ion Ratio Lower Upper
200 100
173 25.8 19.6 29.4
215 48.6 39.1 58.7



#24
Pentachlorophenol
Concen: 5.85 ng
RT: 16.823 min Scan# 1387
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52

Tgt Ion:266 Resp: 255745
Ion Ratio Lower Upper
266 100
264 62.8 50.1 75.1
268 63.9 51.4 77.2









#25
Phenanthrene
Concen: 4.32 ng
RT: 17.212 min Scan# 1419
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:178 Resp: 2464108
Ion Ratio Lower Upper
178 100
176 18.9 15.2 22.8
179 15.6 12.3 18.5



#26
Anthracene
Concen: 4.76 ng
RT: 17.297 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52

Tgt Ion:178 Resp: 2384410
Ion Ratio Lower Upper
178 100
176 18.3 14.6 22.0
179 15.7 12.2 18.4









#27
Fluoranthene-d10
Concen: 4.83 ng
RT: 19.212 min Scan# 1726
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:212 Resp: 2560653
Ion Ratio Lower Upper
212 100
106 12.1 9.6 14.4
104 6.7 5.4 8.0



#28
Fluoranthene
Concen: 4.34 ng
RT: 19.242 min Scan# 1732
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52

Tgt Ion:202 Resp: 2759803
Ion Ratio Lower Upper
202 100
101 10.6 8.3 12.5
203 17.7 13.9 20.9









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.376 min Scan# 2015
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:240 Resp: 182548
Ion Ratio Lower Upper
240 100
120 6.2 5.4 8.0
236 25.5 20.3 30.5



#30
Pyrene
Concen: 4.55 ng
RT: 19.604 min Scan# 1805
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52

Tgt Ion:202 Resp: 2894012
Ion Ratio Lower Upper
202 100
200 20.7 16.5 24.7
203 18.1 14.1 21.1









#31
 Terphenyl-d14
 Concen: 4.87 ng
 RT: 19.817 min Scan# 1848
 Delta R.T. -0.000 min
 Lab File: BN015717.D
 Acq: 29 Jul 2021 20:52




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:244 Resp: 2076547
 Ion Ratio Lower Upper
 244 100
 212 7.4 5.8 8.8
 122 10.5 8.4 12.6



#32
 Benzo(a)anthracene
 Concen: 4.77 ng
 RT: 21.355 min Scan# 2013
 Delta R.T. -0.000 min
 Lab File: BN015717.D
 Acq: 29 Jul 2021 20:52

Tgt Ion:228 Resp: 2947505
 Ion Ratio Lower Upper
 228 100
 226 27.1 22.2 33.2
 229 20.3 15.8 23.6









#33
Chrysene
Concen: 4.53 ng
RT: 21.408 min Scan# 2018
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:228 Resp: 2789591
Ion Ratio Lower Upper
228 100
226 30.0 24.1 36.1
229 19.9 15.4 23.0



#34
Bis(2-ethylhexyl)phthalate
Concen: 5.75 ng
RT: 21.302 min Scan# 2008
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52

Tgt Ion:149 Resp: 1904116
Ion Ratio Lower Upper
149 100
167 28.6 22.5 33.7
279 4.3 3.5 5.3









#35
Perylene-d12
Concen: 0.40 ng
RT: 23.714 min Scan# 2494
Delta R.T. -0.004 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:264 Resp: 193440




Ion	Ratio	Lower	Upper
264	100		
260	24.8	19.7	29.5
265	26.0	21.1	31.7




#36
Indeno(1,2,3-cd)pyrene
Concen: 4.57 ng
RT: 26.120 min Scan# 3197
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52

Tgt Ion:276 Resp: 3383000

Ion	Ratio	Lower	Upper
276	100		
138	23.7	19.8	29.6
277	25.2	19.5	29.3









#37
Benzo(b)fluoranthene
Concen: 4.48 ng
RT: 23.009 min Scan# 2288
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:252 Resp: 3015890
Ion Ratio Lower Upper
252 100
253 22.5 18.2 27.4
125 9.2 7.8 11.8



#38
Benzo(k)fluoranthene
Concen: 4.28 ng
RT: 23.056 min Scan# 2302
Delta R.T. 0.003 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52


Tgt Ion:252 Resp: 2950385
Ion Ratio Lower Upper
252 100
253 22.4 18.7 28.1
125 9.4 12.6 18.8#








#39
Benzo(a)pyrene
Concen: 4.60 ng
RT: 23.611 min Scan# 2464
Delta R.T. -0.003 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52


Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0





Tgt Ion:252 Resp: 2783611
Ion Ratio Lower Upper
252 100
253 22.1 18.3 27.5
125 10.2 8.7 13.1




#40
Dibenzo(a,h)anthracene
Concen: 4.76 ng
RT: 26.144 min Scan# 3204
Delta R.T. -0.000 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52




Tgt Ion:278 Resp: 2878211
Ion Ratio Lower Upper
278 100
139 15.8 17.4 26.0#
279 23.5 19.3 28.9





#41
Benzo(g,h,i)perylene
Concen: 4.67 ng
RT: 26.859 min Scan# 3413
Delta R.T. 0.007 min
Lab File: BN015717.D
Acq: 29 Jul 2021 20:52

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0



Tgt Ion:276 Resp: 2881663
Ion Ratio Lower Upper
276 100
277 24.0 19.0 28.6
138 20.1 16.4 24.6

