

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072921\
 Data File : BN015718.D
 Acq On : 29 Jul 2021 21:29
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN072921

Quant Time: Jul 30 02:36:27 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 30 02:33:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	113	0.00
2	1,4-Dioxane	0.140	0.150	-7.1	108	0.00
3	n-Nitrosodimethylamine	0.274	0.251	8.4	116	0.00
4 S	2-Fluorophenol	1.076	1.050	2.4	111	0.00
5 S	Phenol-d6	1.361	1.310	3.7	109	0.00
6	bis(2-Chloroethyl)ether	1.069	1.066	0.3	112	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
8 S	Nitrobenzene-d5	0.268	0.248	7.5	114	0.00
9	Naphthalene	1.038	1.011	2.6	110	0.00
10	Hexachlorobutadiene	0.201	0.198	1.5	111	0.00
11 SURR	2-Methylnaphthalene-d10	0.629	0.621	1.3	111	0.00
12	2-Methylnaphthalene	0.703	0.692	1.6	110	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.00
14 S	2,4,6-Tribromophenol	0.162	0.145	10.5	114	0.00
15 S	2-Fluorobiphenyl	1.524	1.445	5.2	111	0.00
16	Acenaphthylene	1.818	1.711	5.9	108	0.00
17	Acenaphthene	1.132	1.087	4.0	111	0.00
18	Fluorene	1.423	1.365	4.1	110	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
20	4,6-Dinitro-2-methylphenol	0.030	0.029	3.3	133	0.00
21	4-Bromophenyl-phenylether	0.238	0.226	5.0	110	0.00
22	Hexachlorobenzene	0.232	0.228	1.7	113	0.00
23	Atrazine	0.214	0.192	10.3	106	0.00
24	Pentachlorophenol	0.085	0.078	8.2	117	0.00
25	Phenanthrene	1.103	1.072	2.8	112	0.00
26	Anthracene	1.067	1.000	6.3	107	0.00
27 SURR	Fluoranthene-d10	1.132	1.107	2.2	112	0.00
28	Fluoranthene	1.261	1.228	2.6	111	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	112	0.00
30	Pyrene	1.368	1.282	6.3	111	0.00
31 S	Terphenyl-d14	1.017	0.926	8.9	111	0.00
32	Benzo(a)anthracene	1.358	1.276	6.0	113	0.00
33	Chrysene	1.290	1.207	6.4	110	0.00
34	Bis(2-ethylhexyl)phthalate	0.778	0.638	18.0	105	0.00
35 I	Perylene-d12	1.000	1.000	0.0	114	0.00
36	Indeno(1,2,3-cd)pyrene	1.459	1.411	3.3	114	0.00
37	Benzo(b)fluoranthene	1.355	1.292	4.6	113	0.00
38	Benzo(k)fluoranthene	1.336	1.296	3.0	114	0.00
39 C	Benzo(a)pyrene	1.224	1.137	7.1	111	0.00
40	Dibenzo(a,h)anthracene	1.229	1.205	2.0	115	0.00
41	Benzo(g,h,i)perylene	1.240	1.184	4.5	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN072921