

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072921\
 Data File : BN015716.D
 Acq On : 29 Jul 2021 20:15
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Jul 30 02:15:03 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 29 19:08:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.790	152	38560	0.400	ng	0.00
7) Naphthalene-d8	10.571	136	159240	0.400	ng	0.00
13) Acenaphthene-d10	14.421	164	92236	0.400	ng	0.00
19) Phenanthrene-d10	17.164	188	186354	0.400	ng	#-0.01
29) Chrysene-d12	21.376	240	175254	0.400	ng	0.00
35) Perylene-d12	23.717	264	184145	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.393	112	305839	2.999	ng	0.00
5) Phenol-d6	6.960	99	382147	3.114	ng	0.00
8) Nitrobenzene-d5	8.937	82	346051	2.898	ng	0.00
11) 2-Methylnaphthalene-d10	12.167	152	790257	3.247	ng	0.00
14) 2,4,6-Tribromophenol	15.918	330	129466	3.247	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	1067792	2.930	ng	0.00
27) Fluoranthene-d10	19.207	212	1669088	3.267	ng	0.00
31) Terphenyl-d14	19.813	244	1353279	3.306	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.346	88	41013	3.501	ng	# 100
3) n-Nitrosodimethylamine	3.696	42	97374	3.311	ng	# 1
6) bis(2-Chloroethyl)ether	7.228	93	317208	2.887	ng	98
9) Naphthalene	10.622	128	1283705	3.032	ng	99
10) Hexachlorobutadiene	10.926	225	251364	3.166	ng	# 100
12) 2-Methylnaphthalene	12.238	142	847582	3.043	ng	99
16) Acenaphthylene	14.142	152	1325088	3.323	ng	100
17) Acenaphthene	14.484	154	838484	3.150	ng	97
18) Fluorene	15.474	166	1040247	3.196	ng	99
20) 4,6-Dinitro-2-methylph...	15.550	198	74033	7.463	ng	# 83
21) 4-Bromophenyl-phenylether	16.373	248	350026	3.245	ng	95
22) Hexachlorobenzene	16.482	284	343548	2.898	ng	99
23) Atrazine	16.640	200	325802	3.987	ng	99
24) Pentachlorophenol	16.823	266	151991	3.604	ng	99
25) Phenanthrene	17.212	178	1610090	2.926	ng	100
26) Anthracene	17.297	178	1542207	3.191	ng	100
28) Fluoranthene	19.237	202	1786555	2.916	ng	100
30) Pyrene	19.604	202	1887986	3.090	ng	100
32) Benzo(a)anthracene	21.355	228	1915109	3.228	ng	99
33) Chrysene	21.408	228	1787428	3.021	ng	99
34) Bis(2-ethylhexyl)phtha...	21.302	149	1194432	3.754	ng	99
36) Indeno(1,2,3-cd)pyrene	26.120	276	2157175	3.064	ng	98
37) Benzo(b)fluoranthene	23.009	252	1918528	2.994	ng	98
38) Benzo(k)fluoranthene	23.053	252	1924114	2.935	ng	# 92
39) Benzo(a)pyrene	23.611	252	1781838	3.092	ng	98
40) Dibenzo(a,h)anthracene	26.141	278	1811474	3.146	ng	# 94
41) Benzo(g,h,i)perylene	26.856	276	1838026	3.126	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072921\
Data File : BN015716.D
Acq On : 29 Jul 2021 20:15
Operator : CG/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Quant Time: Jul 30 02:15:03 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 29 19:08:51 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.790 min Scan# 511
 Delta R.T. -0.000 min
 Lab File: BN015716.D
 Acq: 29 Jul 2021 20:15




Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC3.2


Tgt Ion:152 Resp: 38560
 Ion Ratio Lower Upper
 152 100
 150 156.8 124.8 187.2
 115 61.0 49.3 73.9



#2
 1,4-Dioxane
 Concen: 3.501 ng
 RT: 3.346 min Scan# 29
 Delta R.T. -0.009 min
 Lab File: BN015716.D
 Acq: 29 Jul 2021 20:15

Tgt Ion: 88 Resp: 41013
 Ion Ratio Lower Upper
 88 100
 43 28.2 0.0 0.0#
 58 78.3 0.0 0.0#









#3
n-Nitrosodimethylamine
Concen: 3.311 ng
RT: 3.696 min Scan# 67
Delta R.T. -0.028 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 97374
Ion Ratio Lower Upper
42 100
74 148.0 34.8 52.2#
44 6.6 2.6 4.0#



#4
2-Fluorophenol
Concen: 2.999 ng
RT: 5.393 min Scan# 251
Delta R.T. 0.009 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion: 112 Resp: 305839
Ion Ratio Lower Upper
112 100
64 55.4 44.6 67.0
63 28.5 23.1 34.7









#5
Phenol-d6
Concen: 3.114 ng
RT: 6.960 min Scan# 421
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 382147
Ion Ratio Lower Upper
99 100
42 17.0 13.8 20.6
71 29.8 24.4 36.6



#6
bis(2-Chloroethyl)ether
Concen: 2.887 ng
RT: 7.228 min Scan# 450
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion: 93 Resp: 317208
Ion Ratio Lower Upper
93 100
63 74.9 61.4 92.2
95 32.8 27.0 40.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.571 min Scan# 746
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	159240
Ion	Ratio	Lower	Upper
136	100		
137	11.1	9.0	13.4
54	8.6	7.4	11.2
68	6.9	6.0	9.0



#8
Nitrobenzene-d5
Concen: 2.898 ng
RT: 8.937 min Scan# 617
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:	82	Resp:	346051
Ion	Ratio	Lower	Upper
82	100		
128	41.2	31.8	47.8
54	49.1	42.1	63.1









#9
Naphthalene
Concen: 3.032 ng
RT: 10.622 min Scan# 750
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:128 Resp: 1283705
Ion Ratio Lower Upper
128 100
129 11.1 9.0 13.6
127 12.9 10.6 15.8



#10
Hexachlorobutadiene
Concen: 3.166 ng
RT: 10.926 min Scan# 774
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:225 Resp: 251364
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.9 51.0 76.4









#11
2-Methylnaphthalene-d10
Concen: 3.247 ng
RT: 12.167 min Scan# 968
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 790257
Ion Ratio Lower Upper
152 100
151 19.9 16.0 24.0



#12
2-Methylnaphthalene
Concen: 3.043 ng
RT: 12.238 min Scan# 984
Delta R.T. -0.004 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:142 Resp: 847582
Ion Ratio Lower Upper
142 100
141 88.0 69.9 104.9
115 33.4 26.8 40.2









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.421 min Scan# 1195
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:164 Resp: 92236
Ion Ratio Lower Upper
164 100
162 99.9 80.1 120.1
160 43.7 35.2 52.8



#14
2,4,6-Tribromophenol
Concen: 3.247 ng
RT: 15.918 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:330 Resp: 129466
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 35.7 29.1 43.7









#15
2-Fluorobiphenyl
Concen: 2.930 ng
RT: 13.051 min Scan# 1087
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:172 Resp: 1067792




Ion	Ratio	Lower	Upper
172	100		
171	33.2	27.0	40.4
170	21.2	17.6	26.4




#16
Acenaphthylene
Concen: 3.323 ng
RT: 14.142 min Scan# 1173
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:152 Resp: 1325088

Ion	Ratio	Lower	Upper
152	100		
151	19.4	15.7	23.5
153	13.2	10.4	15.6









#17
Acenaphthene
Concen: 3.150 ng
RT: 14.484 min Scan# 1200
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 838484
Ion Ratio Lower Upper
154 100
153 108.4 89.6 134.4
152 51.0 42.6 63.8



#18
Fluorene
Concen: 3.196 ng
RT: 15.474 min Scan# 1278
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:166 Resp: 1040247
Ion Ratio Lower Upper
166 100
165 96.2 78.0 117.0
167 13.6 10.8 16.2









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.164 min Scan# 1415
Delta R.T. -0.012 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:188 Resp: 186354
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.9 5.0 7.4#



#20
4,6-Dinitro-2-methylphenol
Concen: 7.463 ng
RT: 15.550 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:198 Resp: 74033
Ion Ratio Lower Upper
198 100
182 5.9 10.1 15.1#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 3.245 ng
RT: 16.373 min Scan# 1350
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:248 Resp: 350026
Ion Ratio Lower Upper
248 100
250 106.5 82.2 123.4
141 28.3 26.9 40.3



#22
Hexachlorobenzene
Concen: 2.898 ng
RT: 16.482 min Scan# 1359
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:284 Resp: 343548
Ion Ratio Lower Upper
284 100
142 34.9 28.6 43.0
249 30.8 24.9 37.3









#23
Atrazine
Concen: 3.987 ng
RT: 16.640 min Scan# 1372
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:200 Resp: 325802
Ion Ratio Lower Upper
200 100
173 23.6 19.6 29.4
215 48.8 39.1 58.7



#24
Pentachlorophenol
Concen: 3.604 ng
RT: 16.823 min Scan# 1387
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:266 Resp: 151991
Ion Ratio Lower Upper
266 100
264 62.8 50.1 75.1
268 63.5 51.4 77.2









#25
Phenanthrene
Concen: 2.926 ng
RT: 17.212 min Scan# 1419
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:178 Resp: 1610090
Ion Ratio Lower Upper
178 100
176 18.8 15.2 22.8
179 15.5 12.3 18.5



#26
Anthracene
Concen: 3.191 ng
RT: 17.297 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:178 Resp: 1542207
Ion Ratio Lower Upper
178 100
176 18.3 14.6 22.0
179 15.5 12.2 18.4









#27
Fluoranthene-d10
Concen: 3.267 ng
RT: 19.207 min Scan# 1725
Delta R.T. -0.005 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:212 Resp: 1669088
Ion Ratio Lower Upper
212 100
106 12.0 9.6 14.4
104 6.6 5.4 8.0



#28
Fluoranthene
Concen: 2.916 ng
RT: 19.237 min Scan# 1731
Delta R.T. -0.005 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:202 Resp: 1786555
Ion Ratio Lower Upper
202 100
101 10.5 8.3 12.5
203 17.6 13.9 20.9









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.376 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:240 Resp: 175254
Ion Ratio Lower Upper
240 100
120 6.0 5.4 8.0
236 25.5 20.3 30.5



#30
Pyrene
Concen: 3.090 ng
RT: 19.604 min Scan# 1805
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:202 Resp: 1887986
Ion Ratio Lower Upper
202 100
200 20.6 16.5 24.7
203 17.8 14.1 21.1









#31
 Terphenyl-d14
 Concen: 3.306 ng
 RT: 19.813 min Scan# 1847
 Delta R.T. -0.005 min
 Lab File: BN015716.D
 Acq: 29 Jul 2021 20:15




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:244 Resp: 1353279
 Ion Ratio Lower Upper
 244 100
 212 7.6 5.8 8.8
 122 12.0 8.4 12.6



#32
 Benzo(a)anthracene
 Concen: 3.228 ng
 RT: 21.355 min Scan# 2013
 Delta R.T. 0.000 min
 Lab File: BN015716.D
 Acq: 29 Jul 2021 20:15

Tgt Ion:228 Resp: 1915109
 Ion Ratio Lower Upper
 228 100
 226 26.8 22.2 33.2
 229 20.0 15.8 23.6









#33
Chrysene
Concen: 3.021 ng
RT: 21.408 min Scan# 2018
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:228 Resp: 1787428
Ion Ratio Lower Upper
228 100
226 29.8 24.1 36.1
229 20.0 15.4 23.0



#34
Bis(2-ethylhexyl)phthalate
Concen: 3.754 ng
RT: 21.302 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15


Tgt Ion:149 Resp: 1194432
Ion Ratio Lower Upper
149 100
167 28.7 22.5 33.7
279 4.3 3.5 5.3







#35
Perylene-d12
Concen: 0.400 ng
RT: 23.717 min Scan# 2495
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2







Tgt Ion:264 Resp: 184145
Ion Ratio Lower Upper
264 100
260 24.3 19.7 29.5
265 26.4 21.1 31.7



#36
Indeno(1,2,3-cd)pyrene
Concen: 3.064 ng
RT: 26.120 min Scan# 3197
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Tgt Ion:276 Resp: 2157175
Ion Ratio Lower Upper
276 100
138 23.5 19.8 29.6
277 25.1 19.5 29.3









#37
Benzo(b)fluoranthene
Concen: 2.994 ng
RT: 23.009 min Scan# 2288
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 1918528
Ion Ratio Lower Upper
252 100
253 22.1 18.2 27.4
125 9.1 7.8 11.8



#38
Benzo(k)fluoranthene
Concen: 2.935 ng
RT: 23.053 min Scan# 2301
Delta R.T. 0.000 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15


Tgt Ion:252 Resp: 1924114
Ion Ratio Lower Upper
252 100
253 22.0 18.7 28.1
125 9.5 12.6 18.8#








#39
Benzo(a)pyrene
Concen: 3.092 ng
RT: 23.611 min Scan# 2464
Delta R.T. -0.003 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2







Tgt Ion:252 Resp: 1781838
Ion Ratio Lower Upper
252 100
253 22.3 18.3 27.5
125 10.2 8.7 13.1



#40
Dibenzo(a,h)anthracene
Concen: 3.146 ng
RT: 26.141 min Scan# 3203
Delta R.T. -0.003 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15


Tgt Ion:278 Resp: 1811474
Ion Ratio Lower Upper
278 100
139 15.8 17.4 26.0#
279 23.7 19.3 28.9





#41
Benzo(g,h,i)perylene
Concen: 3.126 ng
RT: 26.856 min Scan# 3412
Delta R.T. 0.004 min
Lab File: BN015716.D
Acq: 29 Jul 2021 20:15

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2



Tgt Ion:276 Resp: 1838026
Ion Ratio Lower Upper
276 100
277 23.8 19.0 28.6
138 20.0 16.4 24.6

