

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN073020\
 Data File : BN011306.D
 Acq On : 30 Jul 2020 19:52
 Operator : CG/JU
 Sample : L3388-11
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AD7

Quant Time: Jul 31 03:00:31 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN073020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 31 02:58:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.47	152	1327	0.40	ng/ul	0.00
2) Naphthalene-d8	10.21	136	5113	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.10	164	3158	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.86	188	7177	0.40	ng/ul	0.00
16) Chrysene-d12	21.07	240	7151	0.40	ng/ul	0.00
20) Perylene-d12	23.19	264	7899	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.81	152	2653	0.21	ng/ul	0.00
14) Fluoranthene-d10	18.89	212	7162	0.21	ng/ul	0.00
Target Compounds				Ovalue		
11) Pentachlorophenol	16.52	266	376	0.214	ng/ul	96
12) Phenanthrene	16.89	178	892	0.038	ng/ul#	84
15) Fluoranthene	18.93	202	2163	0.074	ng/ul	88
17) Pyrene	19.29	202	2341	0.072	ng/ul#	90
18) Benzo(a)anthracene	21.06	228	1769	0.064	ng/ul#	91
19) Chrysene	21.11	228	1971	0.063	ng/ul#	91
21) Benzo(b)fluoranthene	22.57	252	2332	0.071	ng/ul#	55
22) Benzo(k)fluoranthene	22.61	252	2130	0.058	ng/ul#	5
23) Benzo(a)pyrene	23.10	252	1788	0.062	ng/ul#	50
24) Indeno(1,2,3-cd)pyrene	25.27	276	2149	0.071	ng/ul#	96
25) Dibenzo(a,h)anthracene	25.29	278	1495	0.062	ng/ul#	53
26) Benzo(g,h,i)perylene	25.90	276	1734	0.064	ng/ul#	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN073020\
Data File : BN011306.D
Acq On : 30 Jul 2020 19:52
Operator : CG/JU
Sample : L3388-11
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AD7

Quant Time: Jul 31 03:00:31 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN073020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 31 02:58:38 2020
Response via : Initial Calibration

