

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN073020\
 Data File : BN011297.D
 Acq On : 30 Jul 2020 13:41
 Operator : CG/JU
 Sample : SST00.402
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST00.402

Quant Time: Jul 30 14:14:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN073020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 30 14:13:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.47	152	1332	0.40	ng/ul	0.00
2) Naphthalene-d8	10.22	136	4521	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.10	164	2795	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.86	188	6308	0.40	ng/ul	0.00
16) Chrysene-d12	21.08	240	5990	0.40	ng/ul	0.00
20) Perylene-d12	23.20	264	5992	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.82	152	4816	0.43	ng/ul	0.00
14) Fluoranthene-d10	18.90	212	12832	0.47	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.27	128	5714	0.422	ng/ul	100
5) 2-Methylnaphthalene	11.89	142	3791	0.429	ng/ul	100
7) Acenaphthylene	13.82	152	5585	0.412	ng/ul	100
8) Acenaphthene	14.16	153	4463	0.419	ng/ul	100
9) Fluorene	15.16	166	5193	0.430	ng/ul	100
11) Pentachlorophenol	16.52	266	621	0.461	ng/ul	100
12) Phenanthrene	16.90	178	8722	0.419	ng/ul	100
13) Anthracene	16.99	178	7610	0.444	ng/ul	100
15) Fluoranthene	18.93	202	10843	0.467	ng/ul	100
17) Pyrene	19.29	202	10950	0.342	ng/ul	100
18) Benzo(a)anthracene	21.06	228	9687	0.457	ng/ul	100
19) Chrysene	21.11	228	11220	0.398	ng/ul	100
21) Benzo(b)fluoranthene	22.58	252	10172	0.393	ng/ul	100
22) Benzo(k)fluoranthene	22.62	252	11887	0.400	ng/ul	100
23) Benzo(a)pyrene	23.11	252	9178	0.422	ng/ul	100
24) Indeno(1,2,3-cd)pyrene	25.28	276	9678	0.377	ng/ul	98
25) Dibenzo(a,h)anthracene	25.29	278	7605	0.393	ng/ul	100
26) Benzo(g,h,i)perylene	25.91	276	8623	0.338	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN073020\
Data File : BN011297.D
Acq On : 30 Jul 2020 13:41
Operator : CG/JU
Sample : SST00.402
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.402

Quant Time: Jul 30 14:14:39 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN073020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 30 14:13:20 2020
Response via : Initial Calibration

