


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN073020\
 Data File : BN011294.D
 Acq On : 30 Jul 2020 11:53
 Operator : CG/JU
 Sample : DFTPP58
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP58

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN073020.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jul 31 02:58:38 2020



AutoFind: Scans 2438, 2439, 2440; Background Corrected with Scan 2431

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.1	162389	PASS
68	69	0.00	2	2.0	5529	PASS
69	198	0.00	100	42.9	277674	PASS
70	69	0.00	2	0.6	1608	PASS
127	198	10	80	49.3	319232	PASS
197	198	0.00	2	0.1	781	PASS
198	198	100	100	100.0	647253	PASS
199	198	5	9	6.5	41909	PASS
275	198	10	60	27.0	174698	PASS
365	198	1	100	4.4	28424	PASS
441	198	0.01	100	11.2	72573	PASS
442	198	50	100	66.6	431104	PASS
443	442	15	24	20.0	86186	PASS

Scan 2431 (17.286 min): BN011294.D
DFTPP58

m/z	Abundance
39.90	594.0
41.10	336.0
43.00	360.0
44.00	1193.0
73.00	1222.0
78.00	371.0
83.70	411.0
147.00	468.0
192.90	354.0
207.00	1751.0
208.00	616.0
280.90	667.0

Instrument :
BNA_N
ClientSampleId :
DFTPP58

m/z	Abundance
35.90	316.0
36.90	1078.0
38.00	2412.0
39.00	13373.0
40.00	997.0
41.00	440.0
42.90	326.0
43.90	1082.0
49.10	1717.0
50.00	42968.0
51.00	158272.0
52.00	8453.0
53.00	382.0
55.00	1504.0
55.90	4781.0
57.00	10500.0
58.00	490.0
61.00	2555.0
61.90	3084.0
63.00	9066.0
63.90	1566.0
65.00	3583.0
66.10	592.0
67.00	480.0
68.00	5802.0
68.90	255040.0
70.00	1824.0
71.00	488.0
73.10	3742.0
74.00	31960.0
75.00	45656.0
76.00	15350.0
77.00	293504.0
78.00	21184.0
79.00	22160.0
80.00	16288.0
80.90	22280.0
81.90	5423.0
82.90	3708.0
83.90	1139.0
85.00	3984.0
85.90	8169.0
87.00	2870.0
87.90	973.0
88.90	612.0
91.00	5017.0
92.00	6001.0
93.00	40496.0
94.00	2148.0
95.00	766.0
96.10	1302.0
97.00	913.0
98.00	33672.0
98.90	24584.0
100.00	1944.0
101.00	11579.0
101.90	666.0
102.90	4459.0
104.00	8178.0
105.00	7155.0
105.90	1955.0
107.00	90768.0
108.00	13881.0
109.10	2581.0
110.00	143040.0
111.00	24208.0
112.00	3440.0
112.90	950.0
114.90	637.0
116.00	5292.0
117.00	104968.0
117.90	6613.0
119.00	466.0
119.90	1034.0
122.00	7062.0
123.00	9150.0
124.00	4393.0
125.00	3712.0
127.00	273536.0
128.00	22760.0
129.00	109952.0
129.90	10409.0
131.00	1719.0
131.90	1352.0
132.90	624.0
133.90	3864.0
135.00	9757.0
135.90	4079.0
137.00	6186.0
137.90	1263.0
138.90	492.0
140.10	1617.0

Instrument :
BNA_N
ClientSampleId :
DFTPP58

140.90	17672.0
142.00	4954.0
143.00	3855.0
144.00	1028.0
145.00	629.0
146.00	3734.0
147.00	8428.0
147.90	23872.0
148.90	4657.0
150.10	1343.0
151.10	1710.0
151.90	1051.0
153.00	5122.0
154.00	3316.0
155.00	9438.0
156.00	12042.0
157.10	2306.0
157.90	3173.0
159.00	2182.0
159.90	5004.0
161.00	7937.0
162.00	1940.0
163.00	595.0
163.90	1031.0
165.00	7598.0
165.90	4422.0
167.00	37216.0
168.00	21880.0
168.90	2733.0
169.90	1187.0
171.10	959.0
171.90	3440.0
173.00	4409.0
174.00	7054.0
175.00	12721.0
175.90	3265.0
176.90	5740.0
178.00	2326.0
179.00	21856.0
180.00	17368.0
181.00	7484.0
181.90	1194.0
182.90	587.0
183.90	1409.0
185.00	11445.0
186.00	80064.0
187.00	24320.0
188.10	2883.0
189.00	5927.0
190.10	1174.0
190.90	2372.0
192.00	8430.0
193.00	8401.0
193.90	3062.0
194.80	923.0
195.00	779.0
196.00	14705.0
196.70	2345.0
197.90	515008.0
199.00	35040.0
200.00	3057.0
201.50	2581.0
203.00	4807.0
204.00	23968.0
205.00	40208.0
206.00	153920.0
207.00	20528.0
208.00	6014.0
208.90	2349.0
210.10	3156.0
211.00	6536.0
211.90	1543.0
213.00	437.0
213.90	335.0
215.00	1987.0
216.00	3642.0
217.00	45784.0
218.00	5731.0
219.00	518.0
221.00	19336.0
221.90	5812.0
223.00	10013.0
224.00	83664.0
225.00	23376.0
226.00	3438.0
227.00	40904.0
227.90	7116.0
229.00	10653.0
229.90	1347.0
231.00	2548.0
232.10	556.0
232.90	906.0
233.90	2484.0
235.00	3297.0
235.90	2817.0
237.00	4487.0

Instrument :
BNA_N
ClientSampleId :
DFTPP58

238.00	798.0
239.00	1569.0
240.00	1611.0
240.90	1759.0
242.00	4396.0
243.00	4877.0
244.00	59328.0
245.00	8789.0
246.00	13764.0
247.00	3251.0
248.00	508.0
249.00	1801.0
249.90	597.0
250.90	921.0
252.00	840.0
253.00	1519.0
254.00	2758.0
255.00	299584.0
256.00	51488.0
257.00	3948.0
258.00	23584.0
259.00	4445.0
260.00	529.0
260.90	758.0
263.00	344.0
263.80	420.0
265.00	10103.0
265.90	1524.0
266.90	442.0
270.00	690.0
271.00	950.0
272.00	869.0
273.00	9101.0
274.00	25128.0
275.00	133376.0
276.00	16704.0
277.00	14952.0
277.90	3468.0
278.80	799.0
281.00	921.0
281.90	482.0
283.00	1203.0
284.00	1113.0
285.00	2130.0
285.90	395.0
288.00	384.0
288.70	454.0
289.90	719.0
291.00	437.0
291.80	350.0
292.10	342.0
292.90	3005.0
294.00	848.0
295.10	1286.0
295.90	45824.0
297.00	6044.0
297.80	558.0
300.90	548.0
301.90	520.0
303.00	5024.0
304.00	947.0
308.00	754.0
308.90	474.0
309.90	526.0
313.00	625.0
313.90	1733.0
315.00	5108.0
316.00	2492.0
316.90	647.0
320.90	1246.0
322.00	782.0
323.00	11522.0
323.90	2011.0
325.90	378.0
326.90	3139.0
327.80	1115.0
331.90	1082.0
333.00	1507.0
334.00	8516.0
335.00	2129.0
335.90	425.0
338.90	328.0
340.90	1652.0
342.10	342.0
346.00	3270.0
347.00	374.0
349.80	315.0
351.20	429.0
351.90	4336.0
353.00	2904.0
354.00	3656.0
355.00	992.0
358.90	424.0
365.00	21840.0
365.90	2757.0
369.90	465.0

Instrument :
BNA_N
ClientSampleId :
DFTPP58

370.90	1038.0
371.90	5586.0
372.90	1999.0
374.00	370.0
382.90	2243.0
383.90	470.0
390.10	631.0
391.00	389.0
391.90	646.0
400.70	505.0
402.00	1946.0
403.00	3085.0
403.90	951.0
421.00	2430.0
421.90	2492.0
423.00	19568.0
424.00	5275.0
424.90	887.0
441.00	43888.0
442.00	262592.0
443.00	56672.0
444.00	4515.0

Instrument :
BNA_N
ClientSampleId :
DFTPP58

m/z	Abundance
37.00	1400.0
38.00	3035.0
39.00	16472.0
39.90	1883.0
41.00	940.0
43.00	387.0
43.90	918.0
44.80	574.0
49.00	2058.0
50.00	59552.0
51.00	196160.0
52.00	10623.0
52.90	326.0
54.90	1523.0
56.00	6630.0
57.00	13450.0
57.80	651.0
60.90	4033.0
62.00	4091.0
63.00	10580.0
64.00	2187.0
65.00	6086.0
66.80	523.0
67.90	6487.0
68.90	345856.0
70.00	1962.0
72.00	392.0
73.00	5037.0
74.00	41816.0
75.00	59624.0
76.00	20896.0
77.00	388928.0
78.00	26792.0
79.00	29536.0
80.00	20952.0
80.90	28520.0
81.90	7272.0
83.00	6173.0
84.00	1494.0
85.00	5163.0
85.90	9287.0
87.00	4245.0
87.90	1353.0
88.80	687.0
90.90	7610.0
92.00	7550.0
93.00	55352.0
93.90	3434.0
94.90	698.0
96.00	2492.0
97.00	1448.0
97.90	42504.0
99.00	30360.0
100.00	2772.0
101.00	15260.0
101.90	1122.0
102.90	4874.0
103.90	10792.0
105.00	9345.0
105.90	2686.0
107.00	122072.0
108.00	18104.0
109.10	3815.0
110.00	194496.0
111.00	33256.0
112.00	3617.0
113.00	1218.0
115.10	773.0
116.00	7556.0
117.00	142144.0
118.00	8576.0
119.00	1166.0
119.90	1485.0
121.10	790.0
122.00	9282.0
123.00	13217.0
124.00	5727.0
124.90	5964.0
127.00	387584.0
128.00	30640.0
129.00	159744.0
130.00	14088.0
131.00	2933.0
132.00	1289.0
132.90	877.0
134.00	5048.0
135.00	13922.0
136.00	4720.0
137.00	7154.0
137.90	1182.0
138.70	625.0
139.00	613.0

Instrument :
BNA_N
ClientSampleId :
DFTPP58

140.00	1804.0
141.00	23416.0
142.00	6966.0
142.90	4926.0
144.10	1503.0
144.90	636.0
145.90	4480.0
147.00	12181.0
148.00	30480.0
149.00	6459.0
149.90	1806.0
151.10	2048.0
151.70	1904.0
153.00	8591.0
153.90	6122.0
155.00	13851.0
156.00	18112.0
157.00	2780.0
157.90	4832.0
158.90	4466.0
159.90	7654.0
161.00	10369.0
161.90	3590.0
163.00	1030.0
164.00	1772.0
164.90	9265.0
166.00	7665.0
167.00	52464.0
168.00	27752.0
169.00	4597.0
169.90	1635.0
171.00	2182.0
171.90	4184.0
172.90	5297.0
174.00	10567.0
175.00	18208.0
176.00	4409.0
177.00	9564.0
177.90	2732.0
178.90	38760.0
180.00	25648.0
181.00	10093.0
182.20	2224.0
183.10	1199.0
184.00	2224.0
185.00	17368.0
186.00	124296.0
187.00	36856.0
188.00	3747.0
189.00	9452.0
189.90	1462.0
191.00	4545.0
192.00	12922.0
193.00	12600.0
194.00	2639.0
195.00	1673.0
195.90	24440.0
197.90	768576.0
198.90	51960.0
200.00	3426.0
201.60	3034.0
203.00	7165.0
204.00	34032.0
205.00	60648.0
206.00	230784.0
207.00	32568.0
208.00	9024.0
209.00	3437.0
210.10	4254.0
211.00	8868.0
211.90	1954.0
213.00	1082.0
214.90	3816.0
216.00	5947.0
216.90	70232.0
218.00	9126.0
219.00	1019.0
221.00	30376.0
221.90	8482.0
222.90	16056.0
224.00	130184.0
225.00	33496.0
226.00	4205.0
227.00	62464.0
227.90	8892.0
229.00	11278.0
230.00	1530.0
231.00	4992.0
232.00	1254.0
233.00	1273.0
234.00	4386.0
235.00	4208.0
236.00	2910.0
237.00	5276.0
238.10	800.0
238.90	2608.0

Instrument :
BNA_N
ClientSampleId :
DFTPP58

240.00	2148.0
241.00	4042.0
242.00	6525.0
243.00	7233.0
244.00	86392.0
245.00	12266.0
245.90	21280.0
247.00	5090.0
248.00	1069.0
249.00	3380.0
249.80	695.0
251.00	779.0
251.80	992.0
253.00	2044.0
255.00	529728.0
256.00	72288.0
257.00	6199.0
257.90	36080.0
258.90	4836.0
260.10	831.0
260.80	641.0
261.00	593.0
262.90	308.0
264.00	1241.0
265.00	14546.0
266.00	2372.0
267.10	338.0
269.90	503.0
270.90	1286.0
272.00	1622.0
273.00	14646.0
274.00	37400.0
275.00	208256.0
276.00	26880.0
277.00	20576.0
278.00	3871.0
278.90	750.0
280.90	920.0
282.00	899.0
283.00	2829.0
284.00	2089.0
284.90	3718.0
285.80	631.0
289.10	1131.0
289.90	563.0
290.90	660.0
292.00	679.0
293.00	5171.0
293.90	1172.0
294.30	621.0
295.00	1885.0
296.00	71056.0
296.90	9716.0
298.10	610.0
301.00	897.0
302.00	1260.0
303.00	7886.0
304.00	1910.0
308.00	1272.0
308.90	634.0
310.00	772.0
313.10	479.0
313.90	3044.0
315.00	7657.0
316.00	3755.0
316.90	808.0
319.90	489.0
321.00	1968.0
321.90	1227.0
323.00	18544.0
324.00	4438.0
325.00	377.0
326.00	456.0
327.00	4103.0
328.00	2410.0
329.00	501.0
332.00	2001.0
333.00	1930.0
334.00	14632.0
334.90	3727.0
340.10	336.0
340.90	2492.0
342.10	802.0
346.00	5582.0
346.80	1074.0
350.80	682.0
352.00	5959.0
352.90	3864.0
354.00	6156.0
355.00	1037.0
359.00	316.0
363.70	335.0
364.90	27216.0
366.00	5258.0
369.80	523.0
371.00	1463.0

Instrument :
BNA_N
ClientSampleId :
DFTPP58

372.00	7588.0
372.90	2607.0
382.80	2320.0
384.00	694.0
384.80	376.0
390.00	1492.0
390.90	947.0
392.00	662.0
401.00	910.0
401.90	4569.0
403.00	5271.0
404.00	2168.0
404.90	847.0
420.90	5540.0
422.00	5451.0
423.00	35992.0
424.00	8319.0
424.90	1180.0
439.20	398.0
441.00	83568.0
442.00	481600.0
443.00	95736.0
444.00	9773.0
445.10	880.0

Instrument :
BNA_N
ClientSampleId :
DFTPP58

m/z	Abundance
36.90	897.0
37.90	2373.0
39.00	11317.0
40.00	1328.0
41.10	577.0
42.90	482.0
44.00	747.0
44.90	327.0
49.10	1794.0
50.00	39080.0
51.00	132736.0
52.00	5800.0
53.00	388.0
55.00	850.0
56.00	5025.0
57.00	11166.0
58.00	781.0
60.90	2186.0
61.90	4349.0
63.00	7084.0
64.10	1326.0
64.90	3701.0
67.10	758.0
68.00	4300.0
68.90	232128.0
70.00	1039.0
70.90	304.0
73.00	3809.0
74.00	31208.0
75.00	40952.0
76.00	15430.0
77.00	274944.0
78.00	18648.0
79.00	20440.0
80.00	16520.0
81.00	20776.0
81.90	5323.0
83.00	4807.0
83.90	1213.0
85.00	3363.0
85.90	6213.0
87.00	2544.0
87.90	1424.0
88.90	565.0
91.00	4188.0
92.00	5388.0
93.00	39040.0
93.90	2731.0
94.90	470.0
96.00	1691.0
97.10	704.0
98.00	30960.0
99.00	22600.0
99.90	1214.0
101.00	11419.0
101.90	706.0
102.90	2814.0
104.00	7510.0
105.00	6857.0
106.00	2818.0
107.00	91368.0
108.00	13265.0
109.00	3062.0
110.00	147456.0
111.00	22112.0
111.90	3550.0
113.00	1386.0
114.70	361.0
115.00	408.0
116.00	4785.0
117.00	105968.0
118.00	8438.0
118.90	1094.0
119.90	1192.0
120.80	438.0
122.00	6191.0
123.00	10828.0
124.00	4736.0
125.00	3889.0
127.00	296576.0
128.00	22688.0
129.00	121320.0
130.00	10352.0
130.90	1481.0
132.00	1397.0
132.90	691.0
134.00	3688.0
135.00	10844.0
135.90	4171.0
137.00	5786.0
138.00	1231.0
138.90	682.0

Instrument :
BNA_N
ClientSampleId :
DFTPP58

140.00	1704.0
141.00	16309.0
142.00	4655.0
143.00	3668.0
143.90	1012.0
145.00	1001.0
146.00	3101.0
147.00	9182.0
147.90	24728.0
149.00	5018.0
150.20	1363.0
151.00	2311.0
151.90	1702.0
153.00	4871.0
154.00	4273.0
155.00	10735.0
156.00	14637.0
157.10	2722.0
158.00	3619.0
159.00	2788.0
160.00	5867.0
161.00	8614.0
162.10	1808.0
163.00	497.0
163.90	371.0
165.00	6864.0
166.00	4511.0
167.00	44552.0
168.00	22992.0
169.00	2782.0
169.90	1241.0
171.00	1917.0
172.00	3653.0
173.00	4451.0
174.00	7614.0
175.00	14607.0
176.00	3837.0
177.00	7474.0
178.00	2310.0
179.00	31112.0
180.00	19712.0
180.90	8271.0
182.00	1708.0
182.90	779.0
184.00	2491.0
184.90	12137.0
186.00	101256.0
187.00	28704.0
188.00	3425.0
188.90	8151.0
189.90	1153.0
190.90	3440.0
191.90	9147.0
193.00	10888.0
194.00	2757.0
195.00	1003.0
196.00	19384.0
197.90	658176.0
198.90	38728.0
199.90	3994.0
201.50	2865.0
203.00	5815.0
204.00	29112.0
205.00	47664.0
206.00	203968.0
207.00	24400.0
208.00	7010.0
208.90	2535.0
210.30	3606.0
211.00	8137.0
212.00	979.0
213.10	784.0
215.00	2157.0
216.00	5025.0
216.90	55536.0
217.90	6368.0
219.00	1109.0
221.00	23704.0
221.90	6641.0
223.00	13288.0
224.00	116096.0
225.00	29088.0
226.00	3373.0
227.00	58144.0
228.00	6901.0
229.00	13176.0
230.00	1614.0
231.00	3780.0
232.00	1159.0
232.90	1079.0
234.00	4180.0
235.00	3363.0
236.00	3475.0
237.00	4493.0
238.00	889.0
239.00	2077.0

Instrument :
BNA_N
ClientSampleId :
DFTPP58

240.00	2092.0
241.00	3291.0
242.00	5654.0
243.10	5756.0
244.00	86424.0
245.00	10641.0
246.00	21984.0
247.00	4972.0
247.90	909.0
248.90	3233.0
249.90	673.0
250.90	1015.0
251.90	1308.0
252.90	2244.0
254.10	3912.0
255.00	457984.0
256.00	66080.0
257.00	5852.0
258.00	31352.0
258.90	4665.0
259.90	1088.0
261.00	873.0
263.90	902.0
265.00	13350.0
265.80	1671.0
269.70	419.0
271.00	1134.0
271.90	1271.0
273.00	13132.0
274.00	40296.0
274.90	182464.0
276.00	24048.0
277.00	20328.0
278.00	3042.0
279.00	923.0
281.00	905.0
282.00	655.0
283.00	2330.0
283.90	1719.0
285.00	3890.0
286.00	799.0
288.80	856.0
289.80	428.0
290.90	545.0
292.10	787.0
292.90	4578.0
294.10	1292.0
295.10	1740.0
296.00	66328.0
296.90	8321.0
298.00	417.0
301.00	817.0
301.80	780.0
303.00	7518.0
304.00	1472.0
305.00	368.0
307.90	912.0
308.90	851.0
309.90	595.0
313.00	946.0
314.00	2919.0
315.00	6951.0
316.00	4190.0
317.00	801.0
318.80	310.0
321.00	2210.0
321.90	970.0
323.00	20120.0
323.90	3608.0
324.90	360.0
326.10	570.0
326.90	4890.0
328.00	1929.0
329.10	363.0
331.90	1736.0
333.00	2191.0
334.00	14233.0
335.10	3487.0
339.00	538.0
339.80	370.0
340.90	2929.0
341.90	611.0
345.80	4552.0
346.80	527.0
352.00	6458.0
352.90	3179.0
354.00	6010.0
355.00	1158.0
356.10	316.0
358.90	581.0
364.90	36216.0
365.90	4909.0
366.90	442.0
370.00	940.0
370.90	1645.0
371.90	10050.0

Instrument :
BNA_N
ClientSampleId :
DFTPP58

373.10	2382.0
374.10	491.0
382.90	3955.0
383.90	830.0
390.00	1575.0
390.90	1081.0
392.10	734.0
401.00	697.0
402.00	3699.0
402.90	5992.0
404.00	1929.0
404.90	391.0
415.20	349.0
421.00	5588.0
422.00	5092.0
423.00	30992.0
424.00	7438.0
425.00	1257.0
441.00	90264.0
442.00	549120.0
443.00	106152.0
444.00	9556.0
444.90	410.0

Instrument :
BNA_N
ClientSampleId :
DFTPP58