

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN073121\
 Data File : BN015744.D
 Acq On : 31 Jul 2021 21:42
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Aug 02 04:04:05 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN073021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 31 04:53:33 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.790	152	39560	0.400	ng	0.00
7) Naphthalene-d8	10.571	136	175352	0.400	ng	# 0.00
13) Acenaphthene-d10	14.421	164	93228	0.400	ng	0.00
19) Phenanthrene-d10	17.163	188	173131	0.400	ng	# 0.00
29) Chrysene-d12	21.376	240	161501	0.400	ng	# 0.00
35) Perylene-d12	23.714	264	159422	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.374	112	38132	0.359	ng	0.00
5) Phenol-d6	6.951	99	45493	0.354	ng	0.00
8) Nitrobenzene-d5	8.936	82	38209	0.347	ng	0.00
11) 2-Methylnaphthalene-d10	12.163	152	102340	0.388	ng	0.00
14) 2,4,6-Tribromophenol	15.905	330	11721	0.291	ng	-0.01
15) 2-Fluorobiphenyl	13.038	172	132720	0.372	ng	-0.01
27) Fluoranthene-d10	19.207	212	164197	0.360	ng	0.00
31) Terphenyl-d14	19.812	244	141377	0.360	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.355	88	6284	0.462	ng	87
3) n-Nitrosodimethylamine	3.714	42	9127	0.347	ng	# 99
6) bis(2-Chloroethyl)ether	7.218	93	42214	0.401	ng	100
9) Naphthalene	10.622	128	177799	0.389	ng	100
10) Hexachlorobutadiene	10.914	225	31100	0.387	ng	# 100
12) 2-Methylnaphthalene	12.234	142	115561	0.390	ng	100
16) Acenaphthylene	14.129	152	137942	0.337	ng	100
17) Acenaphthene	14.472	154	99186	0.372	ng	100
18) Fluorene	15.461	166	118159	0.364	ng	100
20) 4,6-Dinitro-2-methylph...	15.537	198	3624	0.351	ng	97
21) 4-Bromophenyl-phenylether	16.360	248	36291	0.359	ng	# 77
22) Hexachlorobenzene	16.482	284	42423	0.380	ng	99
23) Atrazine	16.628	200	28057	0.363	ng	# 92
24) Pentachlorophenol	16.823	266	19572	0.467	ng	100
25) Phenanthrene	17.212	178	188478	0.380	ng	100
26) Anthracene	17.297	178	153205	0.342	ng	100
28) Fluoranthene	19.237	202	198647	0.375	ng	100
30) Pyrene	19.599	202	194241	0.360	ng	100
32) Benzo(a)anthracene	21.355	228	182599	0.355	ng	99
33) Chrysene	21.408	228	194719	0.372	ng	99
34) Bis(2-ethylhexyl)phtha...	21.302	149	57450	0.335	ng	99
36) Indeno(1,2,3-cd)pyrene	26.117	276	213935	0.373	ng	100
37) Benzo(b)fluoranthene	23.005	252	193459	0.377	ng	99
38) Benzo(k)fluoranthene	23.053	252	209391	0.379	ng	99
39) Benzo(a)pyrene	23.611	252	162279	0.347	ng	100
40) Dibenzo(a,h)anthracene	26.137	278	182867	0.378	ng	99
41) Benzo(g,h,i)perylene	26.846	276	182740	0.366	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN073121\
Data File : BN015744.D
Acq On : 31 Jul 2021 21:42
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Aug 02 04:04:05 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN073021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jul 31 04:53:33 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.790 min Scan# 511
 Delta R.T. -0.000 min
 Lab File: BN015744.D
 Acq: 31 Jul 2021 21:42




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion: 152 Resp: 39560
 Ion Ratio Lower Upper
 152 100
 150 154.8 123.9 185.9
 115 58.4 47.8 71.6



#2
 1,4-Dioxane
 Concen: 0.462 ng
 RT: 3.355 min Scan# 30
 Delta R.T. 0.000 min
 Lab File: BN015744.D
 Acq: 31 Jul 2021 21:42

Tgt Ion: 88 Resp: 6284
 Ion Ratio Lower Upper
 88 100
 43 22.4 17.2 25.8
 58 63.5 62.2 93.2









#3
n-Nitrosodimethylamine
Concen: 0.347 ng
RT: 3.714 min Scan# 69
Delta R.T. 0.000 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 9127
Ion Ratio Lower Upper
42 100
74 146.6 116.6 174.8
44 5.1 6.6 10.0#



#4
2-Fluorophenol
Concen: 0.359 ng
RT: 5.374 min Scan# 249
Delta R.T. -0.009 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42

Tgt Ion: 112 Resp: 38132
Ion Ratio Lower Upper
112 100
64 52.1 42.1 63.1
63 27.4 21.9 32.9









#5
Phenol-d6
Concen: 0.354 ng
RT: 6.951 min Scan# 420
Delta R.T. 0.000 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	99	Resp:	45493
Ion	Ratio	Lower	Upper
99	100		
42	16.1	12.9	19.3
71	29.5	23.8	35.6



#6
bis(2-Chloroethyl)ether
Concen: 0.401 ng
RT: 7.218 min Scan# 449
Delta R.T. 0.000 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42


Tgt Ion:	93	Resp:	42214
Ion	Ratio	Lower	Upper
93	100		
63	73.3	58.6	87.8
95	33.2	26.9	40.3








#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.571 min Scan# 746
Delta R.T. 0.000 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.7	13.1
54	5.3	5.5	8.3#
68	4.3	4.5	6.7#



#8
Nitrobenzene-d5
Concen: 0.347 ng
RT: 8.936 min Scan# 617
Delta R.T. 0.000 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42

Tgt Ion	Ratio	Lower	Upper
82	100		
128	45.6	33.2	49.8
54	45.0	38.4	57.6









#9
Naphthalene
Concen: 0.389 ng
RT: 10.622 min Scan# 750
Delta R.T. 0.000 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:128 Resp: 177799
Ion Ratio Lower Upper
128 100
129 11.1 8.9 13.3
127 12.7 10.4 15.6



#10
Hexachlorobutadiene
Concen: 0.387 ng
RT: 10.914 min Scan# 773
Delta R.T. 0.000 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42

Tgt Ion:225 Resp: 31100
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.9 51.0 76.4









#11
2-Methylnaphthalene-d10
Concen: 0.388 ng
RT: 12.163 min Scan# 967
Delta R.T. -0.004 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 102340
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.390 ng
RT: 12.234 min Scan# 983
Delta R.T. -0.004 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42

Tgt Ion:142 Resp: 115561
Ion Ratio Lower Upper
142 100
141 88.5 70.5 105.7
115 33.9 27.3 40.9









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.421 min Scan# 1195
Delta R.T. 0.000 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 93228
Ion Ratio Lower Upper
164 100
162 97.6 79.8 119.6
160 42.8 35.8 53.6



#14
2,4,6-Tribromophenol
Concen: 0.291 ng
RT: 15.905 min Scan# 1312
Delta R.T. -0.013 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42

Tgt Ion:330 Resp: 11721
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 34.9 28.8 43.2









#15
2-Fluorobiphenyl
Concen: 0.372 ng
RT: 13.038 min Scan# 1086
Delta R.T. -0.013 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:172 Resp: 132720
Ion Ratio Lower Upper
172 100
171 35.4 26.9 40.3
170 23.9 17.4 26.0



#16
Acenaphthylene
Concen: 0.337 ng
RT: 14.129 min Scan# 1172
Delta R.T. -0.013 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42

Tgt Ion:152 Resp: 137942
Ion Ratio Lower Upper
152 100
151 19.4 15.7 23.5
153 12.9 10.4 15.6









#17
Acenaphthene
Concen: 0.372 ng
RT: 14.472 min Scan# 1199
Delta R.T. -0.013 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 99186
Ion Ratio Lower Upper
154 100
153 112.1 89.7 134.5
152 55.0 43.8 65.6



#18
Fluorene
Concen: 0.364 ng
RT: 15.461 min Scan# 1277
Delta R.T. -0.013 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42

Tgt Ion:166 Resp: 118159
Ion Ratio Lower Upper
166 100
165 97.8 78.1 117.1
167 13.5 10.8 16.2









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.163 min Scan# 1415
Delta R.T. 0.000 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 173131
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 9.9 10.2 15.4#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.351 ng
RT: 15.537 min Scan# 1283
Delta R.T. -0.013 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42

Tgt Ion:198 Resp: 3624
Ion Ratio Lower Upper
198 100
182 12.9 9.4 14.0
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.359 ng
RT: 16.360 min Scan# 1349
Delta R.T. -0.012 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:248 Resp: 36291
Ion Ratio Lower Upper
248 100
250 93.9 83.9 125.9
141 69.2 26.9 40.3#



#22
Hexachlorobenzene
Concen: 0.380 ng
RT: 16.482 min Scan# 1359
Delta R.T. 0.000 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42

Tgt Ion:284 Resp: 42423
Ion Ratio Lower Upper
284 100
142 36.2 29.4 44.0
249 29.7 23.8 35.6









#23
Atrazine
Concen: 0.363 ng
RT: 16.628 min Scan# 1371
Delta R.T. -0.012 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:200 Resp: 28057
Ion Ratio Lower Upper
200 100
173 31.8 18.9 28.3#
215 45.6 38.1 57.1



#24
Pentachlorophenol
Concen: 0.467 ng
RT: 16.823 min Scan# 1387
Delta R.T. 0.000 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42

Tgt Ion:266 Resp: 19572
Ion Ratio Lower Upper
266 100
264 62.6 50.2 75.2
268 64.0 51.1 76.7









#25
Phenanthrene
Concen: 0.380 ng
RT: 17.212 min Scan# 1419
Delta R.T. 0.000 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 188478
Ion Ratio Lower Upper
178 100
176 19.0 15.3 22.9
179 15.4 12.4 18.6



#26
Anthracene
Concen: 0.342 ng
RT: 17.297 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42

Tgt Ion:178 Resp: 153205
Ion Ratio Lower Upper
178 100
176 18.4 14.6 22.0
179 15.4 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.360 ng
RT: 19.207 min Scan# 1725
Delta R.T. -0.005 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 164197
Ion Ratio Lower Upper
212 100
106 13.4 10.6 16.0
104 7.4 5.9 8.9



#28
Fluoranthene
Concen: 0.375 ng
RT: 19.237 min Scan# 1731
Delta R.T. -0.005 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42

Tgt Ion:202 Resp: 198647
Ion Ratio Lower Upper
202 100
101 11.3 9.0 13.6
203 17.3 13.8 20.8





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.376 min Scan# 2015

Delta R.T. 0.000 min

Lab File: BN015744.D

Acq: 31 Jul 2021 21:42

Tgt Ion:240 Resp: 161501

Ion Ratio Lower Upper

240 100

120 6.1 7.4 11.0#





236 25.3 20.9 31.3

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4



#30

Pyrene

Concen: 0.360 ng

RT: 19.599 min Scan# 1804

Delta R.T. -0.005 min

Lab File: BN015744.D

Acq: 31 Jul 2021 21:42




Tgt Ion:202 Resp: 194241


Ion Ratio Lower Upper

202 100

200 20.6 16.5 24.7

203 17.5 14.0 21.0









#31
Terphenyl-d14
Concen: 0.360 ng
RT: 19.812 min Scan# 1847
Delta R.T. -0.005 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:244 Resp: 141377
Ion Ratio Lower Upper
244 100
212 7.5 5.9 8.9
122 13.2 9.6 14.4



#32
Benzo(a)anthracene
Concen: 0.355 ng
RT: 21.355 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42

Tgt Ion:228 Resp: 182599
Ion Ratio Lower Upper
228 100
226 26.8 21.9 32.9
229 19.6 15.4 23.0





#33

Chrysene

Concen: 0.372 ng

RT: 21.408 min Scan# 2018

Delta R.T. 0.000 min

Lab File: BN015744.D

Acq: 31 Jul 2021 21:42

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:228 Resp: 194719

Ion Ratio Lower Upper

228 100

226 29.3 24.3 36.5

229 19.4 15.3 22.9



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.335 ng

RT: 21.302 min Scan# 2008

Delta R.T. -0.010 min

Lab File: BN015744.D

Acq: 31 Jul 2021 21:42




Tgt Ion:149 Resp: 57450


Ion Ratio Lower Upper

149 100

167 26.0 21.5 32.3

279 4.0 3.2 4.8









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.714 min Scan# 2494
 Delta R.T. -0.007 min
 Lab File: BN015744.D
 Acq: 31 Jul 2021 21:42




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:264 Resp: 159422
 Ion Ratio Lower Upper
 264 100
 260 24.3 19.4 29.2
 265 23.8 19.9 29.9



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.373 ng
 RT: 26.117 min Scan# 3196
 Delta R.T. -0.010 min
 Lab File: BN015744.D
 Acq: 31 Jul 2021 21:42


Tgt Ion:276 Resp: 213935
 Ion Ratio Lower Upper
 276 100
 138 25.3 20.6 30.8
 277 25.0 20.1 30.1








#37
Benzo(b)fluoranthene
Concen: 0.377 ng
RT: 23.005 min Scan# 2287
Delta R.T. -0.007 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:252 Resp: 193459
Ion Ratio Lower Upper
252 100
253 22.1 18.1 27.1
125 10.3 8.3 12.5



#38
Benzo(k)fluoranthene
Concen: 0.379 ng
RT: 23.053 min Scan# 2301
Delta R.T. -0.003 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42

Tgt Ion:252 Resp: 209391
Ion Ratio Lower Upper
252 100
253 22.7 18.1 27.1
125 11.8 10.4 15.6









#39
Benzo(a)pyrene
Concen: 0.347 ng
RT: 23.611 min Scan# 2464
Delta R.T. -0.007 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 162279
Ion Ratio Lower Upper
252 100
253 22.4 18.1 27.1
125 11.3 9.2 13.8



#40
Dibenzo(a,h)anthracene
Concen: 0.378 ng
RT: 26.137 min Scan# 3202
Delta R.T. -0.010 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42


Tgt Ion:278 Resp: 182867
Ion Ratio Lower Upper
278 100
139 19.3 16.0 24.0
279 23.9 19.2 28.8





#41
Benzo(g,h,i)perylene
Concen: 0.366 ng
RT: 26.846 min Scan# 3409
Delta R.T. -0.014 min
Lab File: BN015744.D
Acq: 31 Jul 2021 21:42

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:276 Resp: 182740
Ion Ratio Lower Upper
276 100
277 23.7 19.2 28.8
138 20.7 16.8 25.2

