

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080221\
 Data File : BN015759.D
 Acq On : 02 Aug 2021 10:21
 Operator : CG/JU
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Quant Time: Aug 02 11:48:38 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 02 11:44:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.790	152	36600	0.400	ng	0.00
7) Naphthalene-d8	10.571	136	159651	0.400	ng	0.00
13) Acenaphthene-d10	14.408	164	79857	0.400	ng	0.00
19) Phenanthrene-d10	17.164	188	147648	0.400	ng	0.00
29) Chrysene-d12	21.366	240	129325	0.400	ng	0.00
35) Perylene-d12	23.710	264	123158	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.383	112	17102	0.174	ng	0.00
5) Phenol-d6	6.951	99	20058	0.169	ng	0.00
8) Nitrobenzene-d5	8.924	82	19087	0.190	ng	0.00
11) 2-Methylnaphthalene-d10	12.158	152	47793	0.199	ng	0.00
14) 2,4,6-Tribromophenol	15.905	330	4591	0.133	ng	0.00
15) 2-Fluorobiphenyl	13.038	172	68830	0.225	ng	0.00
27) Fluoranthene-d10	19.207	212	71917	0.185	ng	0.00
31) Terphenyl-d14	19.813	244	67558	0.215	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.355	88	2127	0.169	ng	99
3) n-Nitrosodimethylamine	3.724	42	3747	0.154	ng	# 100
6) bis(2-Chloroethyl)ether	7.218	93	19843	0.204	ng	100
9) Naphthalene	10.622	128	83937	0.202	ng	100
10) Hexachlorobutadiene	10.914	225	14650	0.200	ng	# 99
12) 2-Methylnaphthalene	12.234	142	53863	0.199	ng	99
16) Acenaphthylene	14.129	152	57468	0.164	ng	100
17) Acenaphthene	14.472	154	44926	0.197	ng	100
18) Fluorene	15.461	166	52815	0.190	ng	100
20) 4,6-Dinitro-2-methylph...	15.550	198	1134	0.100	ng	# 85
21) 4-Bromophenyl-phenylether	16.360	248	16312	0.189	ng	97
22) Hexachlorobenzene	16.470	284	19534	0.205	ng	100
23) Atrazine	16.628	200	8725	0.133	ng	98
24) Pentachlorophenol	16.823	266	4178	0.117	ng	99
25) Phenanthrene	17.200	178	86225	0.204	ng	100
26) Anthracene	17.297	178	65452	0.171	ng	100
28) Fluoranthene	19.232	202	86376	0.191	ng	100
30) Pyrene	19.599	202	85161	0.197	ng	100
32) Benzo(a)anthracene	21.355	228	70790	0.172	ng	99
33) Chrysene	21.408	228	86652	0.207	ng	100
34) Bis(2-ethylhexyl)phtha...	21.302	149	20645	0.094	ng	99
36) Indeno(1,2,3-cd)pyrene	26.113	276	79124	0.179	ng	99
37) Benzo(b)fluoranthene	23.002	252	82409	0.208	ng	98
38) Benzo(k)fluoranthene	23.050	252	87162	0.204	ng	98
39) Benzo(a)pyrene	23.608	252	61804	0.171	ng	98
40) Dibenzo(a,h)anthracene	26.130	278	66336	0.177	ng	96
41) Benzo(g,h,i)perylene	26.842	276	70796	0.183	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080221\
Data File : BN015759.D
Acq On : 02 Aug 2021 10:21
Operator : CG/JU
Sample : SSTDICC0.2
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2

Quant Time: Aug 02 11:48:38 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 02 11:44:42 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.790 min Scan# 511
 Delta R.T. -0.000 min
 Lab File: BN015759.D
 Acq: 02 Aug 2021 10:21




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2


Tgt Ion:	152	Resp:	36600
Ion	Ratio	Lower	Upper
152	100		
150	153.8	123.5	185.3
115	58.2	46.8	70.2



1,4-Dioxane
 Concen: 0.169 ng
 RT: 3.355 min Scan# 30
 Delta R.T. -0.000 min
 Lab File: BN015759.D
 Acq: 02 Aug 2021 10:21

Tgt Ion:	88	Resp:	2127
Ion	Ratio	Lower	Upper
88	100		
43	27.7	23.1	34.7
58	78.0	62.2	93.4









n-Nitrosodimethylamine
Concen: 0.154 ng
RT: 3.724 min Scan# 70
Delta R.T. 0.009 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion: 42 Resp: 3747
Ion Ratio Lower Upper
42 100
74 143.5 114.4 171.6
44 3.5 3.6 5.4#



2-Fluorophenol
Concen: 0.174 ng
RT: 5.383 min Scan# 250
Delta R.T. -0.000 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21

Tgt Ion: 112 Resp: 17102
Ion Ratio Lower Upper
112 100
64 52.8 42.2 63.4
63 28.1 22.2 33.2









Phenol-d6
 Concen: 0.169 ng
 RT: 6.951 min Scan# 420
 Delta R.T. -0.000 min
 Lab File: BN015759.D
 Acq: 02 Aug 2021 10:21




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Tgt Ion:	99	Resp:	20058
Ion	Ratio	Lower	Upper
99	100		
42	16.5	13.2	19.8
71	30.1	23.8	35.8



bis(2-Chloroethyl)ether
 Concen: 0.204 ng
 RT: 7.218 min Scan# 449
 Delta R.T. -0.000 min
 Lab File: BN015759.D
 Acq: 02 Aug 2021 10:21

Tgt Ion:	93	Resp:	19843
Ion	Ratio	Lower	Upper
93	100		
63	73.7	58.8	88.2
95	33.2	26.6	39.8



Abundance Scan 746 (10.572 min): BN015760.D\data.ms (-747) (-)

Naphthalene-d8


Concen: 0.400 ng

RT: 10.571 min Scan# 746

Delta R.T. -0.000 min

Lab File: BN015759.D

Acq: 02 Aug 2021 10:21



Tgt Ion: 136 Resp: 159651




Ion Ratio Lower Upper

136 100

137 10.9 8.7 13.1

54 4.9 3.9 5.9

68 4.0 3.3 4.9



Abundance Scan 616 (8.924 min): BN015760.D\data.ms (-613) (-)

Nitrobenzene-d5


Concen: 0.190 ng

RT: 8.924 min Scan# 616

Delta R.T. -0.000 min

Lab File: BN015759.D

Acq: 02 Aug 2021 10:21






Tgt Ion: 82 Resp: 19087

Ion Ratio Lower Upper

82 100

128 34.0 26.2 39.4


54 55.4 45.1 67.7





Abundance Scan 750 (10.622 min): BN015760.D\data.ms (-749) (-)

Naphthalene
Concen: 0.202 ng
RT: 10.622 min Scan# 750
Delta R.T. -0.000 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21


Instrument :
BNA_N
ClientSampled :
SSTDIC0.2



Tgt Ion:	128	Resp:	83937
Ion Ratio	Lower	Upper	
128	100		
129	10.9	8.8	13.2
127	12.6	10.2	15.2




Abundance Scan 750 (10.622 min): BN015759.D\data.ms (-742) (-)





Abundance Scan 773 (10.914 min): BN015760.D\data.ms (-766) (-)


Hexachlorobutadiene
Concen: 0.200 ng
RT: 10.914 min Scan# 773
Delta R.T. -0.000 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21




Tgt Ion:	225	Resp:	14650
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.9	51.5	77.3



Abundance Scan 773 (10.914 min): BN015759.D\data.ms (-765) (-)






Abundance Scan 967 (12.163 min): BN015760.D\data.ms (-953) (-)




2-Methylnaphthalene-d10
Concen: 0.199 ng
RT: 12.158 min Scan# 966
Delta R.T. -0.005 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Tgt Ion:152 Resp: 47793
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.4







Abundance Scan 983 (12.234 min): BN015760.D\data.ms (-973) (-)



2-Methylnaphthalene
Concen: 0.199 ng
RT: 12.234 min Scan# 983
Delta R.T. -0.000 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21

Tgt Ion:142 Resp: 53863
Ion Ratio Lower Upper
142 100
141 88.9 70.5 105.7
115 34.5 27.4 41.0









Scan 1194 (14.408 min): BN015760.D\data.ms (-1178) (-)
 Acenaphthene-d10
 Concen: 0.400 ng
 RT: 14.408 min Scan# 1194
 Delta R.T. -0.000 min
 Lab File: BN015759.D
 Acq: 02 Aug 2021 10:21




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2


Tgt Ion:	164	Resp:	79857
Ion	Ratio	Lower	Upper
164	100		
162	104.4	84.2	126.2
160	48.8	39.8	59.6



Scan 1312 (15.905 min): BN015760.D\data.ms (-1304) (-)
 2,4,6-Tribromophenol
 Concen: 0.133 ng
 RT: 15.905 min Scan# 1312
 Delta R.T. -0.000 min
 Lab File: BN015759.D
 Acq: 02 Aug 2021 10:21

Tgt Ion:	330	Resp:	4591
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	34.7	28.2	42.2









Scan 1086 (13.038 min): BN015760.D\data.ms (-1095) (-)
 2-Fluorobiphenyl
 Concen: 0.225 ng
 RT: 13.038 min Scan# 1086
 Delta R.T. -0.000 min
 Lab File: BN015759.D
 Acq: 02 Aug 2021 10:21




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Tgt Ion:	172	Resp:	68830
Ion	Ratio	Lower	Upper
172	100		
171	34.8	27.8	41.8
170	23.5	18.7	28.1



Scan 1172 (14.129 min): BN015760.D\data.ms (-1144) (-)
 Acenaphthylene
 Concen: 0.164 ng
 RT: 14.129 min Scan# 1172
 Delta R.T. -0.000 min
 Lab File: BN015759.D
 Acq: 02 Aug 2021 10:21


Tgt Ion:	152	Resp:	57468
Ion	Ratio	Lower	Upper
152	100		
151	19.5	15.7	23.5
153	12.9	10.5	15.7






Abundance Scan 1199 (14.472 min): BN015760.D\data.ms (-1191) (-)

Acenaphthene
Concen: 0.197 ng
RT: 14.472 min Scan# 1199
Delta R.T. -0.000 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2






Tgt Ion:	154	Resp:	44926
Ion Ratio	Lower	Upper	
154	100		
153	113.4	90.5	135.7
152	56.0	44.4	66.6



Abundance Scan 1277 (15.461 min): BN015760.D\data.ms (-1261) (-)

Fluorene
Concen: 0.190 ng
RT: 15.461 min Scan# 1277
Delta R.T. -0.000 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21


Tgt Ion:	166	Resp:	52815
Ion Ratio	Lower	Upper	
166	100		
165	97.8	78.2	117.2
167	13.5	10.7	16.1






Abundance Scan 1415 (17.164 min): BN015760.D\data.ms (-1419) (-)

Phenanthrene-d10
Concen: 0.400 ng
RT: 17.164 min Scan# 1415
Delta R.T. -0.000 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21

Instrument :
BNA_N
ClientSampled :
SSTDIC0.2




Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	9.3	7.7	11.5






Abundance Scan 1283 (15.538 min): BN015760.D\data.ms (-1228) (-)


4,6-Dinitro-2-methylphenol
Concen: 0.100 ng
RT: 15.550 min Scan# 1284
Delta R.T. 0.013 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21



Tgt Ion	Ratio	Lower	Upper
198	100		
182	19.9	11.1	16.7#
77	0.0	0.0	0.0





Abundance Scan 1349 (16.360 min): BN015760.D\data.ms (-1341) (-)




4-Bromophenyl-phenylether
Concen: 0.189 ng
RT: 16.360 min Scan# 1349
Delta R.T. -0.000 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:	248	Resp:	16312
Ion	Ratio	Lower	Upper
248	100		
250	95.5	75.8	113.6
141	61.8	52.8	79.2



Abundance Scan 1349 (16.360 min): BN015759.D\data.ms (-1341) (-)





Abundance Scan 1358 (16.470 min): BN015760.D\data.ms (-1324) (-)




Hexachlorobenzene
Concen: 0.205 ng
RT: 16.470 min Scan# 1358
Delta R.T. -0.000 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21

Tgt Ion:	284	Resp:	19534
Ion	Ratio	Lower	Upper
284	100		
142	36.5	29.4	44.2
249	29.8	23.8	35.6




Abundance Scan 1358 (16.470 min): BN015759.D\data.ms (-1350) (-)





Abundance Scan 1371 (16.628 min): BN015760.D\data.ms (-1322) (-)

Atrazine
Concen: 0.133 ng
RT: 16.628 min Scan# 1371
Delta R.T. -0.000 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2



Tgt Ion	Ratio	Lower	Upper
200	100		
173	30.3	25.0	37.6
215	46.3	36.3	54.5






Abundance Scan 1371 (16.628 min): BN015759.D\data.ms (-1330) (-)




Abundance Scan 1387 (16.823 min): BN015760.D\data.ms (-1324) (-)

Pentachlorophenol
Concen: 0.117 ng
RT: 16.823 min Scan# 1387
Delta R.T. -0.000 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21


Tgt Ion	Ratio	Lower	Upper
266	100		
264	63.1	49.7	74.5
268	64.4	51.8	77.6



Abundance Scan 1387 (16.823 min): BN015759.D\data.ms (-1379) (-)



Abundance Scan 1419 (17.212 min): BN015760.D\data.ms (-1425) (-)





Phenanthrene
Concen: 0.204 ng
RT: 17.200 min Scan# 1418
Delta R.T. -0.012 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21


Tgt Ion:	178	Resp:	86225
Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.1	22.7
179	15.6	12.2	18.4

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Abundance Scan 1418 (17.200 min): BN015759.D\data.ms



Abundance Scan 1418 (17.200 min): BN015759.D\data.ms (-1394) (-)





Abundance Scan 1426 (17.298 min): BN015760.D\data.ms (-1426) (-)




Anthracene
Concen: 0.171 ng
RT: 17.297 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21

Tgt Ion:	178	Resp:	65452
Ion	Ratio	Lower	Upper
178	100		
176	18.3	14.6	22.0
179	15.3	12.2	18.4


Abundance Scan 1426 (17.297 min): BN015759.D\data.ms



Abundance Scan 1426 (17.297 min): BN015759.D\data.ms (-1418) (-)






Abundance Scan 1725 (19.207 min): BN015760.D\data.ms (-1725) (-)




Fluoranthene-d10
Concen: 0.185 ng
RT: 19.207 min Scan# 1725
Delta R.T. -0.000 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Tgt Ion:	212	Resp:	71917
Ion	Ratio	Lower	Upper
212	100		
106	13.4	10.8	16.2
104	7.4	6.0	9.0






Abundance Scan 1731 (19.237 min): BN015760.D\data.ms (-1723) (-)




Fluoranthene
Concen: 0.191 ng
RT: 19.232 min Scan# 1730
Delta R.T. -0.005 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21

Tgt Ion:	202	Resp:	86376
Ion	Ratio	Lower	Upper
202	100		
101	11.3	9.0	13.6
203	17.2	13.8	20.8






Abundance Scan 2014 (21.366 min): BN015760.D\data.ms (-1995) (-)




Chrysene-d12
Concen: 0.400 ng
RT: 21.366 min Scan# 2014
Delta R.T. -0.000 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Tgt Ion:	240	Resp:	129325
Ion	Ratio	Lower	Upper
240	100		
120	14.2	11.4	17.0
236	27.0	21.2	31.8






Abundance Scan 1804 (19.599 min): BN015760.D\data.ms (-1764) (-)



Pyrene
Concen: 0.197 ng
RT: 19.599 min Scan# 1804
Delta R.T. -0.000 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21


Tgt Ion:	202	Resp:	85161
Ion	Ratio	Lower	Upper
202	100		
200	20.5	16.5	24.7
203	17.3	13.9	20.9






Abundance Scan 1847 (19.813 min): BN015760.D\data.ms (-1838) (-)

Terphenyl-d14
Concen: 0.215 ng
RT: 19.813 min Scan# 1847
Delta R.T. -0.000 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2




Tgt Ion	Ratio	Lower	Upper
244	100		
212	7.5	5.9	8.9
122	11.7	9.7	14.5







Abundance Scan 2013 (21.355 min): BN015760.D\data.ms (-2002) (-)

Benzo(a)anthracene
Concen: 0.172 ng
RT: 21.355 min Scan# 2013
Delta R.T. -0.000 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21



Tgt Ion	Ratio	Lower	Upper
228	100		
226	26.3	21.2	31.8
229	19.8	15.6	23.4









Chrysene
 Concen: 0.207 ng
 RT: 21.408 min Scan# 2018
 Delta R.T. -0.000 min
 Lab File: BN015759.D
 Acq: 02 Aug 2021 10:21




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Tgt Ion:	228	Resp:	86652
Ion Ratio	Lower	Upper	
228	100		
226	28.9	23.3	34.9
229	19.8	15.7	23.5



Bis(2-ethylhexyl)phthalate
 Concen: 0.094 ng
 RT: 21.302 min Scan# 2008
 Delta R.T. -0.000 min
 Lab File: BN015759.D
 Acq: 02 Aug 2021 10:21


Tgt Ion:	149	Resp:	20645
Ion Ratio	Lower	Upper	
149	100		
167	26.8	21.2	31.8
279	4.1	3.2	4.8





Abundance Scan 2494 (23.714 min): BN015760.D\data.ms (-2435) (-)

Perylene-d12
Concen: 0.400 ng
RT: 23.710 min Scan# 2493
Delta R.T. -0.004 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2



Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.3	19.5	29.3
265	24.0	19.2	28.8






Abundance Scan 2493 (23.710 min): BN015759.D\data.ms (-2450) (-)




Abundance Scan 3195 (26.113 min): BN015760.D\data.ms (-3126) (-)

Indeno(1,2,3-cd)pyrene
Concen: 0.179 ng
RT: 26.113 min Scan# 3195
Delta R.T. -0.000 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21

Tgt Ion	Ratio	Lower	Upper
276	100		
138	25.5	19.6	29.4
277	25.0	20.0	30.0




Abundance Scan 3195 (26.113 min): BN015759.D\data.ms (-3166) (-)






Abundance Scan 2286 (23.002 min): BN015760.D\data.ms (-2287) (-)

Benzo(b)fluoranthene
Concen: 0.208 ng
RT: 23.002 min Scan# 2286
Delta R.T. -0.000 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2







Tgt Ion:	252	Resp:	82409
Ion Ratio	Lower	Upper	
252	100		
253	22.9	17.7	26.5
125	10.8	8.2	12.4



Abundance Scan 2300 (23.050 min): BN015760.D\data.ms (-2298) (-)

Benzo(k)fluoranthene
Concen: 0.204 ng
RT: 23.050 min Scan# 2300
Delta R.T. -0.000 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21


Tgt Ion:	252	Resp:	87162
Ion Ratio	Lower	Upper	
252	100		
253	23.0	17.8	26.6
125	11.7	8.3	12.5





Abundance Scan 2463 (23.608 min): BN015760.D\data.ms (-2405) (-)

Benzo(a)pyrene
Concen: 0.171 ng
RT: 23.608 min Scan# 2463
Delta R.T. -0.000 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2



Tgt Ion:	252	Resp:	61804
Ion Ratio	Lower	Upper	
252	100		
253	23.4	17.8	26.8
125	12.0	9.0	13.6




Abundance Scan 2463 (23.608 min): BN015759.D\data.ms (-2405) (-)





Abundance Scan 3200 (26.130 min): BN015760.D\data.ms (-3177) (-)


Dibenzo(a,h)anthracene
Concen: 0.177 ng
RT: 26.130 min Scan# 3200
Delta R.T. -0.000 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21




Tgt Ion:	278	Resp:	66336
Ion Ratio	Lower	Upper	
278	100		
139	19.6	13.1	19.7
279	24.4	19.1	28.7



Abundance Scan 3200 (26.130 min): BN015759.D\data.ms (-3177) (-)




Abundance Scan 3409 (26.846 min): BN015760.D\data.ms (-3374) (-)



Benzo(g,h,i)perylene
Concen: 0.183 ng
RT: 26.842 min Scan# 3408
Delta R.T. -0.004 min
Lab File: BN015759.D
Acq: 02 Aug 2021 10:21

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2

Abundance Scan 3408 (26.842 min): BN015759.D\data.ms



Tgt Ion:	276	Resp:	70796
Ion	Ratio	Lower	Upper
276	100		
277	23.8	19.0	28.6
138	20.9	16.6	24.8

Abundance Scan 3408 (26.842 min): BN015759.D\data.ms (-3351) (-)

