

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080221\
 Data File : BN015763.D
 Acq On : 02 Aug 2021 12:47
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Aug 02 13:22:40 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 02 11:44:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.790	152	40390	0.400	ng	0.00
7) Naphthalene-d8	10.571	136	166088	0.400	ng	0.00
13) Acenaphthene-d10	14.421	164	88835	0.400	ng	0.01
19) Phenanthrene-d10	17.163	188	161816	0.400	ng	0.00
29) Chrysene-d12	21.376	240	157300	0.400	ng	# 0.01
35) Perylene-d12	23.714	264	150768	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.383	112	317153	2.924	ng	0.00
5) Phenol-d6	6.951	99	388950	2.963	ng	0.00
8) Nitrobenzene-d5	8.936	82	377015	3.613	ng	0.01
11) 2-Methylnaphthalene-d10	12.163	152	806456	3.224	ng	0.00
14) 2,4,6-Tribromophenol	15.905	330	135067	3.518	ng	0.00
15) 2-Fluorobiphenyl	13.038	172	1053861	3.099	ng	0.00
27) Fluoranthene-d10	19.207	212	1439946	3.377	ng	0.00
31) Terphenyl-d14	19.812	244	1192134	3.114	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.345	88	45463	3.274	ng	# 76
3) n-Nitrosodimethylamine	3.687	42	98394	3.665	ng	# 97
6) bis(2-Chloroethyl)ether	7.218	93	334152	3.111	ng	99
9) Naphthalene	10.622	128	1366162	3.158	ng	100
10) Hexachlorobutadiene	10.914	225	239324	3.144	ng	# 100
12) 2-Methylnaphthalene	12.234	142	878680	3.127	ng	99
16) Acenaphthylene	14.129	152	1308003	3.352	ng	100
17) Acenaphthene	14.472	154	829095	3.262	ng	96
18) Fluorene	15.461	166	1003265	3.240	ng	99
20) 4,6-Dinitro-2-methylph...	15.550	198	55178	4.425	ng	# 81
21) 4-Bromophenyl-phenylether	16.360	248	319727	3.384	ng	100
22) Hexachlorobenzene	16.482	284	351571	3.368	ng	99
23) Atrazine	16.640	200	259172	3.592	ng	# 88
24) Pentachlorophenol	16.823	266	142602	3.639	ng	99
25) Phenanthrene	17.212	178	1538491	3.320	ng	100
26) Anthracene	17.297	178	1416876	3.383	ng	100
28) Fluoranthene	19.237	202	1629472	3.289	ng	99
30) Pyrene	19.599	202	1705423	3.246	ng	100
32) Benzo(a)anthracene	21.355	228	1691655	3.378	ng	99
33) Chrysene	21.408	228	1619977	3.179	ng	99
34) Bis(2-ethylhexyl)phtha...	21.302	149	904684	3.397	ng	99
36) Indeno(1,2,3-cd)pyrene	26.120	276	1833446	3.384	ng	100
37) Benzo(b)fluoranthene	23.005	252	1721187	3.551	ng	99
38) Benzo(k)fluoranthene	23.057	252	1697064	3.248	ng	100
39) Benzo(a)pyrene	23.611	252	1521423	3.436	ng	99
40) Dibenzo(a,h)anthracene	26.141	278	1542235	3.368	ng	100
41) Benzo(g,h,i)perylene	26.852	276	1563454	3.309	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080221\
Data File : BN015763.D
Acq On : 02 Aug 2021 12:47
Operator : CG/JU
Sample : SSTDIC3.2
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Quant Time: Aug 02 13:22:40 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 02 11:44:42 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.790 min Scan# 511
Delta R.T. -0.000 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	152	Resp:	40390
Ion	Ratio	Lower	Upper
152	100		
150	157.6	123.5	185.3
115	58.1	46.8	70.2



1,4-Dioxane
Concen: 3.274 ng
RT: 3.345 min Scan# 29
Delta R.T. -0.010 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47

Tgt Ion:	88	Resp:	45463
Ion	Ratio	Lower	Upper
88	100		
43	19.0	23.1	34.7#
58	55.2	62.2	93.4#









#3
 n-Nitrosodimethylamine
 Concen: 3.665 ng
 RT: 3.687 min Scan# 66
 Delta R.T. -0.028 min
 Lab File: BN015763.D
 Acq: 02 Aug 2021 12:47




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:	42	Resp:	98394
Ion	Ratio	Lower	Upper
42	100		
74	146.6	114.4	171.6
44	6.7	3.6	5.4#



#4
 2-Fluorophenol
 Concen: 2.924 ng
 RT: 5.383 min Scan# 250
 Delta R.T. -0.000 min
 Lab File: BN015763.D
 Acq: 02 Aug 2021 12:47

Tgt Ion:	112	Resp:	317153
Ion	Ratio	Lower	Upper
112	100		
64	52.5	42.2	63.4
63	27.5	22.2	33.2









Phenol-d6
 Concen: 2.963 ng
 RT: 6.951 min Scan# 420
 Delta R.T. -0.000 min
 Lab File: BN015763.D
 Acq: 02 Aug 2021 12:47




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:	99	Resp:	388950
Ion	Ratio	Lower	Upper
99	100		
42	16.2	13.2	19.8
71	29.6	23.8	35.8



bis(2-Chloroethyl)ether
 Concen: 3.111 ng
 RT: 7.218 min Scan# 449
 Delta R.T. -0.000 min
 Lab File: BN015763.D
 Acq: 02 Aug 2021 12:47

Tgt Ion:	93	Resp:	334152
Ion	Ratio	Lower	Upper
93	100		
63	72.6	58.8	88.2
95	33.0	26.6	39.8









Naphthalene-d8
Concen: 0.400 ng
RT: 10.571 min Scan# 746
Delta R.T. -0.000 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	166088
Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.7	13.1
54	5.0	3.9	5.9
68	4.1	3.3	4.9



Nitrobenzene-d5
Concen: 3.613 ng
RT: 8.936 min Scan# 617
Delta R.T. 0.012 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47

Tgt Ion:	82	Resp:	377015
Ion	Ratio	Lower	Upper
82	100		
128	50.4	26.2	39.4#
54	43.2	45.1	67.7#









Naphthalene
 Concen: 3.158 ng
 RT: 10.622 min Scan# 750
 Delta R.T. -0.000 min
 Lab File: BN015763.D
 Acq: 02 Aug 2021 12:47




Instrument :
 BNA_N
 ClientSampled :
 SSTDICC3.2

Tgt Ion:	128	Resp:	1366162
Ion	Ratio	Lower	Upper
128	100		
129	11.1	8.8	13.2
127	12.5	10.2	15.2




Hexachlorobutadiene
 Concen: 3.144 ng
 RT: 10.914 min Scan# 773
 Delta R.T. -0.000 min
 Lab File: BN015763.D
 Acq: 02 Aug 2021 12:47

Tgt Ion:	225	Resp:	239324
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.1	51.5	77.3






Abundance Scan 967 (12.163 min): BN015760.D\data.ms (-953) (-)




2-Methylnaphthalene-d10
Concen: 3.224 ng
RT: 12.163 min Scan# 967
Delta R.T. -0.000 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:152 Resp: 806456
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.4







Abundance Scan 983 (12.234 min): BN015760.D\data.ms (-973) (-)



2-Methylnaphthalene
Concen: 3.127 ng
RT: 12.234 min Scan# 983
Delta R.T. -0.000 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47

Tgt Ion:142 Resp: 878680
Ion Ratio Lower Upper
142 100
141 87.1 70.5 105.7
115 33.2 27.4 41.0









Scan 1194 (14.408 min): BN015760.D\data.ms (-1178) (-)
 Acenaphthene-d10
 Concen: 0.400 ng
 RT: 14.421 min Scan# 1195
 Delta R.T. 0.012 min
 Lab File: BN015763.D
 Acq: 02 Aug 2021 12:47




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:	164	Resp:	88835
Ion	Ratio	Lower	Upper
164	100		
162	98.4	84.2	126.2
160	42.9	39.8	59.6



Scan 1312 (15.906 min): BN015760.D\data.ms (-1304) (-)
 2,4,6-Tribromophenol
 Concen: 3.518 ng
 RT: 15.905 min Scan# 1312
 Delta R.T. -0.000 min
 Lab File: BN015763.D
 Acq: 02 Aug 2021 12:47

Tgt Ion:	330	Resp:	135067
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	34.5	28.2	42.2









2-Fluorobiphenyl
 Concen: 3.099 ng
 RT: 13.038 min Scan# 1086
 Delta R.T. -0.000 min
 Lab File: BN015763.D
 Acq: 02 Aug 2021 12:47




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:	172	Resp:	1053861
Ion	Ratio	Lower	Upper
172	100		
171	34.5	27.8	41.8
170	23.0	18.7	28.1




Acenaphthylene
 Concen: 3.352 ng
 RT: 14.129 min Scan# 1172
 Delta R.T. -0.000 min
 Lab File: BN015763.D
 Acq: 02 Aug 2021 12:47

Tgt Ion:	152	Resp:	1308003
Ion	Ratio	Lower	Upper
152	100		
151	19.3	15.7	23.5
153	13.1	10.5	15.7






Abundance Scan 1199 (14.472 min): BN015760.D\data.ms (-1191) (-)




Acenaphthene
Concen: 3.262 ng
RT: 14.472 min Scan# 1199
Delta R.T. -0.000 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:	154	Resp:	829095
Ion	Ratio	Lower	Upper
154	100		
153	109.2	90.5	135.7
152	52.6	44.4	66.6






Abundance Scan 1277 (15.461 min): BN015760.D\data.ms (-1271) (-)



Fluorene
Concen: 3.240 ng
RT: 15.461 min Scan# 1277
Delta R.T. -0.000 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47


Tgt Ion:	166	Resp:	1003265
Ion	Ratio	Lower	Upper
166	100		
165	96.6	78.2	117.2
167	13.5	10.7	16.1





Abundance Scan 1415 (17.164 min): BN015760.D\data.ms (-1419) (-)

Phenanthrene-d10
Concen: 0.400 ng
RT: 17.163 min Scan# 1415
Delta R.T. -0.000 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47


Instrument :
BNA_N
ClientSampled :
SSTDICC3.2



Tgt Ion:	188	Resp:	161816
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	9.6	7.7	11.5





Abundance Scan 1415 (17.163 min): BN015763.D\data.ms (-1399) (-)




Abundance Scan 1283 (15.538 min): BN015760.D\data.ms (-1278) (-)

4,6-Dinitro-2-methylphenol
Concen: 4.425 ng
RT: 15.550 min Scan# 1284
Delta R.T. 0.012 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47

Tgt Ion:	198	Resp:	55178
Ion	Ratio	Lower	Upper
198	100		
182	6.1	11.1	16.7#
77	0.0	0.0	0.0




Abundance Scan 1284 (15.550 min): BN015763.D\data.ms (-1275) (-)






Abundance Scan 1349 (16.360 min): BN015760.D\data.ms (-1349) (-)

4-Bromophenyl-phenylether
Concen: 3.384 ng
RT: 16.360 min Scan# 1349
Delta R.T. -0.000 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Tgt Ion:	248	Resp:	319727
Ion	Ratio	Lower	Upper
248	100		
250	95.3	75.8	113.6
141	65.7	52.8	79.2






Abundance Scan 1358 (16.470 min): BN015760.D\data.ms (-1358) (-)

Hexachlorobenzene
Concen: 3.368 ng
RT: 16.482 min Scan# 1359
Delta R.T. 0.012 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47




Tgt Ion:	284	Resp:	351571
Ion	Ratio	Lower	Upper
284	100		
142	36.0	29.4	44.2
249	29.4	23.8	35.6






Abundance Scan 1371 (16.628 min): BN015760.D\data.ms (-1322) (-)

Atrazine
Concen: 3.592 ng
RT: 16.640 min Scan# 1372
Delta R.T. 0.012 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2







Tgt Ion	Ratio	Lower	Upper
200	100		
173	20.1	25.0	37.6#
215	49.2	36.3	54.5




Abundance Scan 1387 (16.823 min): BN015760.D\data.ms (-1324) (-)

Pentachlorophenol
Concen: 3.639 ng
RT: 16.823 min Scan# 1387
Delta R.T. -0.000 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47

Tgt Ion	Ratio	Lower	Upper
266	100		
264	62.5	49.7	74.5
268	63.9	51.8	77.6






Abundance Scan 1419 (17.212 min): BN015760.D\data.ms (-1425) (-)




Phenanthrene
Concen: 3.320 ng
RT: 17.212 min Scan# 1419
Delta R.T. -0.000 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:178 Resp: 1538491
Ion Ratio Lower Upper
178 100
176 18.9 15.1 22.7
179 15.5 12.2 18.4






Abundance Scan 1426 (17.297 min): BN015760.D\data.ms (-1426) (-)




Anthracene
Concen: 3.383 ng
RT: 17.297 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47

Tgt Ion:178 Resp: 1416876
Ion Ratio Lower Upper
178 100
176 18.3 14.6 22.0
179 15.5 12.2 18.4





Abundance Scan 1725 (19.207 min): BN015760.D\data.ms (-1725) (-)




Fluoranthene-d10
Concen: 3.377 ng
RT: 19.207 min Scan# 1725
Delta R.T. -0.000 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:212 Resp: 1439946
Ion Ratio Lower Upper
212 100
106 13.8 10.8 16.2
104 7.6 6.0 9.0



Abundance Scan 1725 (19.207 min): BN015763.D\data.ms (-1705) (-)





Abundance Scan 1731 (19.237 min): BN015760.D\data.ms (-1723) (-)




Fluoranthene
Concen: 3.289 ng
RT: 19.237 min Scan# 1731
Delta R.T. -0.000 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47


Tgt Ion:202 Resp: 1629472
Ion Ratio Lower Upper
202 100
101 11.5 9.0 13.6
203 17.6 13.8 20.8



Abundance Scan 1731 (19.237 min): BN015763.D\data.ms (-1691) (-)





Abundance Scan 2014 (21.366 min): BN015760.D\data.ms (-2029) (-)




Chrysene-d12
Concen: 0.400 ng
RT: 21.376 min Scan# 2015
Delta R.T. 0.010 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion	Ratio	Lower	Upper
240	100		
120	7.0	11.4	17.0#
236	25.7	21.2	31.8



Abundance Scan 2015 (21.376 min): BN015763.D\data.ms (-1995) (-)





Abundance Scan 1804 (19.599 min): BN015760.D\data.ms (-1734) (-)




Pyrene
Concen: 3.246 ng
RT: 19.599 min Scan# 1804
Delta R.T. -0.000 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47

Tgt Ion	Ratio	Lower	Upper
202	100		
200	20.6	16.5	24.7
203	17.7	13.9	20.9




Abundance Scan 1804 (19.599 min): BN015763.D\data.ms (-1764) (-)



Abundance Scan 1847 (19.813 min): BN015760.D\data.ms (-1838) (-)




Terphenyl-d14
Concen: 3.114 ng
RT: 19.812 min Scan# 1847
Delta R.T. -0.000 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Tgt Ion:244 Resp: 1192134

Ion	Ratio	Lower	Upper
244	100		
212	7.5	5.9	8.9
122	12.7	9.7	14.5






Abundance Scan 2013 (21.355 min): BN015760.D\data.ms (-2009) (-)


Benzo(a)anthracene
Concen: 3.378 ng
RT: 21.355 min Scan# 2013
Delta R.T. -0.000 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47



Tgt Ion:228 Resp: 1691655

Ion	Ratio	Lower	Upper
228	100		
226	27.4	21.2	31.8
229	19.6	15.6	23.4









Scan 2018 (21.408 min): BN015760.D\data.ms (-2009) (-)
 Chrysene
 Concen: 3.179 ng
 RT: 21.408 min Scan# 2018
 Delta R.T. -0.000 min
 Lab File: BN015763.D
 Acq: 02 Aug 2021 12:47




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:	228	Resp:	1619977
Ion	Ratio	Lower	Upper
228	100		
226	29.7	23.3	34.9
229	19.8	15.7	23.5



Scan 2008 (21.302 min): BN015760.D\data.ms (-2009) (-)
 Bis(2-ethylhexyl)phthalate
 Concen: 3.397 ng
 RT: 21.302 min Scan# 2008
 Delta R.T. -0.000 min
 Lab File: BN015763.D
 Acq: 02 Aug 2021 12:47




Tgt Ion:	149	Resp:	904684
Ion	Ratio	Lower	Upper
149	100		
167	27.2	21.2	31.8
279	3.9	3.2	4.8




Abundance Scan 2494 (23.714 min): BN015760.D\data.ms (-2450) (-)

Perylene-d12
Concen: 0.400 ng
RT: 23.714 min Scan# 2494
Delta R.T. -0.000 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47

Instrument :
BNA_N
ClientSampleId :
SSTDIC3.2






Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.6	19.5	29.3
265	23.6	19.2	28.8




Abundance Scan 3195 (26.113 min): BN015760.D\data.ms (-3126) (-)

Indeno(1,2,3-cd)pyrene
Concen: 3.384 ng
RT: 26.120 min Scan# 3197
Delta R.T. 0.007 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47




Tgt Ion	Ratio	Lower	Upper
276	100		
138	24.6	19.6	29.4
277	25.1	20.0	30.0






Abundance Scan 2286 (23.002 min): BN015760.D\data.ms (-2287) (-)

Benzo(b)fluoranthene
Concen: 3.551 ng
RT: 23.005 min Scan# 2287
Delta R.T. 0.003 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.7	26.5
125	10.0	8.2	12.4






Abundance Scan 2300 (23.050 min): BN015760.D\data.ms (-2288) (-)

Benzo(k)fluoranthene
Concen: 3.248 ng
RT: 23.057 min Scan# 2302
Delta R.T. 0.007 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47




Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.8	26.6
125	10.0	8.3	12.5





Abundance Scan 2463 (23.608 min): BN015760.D\data.ms (-2405) (-)

Benzo(a)pyrene
Concen: 3.436 ng
RT: 23.611 min Scan# 2464
Delta R.T. 0.003 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47


Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2



Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.8	26.8
125	11.0	9.0	13.6





Abundance Scan 2464 (23.611 min): BN015763.D\data.ms (-2405) (-)




Abundance Scan 3200 (26.130 min): BN015760.D\data.ms (-3177) (-)

Dibenzo(a,h)anthracene
Concen: 3.368 ng
RT: 26.141 min Scan# 3203
Delta R.T. 0.010 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47


Tgt Ion	Ratio	Lower	Upper
278	100		
139	16.1	13.1	19.7
279	23.8	19.1	28.7



Abundance Scan 3203 (26.141 min): BN015763.D\data.ms (-3177) (-)




Abundance Scan 3409 (26.846 min): BN015760.D\data.ms (-3374) (-)



Benzo(g,h,i)perylene
Concen: 3.309 ng
RT: 26.852 min Scan# 3411
Delta R.T. 0.007 min
Lab File: BN015763.D
Acq: 02 Aug 2021 12:47

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Abundance Scan 3411 (26.852 min): BN015763.D\data.ms



Tgt Ion:	276	Resp:	1563454
Ion	Ratio	Lower	Upper
276	100		
277	23.9	19.0	28.6
138	20.8	16.6	24.8

Abundance Scan 3411 (26.852 min): BN015763.D\data.ms (-3351) (-)

