

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080221\
 Data File : BN015765.D
 Acq On : 02 Aug 2021 14:19
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN080221

Quant Time: Aug 02 15:01:34 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 02 14:09:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.781	152	37884	0.400	ng	0.00
7) Naphthalene-d8	10.571	136	166278	0.400	ng	0.00
13) Acenaphthene-d10	14.421	164	86782	0.400	ng	0.01
19) Phenanthrene-d10	17.164	188	160934	0.400	ng	0.00
29) Chrysene-d12	21.376	240	148108	0.400	ng	# 0.01
35) Perylene-d12	23.721	264	135044	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.383	112	37312	0.403	ng	0.00
5) Phenol-d6	6.951	99	44412	0.397	ng	0.00
8) Nitrobenzene-d5	8.924	82	41359	0.387	ng	0.00
11) 2-Methylnaphthalene-d10	12.163	152	100606	0.403	ng	0.00
14) 2,4,6-Tribromophenol	15.905	330	10750	0.342	ng	0.00
15) 2-Fluorobiphenyl	13.038	172	133432	0.386	ng	0.00
27) Fluoranthene-d10	19.207	212	164444	0.393	ng	0.00
31) Terphenyl-d14	19.817	244	141074	0.385	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.345	88	4969	0.405	ng	# 83
3) n-Nitrosodimethylamine	3.714	42	9260	0.367	ng	# 96
6) bis(2-Chloroethyl)ether	7.218	93	41270	0.406	ng	100
9) Naphthalene	10.622	128	173753	0.399	ng	100
10) Hexachlorobutadiene	10.914	225	30496	0.401	ng	# 100
12) 2-Methylnaphthalene	12.234	142	113353	0.404	ng	100
16) Acenaphthylene	14.129	152	134831	0.377	ng	100
17) Acenaphthene	14.484	154	96673	0.387	ng	100
18) Fluorene	15.474	166	116584	0.390	ng	99
20) 4,6-Dinitro-2-methylph...	15.550	198	2776	0.360	ng	99
21) 4-Bromophenyl-phenylether	16.360	248	34922	0.369	ng	92
22) Hexachlorobenzene	16.482	284	42514	0.391	ng	99
23) Atrazine	16.640	200	26736	0.460	ng	# 89
24) Pentachlorophenol	16.823	266	11424	0.340	ng	99
25) Phenanthrene	17.212	178	185588	0.388	ng	100
26) Anthracene	17.297	178	149402	0.371	ng	100
28) Fluoranthene	19.237	202	197275	0.396	ng	100
30) Pyrene	19.604	202	192238	0.389	ng	100
32) Benzo(a)anthracene	21.355	228	171909	0.380	ng	98
33) Chrysene	21.408	228	192510	0.393	ng	98
34) Bis(2-ethylhexyl)phtha...	21.302	149	51263	0.429	ng	100
36) Indeno(1,2,3-cd)pyrene	26.123	276	180280	0.377	ng	100
37) Benzo(b)fluoranthene	23.012	252	184952	0.394	ng	100
38) Benzo(k)fluoranthene	23.057	252	193328	0.398	ng	100
39) Benzo(a)pyrene	23.614	252	149905	0.384	ng	100
40) Dibenzo(a,h)anthracene	26.144	278	151914	0.379	ng	100
41) Benzo(g,h,i)perylene	26.853	276	158253	0.379	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080221\
Data File : BN015765.D
Acq On : 02 Aug 2021 14:19
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN080221

Quant Time: Aug 02 15:01:34 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 02 14:09:59 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.781 min Scan# 510
Delta R.T. -0.009 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19




Instrument :
BNA_N
ClientSampleId :
ICVBN080221


Tgt Ion:152 Resp: 37884
Ion Ratio Lower Upper
152 100
150 158.1 123.5 185.3
115 60.6 46.8 70.2



#2
1,4-Dioxane
Concen: 0.405 ng
RT: 3.345 min Scan# 29
Delta R.T. -0.010 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Tgt Ion: 88 Resp: 4969
Ion Ratio Lower Upper
88 100
43 22.2 23.1 34.7#
58 61.6 62.2 93.4#









#3
n-Nitrosodimethylamine
Concen: 0.367 ng
RT: 3.714 min Scan# 69
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19




Instrument :
BNA_N
ClientSampleId :
ICVBN080221


Tgt Ion: 42 Resp: 9260
Ion Ratio Lower Upper
42 100
74 148.0 114.4 171.6
44 6.9 3.6 5.4#



#4
2-Fluorophenol
Concen: 0.403 ng
RT: 5.383 min Scan# 250
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Tgt Ion: 112 Resp: 37312
Ion Ratio Lower Upper
112 100
64 52.7 42.2 63.4
63 27.4 22.2 33.2









#5
Phenol-d6
Concen: 0.397 ng
RT: 6.951 min Scan# 420
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Instrument :
BNA_N
ClientSampleId :
ICVBN080221

Tgt Ion: 99 Resp: 44412




Ion	Ratio	Lower	Upper
99	100		
42	16.1	13.2	19.8
71	29.4	23.8	35.8




#6
bis(2-Chloroethyl)ether
Concen: 0.406 ng
RT: 7.218 min Scan# 449
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Tgt Ion: 93 Resp: 41270

Ion	Ratio	Lower	Upper
93	100		
63	73.2	58.8	88.2
95	33.4	26.6	39.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.571 min Scan# 746
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19




Instrument :
BNA_N
ClientSampleId :
ICVBN080221


Tgt Ion:	136	Resp:	166278
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.7	13.1
54	5.1	3.9	5.9
68	4.2	3.3	4.9



#8
Nitrobenzene-d5
Concen: 0.387 ng
RT: 8.924 min Scan# 616
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Tgt Ion:	82	Resp:	41359
Ion	Ratio	Lower	Upper
82	100		
128	34.9	26.2	39.4
54	54.4	45.1	67.7









#9
Naphthalene
Concen: 0.399 ng
RT: 10.622 min Scan# 750
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19




Instrument :
BNA_N
ClientSampleId :
ICVBN080221


Tgt Ion:128 Resp: 173753
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 12.6 10.2 15.2



#10
Hexachlorobutadiene
Concen: 0.401 ng
RT: 10.914 min Scan# 773
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Tgt Ion:225 Resp: 30496
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.1 51.5 77.3









#11
2-Methylnaphthalene-d10
Concen: 0.403 ng
RT: 12.163 min Scan# 967
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19




Instrument :
BNA_N
ClientSampleId :
ICVBN080221


Tgt Ion:152 Resp: 100606
Ion Ratio Lower Upper
152 100
151 20.4 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.404 ng
RT: 12.234 min Scan# 983
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Tgt Ion:142 Resp: 113353
Ion Ratio Lower Upper
142 100
141 88.3 70.5 105.7
115 34.0 27.4 41.0









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.421 min Scan# 1195
Delta R.T. 0.013 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19




Instrument :
BNA_N
ClientSampleId :
ICVBN080221


Tgt Ion:164 Resp: 86782
Ion Ratio Lower Upper
164 100
162 97.6 84.2 126.2
160 43.1 39.8 59.6



#14
2,4,6-Tribromophenol
Concen: 0.342 ng
RT: 15.905 min Scan# 1312
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Tgt Ion:330 Resp: 10750
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 34.8 28.2 42.2









#15
2-Fluorobiphenyl
Concen: 0.386 ng
RT: 13.038 min Scan# 1086
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19




Instrument :
BNA_N
ClientSampleId :
ICVBN080221


Tgt Ion:172 Resp: 133432
Ion Ratio Lower Upper
172 100
171 35.1 27.8 41.8
170 23.8 18.7 28.1



#16
Acenaphthylene
Concen: 0.377 ng
RT: 14.129 min Scan# 1172
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Tgt Ion:152 Resp: 134831
Ion Ratio Lower Upper
152 100
151 19.5 15.7 23.5
153 13.0 10.5 15.7









#17
Acenaphthene
Concen: 0.387 ng
RT: 14.484 min Scan# 1200
Delta R.T. 0.013 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19




Instrument :
BNA_N
ClientSampleId :
ICVBN080221


Tgt Ion:154 Resp: 96673
Ion Ratio Lower Upper
154 100
153 112.6 90.5 135.7
152 55.4 44.4 66.6



#18
Fluorene
Concen: 0.390 ng
RT: 15.474 min Scan# 1278
Delta R.T. 0.013 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Tgt Ion:166 Resp: 116584
Ion Ratio Lower Upper
166 100
165 97.0 78.2 117.2
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.164 min Scan# 1415
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19




Instrument :
BNA_N
ClientSampleId :
ICVBN080221


Tgt Ion:188 Resp: 160934
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.8 7.7 11.5



#20
4,6-Dinitro-2-methylphenol
Concen: 0.360 ng
RT: 15.550 min Scan# 1284
Delta R.T. 0.012 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Tgt Ion:198 Resp: 2776
Ion Ratio Lower Upper
198 100
182 13.4 11.1 16.7
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.369 ng
RT: 16.360 min Scan# 1349
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19




Instrument :
BNA_N
ClientSampleId :
ICVBN080221


Tgt Ion:248 Resp: 34922
Ion Ratio Lower Upper
248 100
250 92.0 75.8 113.6
141 78.2 52.8 79.2



#22
Hexachlorobenzene
Concen: 0.391 ng
RT: 16.482 min Scan# 1359
Delta R.T. 0.012 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19


Tgt Ion:284 Resp: 42514
Ion Ratio Lower Upper
284 100
142 35.5 29.4 44.2
249 29.6 23.8 35.6








#23
Atrazine
Concen: 0.460 ng
RT: 16.640 min Scan# 1372
Delta R.T. 0.012 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Instrument :
BNA_N
ClientSampleId :
ICVBN080221







Tgt Ion:200 Resp: 26736
Ion Ratio Lower Upper
200 100
173 20.7 25.0 37.6#
215 49.2 36.3 54.5



#24
Pentachlorophenol
Concen: 0.340 ng
RT: 16.823 min Scan# 1387
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Tgt Ion:266 Resp: 11424
Ion Ratio Lower Upper
266 100
264 62.8 49.7 74.5
268 64.2 51.8 77.6









#25
Phenanthrene
Concen: 0.388 ng
RT: 17.212 min Scan# 1419
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19




Instrument :
BNA_N
ClientSampleId :
ICVBN080221


Tgt Ion:178 Resp: 185588
Ion Ratio Lower Upper
178 100
176 19.0 15.1 22.7
179 15.3 12.2 18.4



#26
Anthracene
Concen: 0.371 ng
RT: 17.297 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Tgt Ion:178 Resp: 149402
Ion Ratio Lower Upper
178 100
176 18.4 14.6 22.0
179 15.4 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.393 ng
RT: 19.207 min Scan# 1725
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19




Instrument :
BNA_N
ClientSampleId :
ICVBN080221


Tgt Ion:212 Resp: 164444
Ion Ratio Lower Upper
212 100
106 13.6 10.8 16.2
104 7.5 6.0 9.0



#28
Fluoranthene
Concen: 0.396 ng
RT: 19.237 min Scan# 1731
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Tgt Ion:202 Resp: 197275
Ion Ratio Lower Upper
202 100
101 11.3 9.0 13.6
203 17.3 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.376 min Scan# 2015
Delta R.T. 0.010 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19




Instrument :
BNA_N
ClientSampleId :
ICVBN080221


Tgt Ion:240 Resp: 148108
Ion Ratio Lower Upper
240 100
120 8.5 11.4 17.0#
236 26.0 21.2 31.8



#30
Pyrene
Concen: 0.389 ng
RT: 19.604 min Scan# 1805
Delta R.T. 0.005 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Tgt Ion:202 Resp: 192238
Ion Ratio Lower Upper
202 100
200 20.6 16.5 24.7
203 17.4 13.9 20.9









#31
Terphenyl-d14
Concen: 0.385 ng
RT: 19.817 min Scan# 1848
Delta R.T. 0.005 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19




Instrument :
BNA_N
ClientSampleId :
ICVBN080221


Tgt Ion:244 Resp: 141074
Ion Ratio Lower Upper
244 100
212 7.3 5.9 8.9
122 11.4 9.7 14.5



#32
Benzo(a)anthracene
Concen: 0.380 ng
RT: 21.355 min Scan# 2013
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Tgt Ion:228 Resp: 171909
Ion Ratio Lower Upper
228 100
226 27.9 21.2 31.8
229 19.2 15.6 23.4









#33
Chrysene
Concen: 0.393 ng
RT: 21.408 min Scan# 2018
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Instrument :
BNA_N
ClientSampleId :
ICVBN080221

Tgt Ion:228 Resp: 192510




Ion	Ratio	Lower	Upper
228	100		
226	30.3	23.3	34.9
229	19.1	15.7	23.5




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.429 ng
RT: 21.302 min Scan# 2008
Delta R.T. -0.000 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Tgt Ion:149 Resp: 51263


Ion	Ratio	Lower	Upper
149	100		
167	26.7	21.2	31.8
279	4.1	3.2	4.8








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.721 min Scan# 2496
Delta R.T. 0.007 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Instrument :
BNA_N
ClientSampleId :
ICVBN080221







Tgt Ion:264 Resp: 135044
Ion Ratio Lower Upper
264 100
260 24.3 19.5 29.3
265 24.0 19.2 28.8



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.377 ng
RT: 26.123 min Scan# 3198
Delta R.T. 0.010 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Tgt Ion:276 Resp: 180280
Ion Ratio Lower Upper
276 100
138 24.6 19.6 29.4
277 25.0 20.0 30.0









#37
Benzo(b)fluoranthene
Concen: 0.394 ng
RT: 23.012 min Scan# 2289
Delta R.T. 0.010 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19




Instrument :
BNA_N
ClientSampleId :
ICVBN080221


Tgt Ion:252 Resp: 184952
Ion Ratio Lower Upper
252 100
253 22.2 17.7 26.5
125 10.2 8.2 12.4



#38
Benzo(k)fluoranthene
Concen: 0.398 ng
RT: 23.057 min Scan# 2302
Delta R.T. 0.007 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Tgt Ion:252 Resp: 193328
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.6
125 10.6 8.3 12.5









#39
Benzo(a)pyrene
Concen: 0.384 ng
RT: 23.614 min Scan# 2465
Delta R.T. 0.007 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19




Instrument :
BNA_N
ClientSampleId :
ICVBN080221


Tgt Ion:252 Resp: 149905
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.8
125 11.3 9.0 13.6



#40
Dibenzo(a,h)anthracene
Concen: 0.379 ng
RT: 26.144 min Scan# 3204
Delta R.T. 0.014 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Tgt Ion:278 Resp: 151914
Ion Ratio Lower Upper
278 100
139 16.6 13.1 19.7
279 24.0 19.1 28.7





#41
Benzo(g,h,i)perylene
Concen: 0.379 ng
RT: 26.853 min Scan# 3411
Delta R.T. 0.007 min
Lab File: BN015765.D
Acq: 02 Aug 2021 14:19

Instrument :
BNA_N
ClientSampleId :
ICVBN080221

Tgt Ion:276 Resp: 158253
Ion Ratio Lower Upper
276 100
277 23.8 19.0 28.6
138 21.0 16.6 24.8

