

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN080222\  
 Data File : BN020953.D  
 Acq On : 02 Aug 2022 21:23  
 Operator : CG/JU  
 Sample : N3952-01  
 Misc :  
 ALS Vial : 20 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 COCA0

Quant Time: Aug 02 22:34:32 2022  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN080122.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Aug 02 04:19:04 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|--------|------|----------|-------|-------|----------|
| -----                       |        |      |          |       |       |          |
| Internal Standards          |        |      |          |       |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.805  | 152  | 2607     | 0.400 | ng/u1 | 0.00     |
| 4) Naphthalene-d8           | 10.601 | 136  | 10777    | 0.400 | ng/u1 | 0.00     |
| 9) Acenaphthene-d10         | 14.460 | 164  | 6364     | 0.400 | ng/u1 | 0.00     |
| 13) Phenanthrene-d10        | 17.208 | 188  | 12964    | 0.400 | ng/u1 | 0.00     |
| 17) Chrysene-d12            | 21.418 | 240  | 10646    | 0.400 | ng/u1 | 0.00     |
| 23) Perylene-d12            | 23.751 | 264  | 9592     | 0.400 | ng/u1 | -0.01    |
| System Monitoring Compounds |        |      |          |       |       |          |
| 3) 1,4-Dioxane-d8           | 3.207  | 96   | 18277    | 5.749 | ng/u1 | 0.00     |
| 6) 2-Methylnaphthalene-d10  | 12.206 | 152  | 5582     | 0.351 | ng/u1 | 0.00     |
| 18) Fluoranthene-d10        | 19.248 | 212  | 14187    | 0.435 | ng/u1 | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN080222\  
Data File : BN020953.D  
Acq On : 02 Aug 2022 21:23  
Operator : CG/JU  
Sample : N3952-01  
Misc :  
ALS Vial : 20 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
C0CA0

Quant Time: Aug 02 22:34:32 2022  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN080122.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Aug 02 04:19:04 2022  
Response via : Initial Calibration

