

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080222\
 Data File : BN020945.D
 Acq On : 02 Aug 2022 16:21
 Operator : CG/JU
 Sample : N3943-11
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0C92

Quant Time: Aug 02 22:33:57 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 04:19:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.805	152	3455	0.400	ng/ul	0.00
4) Naphthalene-d8	10.602	136	8993	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.457	164	6971	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.210	188	14292	0.400	ng/ul	0.00
17) Chrysene-d12	21.430	240	11358	0.400	ng/ul	0.00
23) Perylene-d12	23.771	264	11725	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.207	96	18460	4.381	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.208	152	5495	0.414	ng/ul	0.00
18) Fluoranthene-d10	19.254	212	14821	0.426	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080222\
Data File : BN020945.D
Acq On : 02 Aug 2022 16:21
Operator : CG/JU
Sample : N3943-11
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0C92

Quant Time: Aug 02 22:33:57 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 02 04:19:04 2022
Response via : Initial Calibration

