

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080222\
 Data File : BN020948.D
 Acq On : 02 Aug 2022 18:15
 Operator : CG/JU
 Sample : PB146569BL
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK569

Quant Time: Aug 02 22:34:11 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 04:19:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.805	152	2842	0.400	ng/ul	0.00
4) Naphthalene-d8	10.602	136	8916	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.457	164	5234	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.210	188	10362	0.400	ng/ul	0.00
17) Chrysene-d12	21.427	240	7539	0.400	ng/ul	0.00
23) Perylene-d12	23.765	264	8482	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.207	96	18745	5.408	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.213	152	3743	0.285	ng/ul	0.00
18) Fluoranthene-d10	19.254	212	8283	0.359	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080222\
Data File : BN020948.D
Acq On : 02 Aug 2022 18:15
Operator : CG/JU
Sample : PB146569BL
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK569

Quant Time: Aug 02 22:34:11 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 02 04:19:04 2022
Response via : Initial Calibration

