

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080224\
 Data File : BN033175.D
 Acq On : 01 Aug 2024 20:05
 Operator : MA/JU
 Sample : P3401-09
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E1GF4

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 08/02/2024
 Supervised By :mohammad ahmed 08/03/2024

Quant Time: Aug 01 23:16:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN072524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 30 14:49:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.486	152	2748m	0.400	ng/ul	-0.06
4) Naphthalene-d8	10.234	136	8874	0.400	ng/ul	#-0.08
9) Acenaphthene-d10	14.124	164	5545	0.400	ng/ul	#-0.04
13) Phenanthrene-d10	16.895	188	12239m	0.400	ng/ul	-0.05
17) Chrysene-d12	21.154	240	10373	0.400	ng/ul	0.00
23) Perylene-d12	23.320	264	5842m	0.400	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.048	96	15214m	4.377	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	11.879	152	4195	0.290	ng/ul	-0.05
18) Fluoranthene-d10	18.949	212	11570	0.394	ng/ul	-0.02
Target Compounds						Qvalue
5) Naphthalene	10.284	128	973	0.043	ng/ul#	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080224\
Data File : BN033175.D
Acq On : 01 Aug 2024 20:05
Operator : MA/JU
Sample : P3401-09
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
E1GF4

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/02/2024
Supervised By :mohammad ahmed 08/03/2024

Quant Time: Aug 01 23:16:23 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN072524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 30 14:49:49 2024
Response via : Initial Calibration

