

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080224\
 Data File : BN033189.D
 Acq On : 02 Aug 2024 05:08
 Operator : MA/JU
 Sample : P3265-06
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A4D39

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 08/02/2024
 Supervised By :mohammad ahmed 08/03/2024

Quant Time: Aug 02 05:40:24 2024

Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN072524.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Tue Jul 30 14:49:49 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.470	152	4301m	0.400	ng/ul	-0.08
4) Naphthalene-d8	10.229	136	13064	0.400	ng/ul	#-0.09
9) Acenaphthene-d10	14.115	164	6861	0.400	ng/ul	-0.05
13) Phenanthrene-d10	16.882	188	13845	0.400	ng/ul	-0.06
17) Chrysene-d12	21.107	240	10701m	0.400	ng/ul	-0.04
23) Perylene-d12	23.285	264	14271	0.400	ng/ul	-0.06
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.048	96	10399	1.911	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	11.840	152	2571	0.121	ng/ul	-0.09
18) Fluoranthene-d10	18.926	212	4919	0.162	ng/ul	-0.05
Target Compounds					Qvalue	
5) Naphthalene	10.278	128	23712	0.707	ng/ul#	94
7) 2-Methylnaphthalene	11.917	142	8608	0.379	ng/ul	96
8) 1-Methylnaphthalene	12.137	142	6582	0.268	ng/ul	100
10) Acenaphthylene	13.838	152	47610	1.696	ng/ul	98
11) Acenaphthene	14.180	153	27598	1.307	ng/ul	99
12) Fluorene	15.179	166	29794	1.210	ng/ul	100
14) Pentachlorophenol	16.565	266	271	0.096	ng/ul	94
15) Phenanthrene	16.924	178	645377	17.379	ng/ul	97
16) Anthracene	17.017	178	125652	3.763	ng/ul	98
19) Fluoranthene	18.958	202	1330657	37.766	ng/ul	97
20) Pyrene	19.325	202	1056865	27.971	ng/ul	97
21) Benzo(a)anthracene	21.090	228	624135	17.099	ng/ul	100
22) Chrysene	21.145	228	696951m	16.361	ng/ul	
24) Benzo(b)fluoranthene	22.636	252	1067758m	25.613	ng/ul	
25) Benzo(k)fluoranthene	22.671	252	335272m	6.711	ng/ul	
26) Benzo(a)pyrene	23.185	252	505558m	12.885	ng/ul	
27) Indeno(1,2,3-cd)pyrene	25.457	276	462319	7.749	ng/ul	90
28) Dibenzo(a,h)anthracene	25.450	278	136503m	2.887	ng/ul	
29) Benzo(g,h,i)perylene	26.131	276	75440	1.591	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080224\
Data File : BN033189.D
Acq On : 02 Aug 2024 05:08
Operator : MA/JU
Sample : P3265-06
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A4D39

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/02/2024
Supervised By :mohammad ahmed 08/03/2024

Quant Time: Aug 02 05:40:24 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN072524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 30 14:49:49 2024
Response via : Initial Calibration

