

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN080320\
 Data File : BN011331.D
 Acq On : 03 Aug 2020 11:29
 Operator : CG/JU
 Sample : L3388-20
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AE7

Manual Integrations
 APPROVED

mohammad
 8/5/2020 10:27:11 AM

Quant Time: Aug 03 12:15:15 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN073020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 03 10:06:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.47	152	1486	0.40	ng/ul	0.00
2) Naphthalene-d8	10.21	136	5520	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.10	164	3481	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.85	188	8050m	0.40	ng/ul	0.00
16) Chrysene-d12	21.07	240	7665	0.40	ng/ul	0.00
20) Perylene-d12	23.19	264	8604	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.81	152	3968	0.28	ng/ul	0.00
14) Fluoranthene-d10	18.89	212	10285	0.26	ng/ul	0.00
Target Compounds				Qvalue		
12) Phenanthrene	16.89	178	744	0.028	ng/ul#	82
15) Fluoranthene	18.92	202	1094	0.033	ng/ul	81
17) Pyrene	19.28	202	1017	0.029	ng/ul#	66
19) Chrysene	21.10	228	1201	0.036	ng/ul#	83
21) Benzo(b)fluoranthene	22.57	252	1879m	0.052	ng/ul	
23) Benzo(a)pyrene	23.10	252	682	0.022	ng/ul#	10
24) Indeno(1,2,3-cd)pyrene	25.26	276	659	0.020	ng/ul#	78
26) Benzo(g,h,i)perylene	25.88	276	695	0.023	ng/ul#	58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN080320\
Data File : BN011331.D
Acq On : 03 Aug 2020 11:29
Operator : CG/JU
Sample : L3388-20
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AE7

Manual Integrations
APPROVED

mohammad
8/5/2020 10:27:11 AM

Quant Time: Aug 03 12:15:15 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN073020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 03 10:06:56 2020
Response via : Initial Calibration

