

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080321\
 Data File : BN015768.D
 Acq On : 03 Aug 2021 14:50
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Aug 03 16:17:35 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 02 14:09:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.790	152	37382	0.400	ng	0.00
7) Naphthalene-d8	10.571	136	166014	0.400	ng	0.00
13) Acenaphthene-d10	14.408	164	86212	0.400	ng	0.00
19) Phenanthrene-d10	17.163	188	159310	0.400	ng	0.00
29) Chrysene-d12	21.365	240	144109	0.400	ng	0.00
35) Perylene-d12	23.710	264	134831	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.383	112	35238	0.386	ng	0.00
5) Phenol-d6	6.951	99	42308	0.383	ng	0.00
8) Nitrobenzene-d5	8.924	82	38322	0.359	ng	0.00
11) 2-Methylnaphthalene-d10	12.158	152	99387	0.399	ng	0.00
14) 2,4,6-Tribromophenol	15.905	330	10292	0.329	ng	0.00
15) 2-Fluorobiphenyl	13.038	172	132792	0.386	ng	0.00
27) Fluoranthene-d10	19.207	212	156169	0.377	ng	0.00
31) Terphenyl-d14	19.812	244	137424	0.385	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.355	88	4011	0.331	ng	# 85
3) n-Nitrosodimethylamine	3.714	42	8942	0.359	ng	# 100
6) bis(2-Chloroethyl)ether	7.218	93	41312	0.412	ng	100
9) Naphthalene	10.622	128	172886	0.398	ng	100
10) Hexachlorobutadiene	10.913	225	29867	0.393	ng	# 99
12) 2-Methylnaphthalene	12.234	142	112518	0.402	ng	100
16) Acenaphthylene	14.129	152	128757	0.362	ng	100
17) Acenaphthene	14.472	154	95170	0.384	ng	99
18) Fluorene	15.461	166	114833	0.387	ng	100
20) 4,6-Dinitro-2-methylph...	15.537	198	2909	0.375	ng	98
21) 4-Bromophenyl-phenylether	16.360	248	34997	0.374	ng	98
22) Hexachlorobenzene	16.470	284	41749	0.388	ng	100
23) Atrazine	16.628	200	24426	0.424	ng	98
24) Pentachlorophenol	16.811	266	13368	0.402	ng	99
25) Phenanthrene	17.200	178	182931	0.386	ng	100
26) Anthracene	17.297	178	145695	0.366	ng	100
28) Fluoranthene	19.236	202	191694	0.389	ng	100
30) Pyrene	19.599	202	186496	0.388	ng	100
32) Benzo(a)anthracene	21.355	228	161855	0.367	ng	99
33) Chrysene	21.408	228	188327	0.395	ng	100
34) Bis(2-ethylhexyl)phtha...	21.302	149	47909	0.422	ng	100
36) Indeno(1,2,3-cd)pyrene	26.110	276	176973	0.370	ng	99
37) Benzo(b)fluoranthene	23.002	252	181643	0.388	ng	100
38) Benzo(k)fluoranthene	23.050	252	189641	0.391	ng	99
39) Benzo(a)pyrene	23.608	252	144370	0.371	ng	100
40) Dibenzo(a,h)anthracene	26.130	278	147082	0.368	ng	98
41) Benzo(g,h,i)perylene	26.842	276	155188	0.372	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080321\
Data File : BN015768.D
Acq On : 03 Aug 2021 14:50
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Aug 03 16:17:35 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 02 14:09:59 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.790 min Scan# 511
 Delta R.T. -0.000 min
 Lab File: BN015768.D
 Acq: 03 Aug 2021 14:50

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Tgt Ion:152 Resp: 37382




Ion	Ratio	Lower	Upper
152	100		
150	154.7	123.5	185.3
115	58.2	46.8	70.2




#2
 1,4-Dioxane
 Concen: 0.331 ng
 RT: 3.355 min Scan# 30
 Delta R.T. -0.000 min
 Lab File: BN015768.D
 Acq: 03 Aug 2021 14:50

Tgt Ion: 88 Resp: 4011

Ion	Ratio	Lower	Upper
88	100		
43	0.0	23.1	34.7#
58	78.4	62.2	93.4









#3
n-Nitrosodimethylamine
Concen: 0.359 ng
RT: 3.714 min Scan# 69
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 8942
Ion Ratio Lower Upper
42 100
74 143.0 114.4 171.6
44 7.9 3.6 5.4#



#4
2-Fluorophenol
Concen: 0.386 ng
RT: 5.383 min Scan# 250
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50

Tgt Ion: 112 Resp: 35238
Ion Ratio Lower Upper
112 100
64 53.0 42.2 63.4
63 27.9 22.2 33.2









#5
Phenol-d6
Concen: 0.383 ng
RT: 6.951 min Scan# 420
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 42308
Ion Ratio Lower Upper
99 100
42 16.7 13.2 19.8
71 29.9 23.8 35.8



#6
bis(2-Chloroethyl)ether
Concen: 0.412 ng
RT: 7.218 min Scan# 449
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50


Tgt Ion: 93 Resp: 41312
Ion Ratio Lower Upper
93 100
63 74.1 58.8 88.2
95 33.2 26.6 39.8








#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.571 min Scan# 746
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:	136	Resp:	166014
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.7	13.1
54	4.9	3.9	5.9
68	3.9	3.3	4.9



#8
Nitrobenzene-d5
Concen: 0.359 ng
RT: 8.924 min Scan# 616
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50

Tgt Ion:	82	Resp:	38322
Ion	Ratio	Lower	Upper
82	100		
128	33.2	26.2	39.4
54	55.9	45.1	67.7









#9
Naphthalene
Concen: 0.398 ng
RT: 10.622 min Scan# 750
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:128 Resp: 172886
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 12.6 10.2 15.2



#10
Hexachlorobutadiene
Concen: 0.393 ng
RT: 10.913 min Scan# 773
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50

Tgt Ion:225 Resp: 29867
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.8 51.5 77.3









#11
2-Methylnaphthalene-d10
Concen: 0.399 ng
RT: 12.158 min Scan# 966
Delta R.T. -0.005 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 99387
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.402 ng
RT: 12.234 min Scan# 983
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50

Tgt Ion:142 Resp: 112518
Ion Ratio Lower Upper
142 100
141 88.1 70.5 105.7
115 33.9 27.4 41.0









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.408 min Scan# 1194
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 86212
Ion Ratio Lower Upper
164 100
162 104.1 84.2 126.2
160 48.4 39.8 59.6



#14
2,4,6-Tribromophenol
Concen: 0.329 ng
RT: 15.905 min Scan# 1312
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50

Tgt Ion:330 Resp: 10292
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 34.8 28.2 42.2









#15
2-Fluorobiphenyl
Concen: 0.386 ng
RT: 13.038 min Scan# 1086
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:172 Resp: 132792
Ion Ratio Lower Upper
172 100
171 34.5 27.8 41.8
170 23.1 18.7 28.1



#16
Acenaphthylene
Concen: 0.362 ng
RT: 14.129 min Scan# 1172
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50

Tgt Ion:152 Resp: 128757
Ion Ratio Lower Upper
152 100
151 19.5 15.7 23.5
153 13.1 10.5 15.7









#17
Acenaphthene
Concen: 0.384 ng
RT: 14.472 min Scan# 1199
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 95170
Ion Ratio Lower Upper
154 100
153 112.2 90.5 135.7
152 55.7 44.4 66.6



#18
Fluorene
Concen: 0.387 ng
RT: 15.461 min Scan# 1277
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50

Tgt Ion:166 Resp: 114833
Ion Ratio Lower Upper
166 100
165 98.0 78.2 117.2
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.163 min Scan# 1415
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 159310
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 9.2 7.7 11.5



#20
4,6-Dinitro-2-methylphenol
Concen: 0.375 ng
RT: 15.537 min Scan# 1283
Delta R.T. -0.001 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50

Tgt Ion:198 Resp: 2909
Ion Ratio Lower Upper
198 100
182 13.2 11.1 16.7
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.374 ng
RT: 16.360 min Scan# 1349
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion: 248 Resp: 34997




Ion	Ratio	Lower	Upper
248	100		
250	94.2	75.8	113.6
141	62.0	52.8	79.2




#22
Hexachlorobenzene
Concen: 0.388 ng
RT: 16.470 min Scan# 1358
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50

Tgt Ion: 284 Resp: 41749

Ion	Ratio	Lower	Upper
284	100		
142	36.9	29.4	44.2
249	29.5	23.8	35.6









#23
Atrazine
Concen: 0.424 ng
RT: 16.628 min Scan# 1371
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	200	Resp:	24426
Ion Ratio	Lower	Upper	
200	100		
173	29.6	25.0	37.6
215	46.3	36.3	54.5



#24
Pentachlorophenol
Concen: 0.402 ng
RT: 16.811 min Scan# 1386
Delta R.T. -0.012 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50

Tgt Ion:	266	Resp:	13368
Ion Ratio	Lower	Upper	
266	100		
264	62.2	49.7	74.5
268	63.6	51.8	77.6









#25
Phenanthrene
Concen: 0.386 ng
RT: 17.200 min Scan# 1418
Delta R.T. -0.012 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 182931
Ion Ratio Lower Upper
178 100
176 18.9 15.1 22.7
179 15.4 12.2 18.4



#26
Anthracene
Concen: 0.366 ng
RT: 17.297 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50

Tgt Ion:178 Resp: 145695
Ion Ratio Lower Upper
178 100
176 18.3 14.6 22.0
179 15.2 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.377 ng
RT: 19.207 min Scan# 1725
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 156169
Ion Ratio Lower Upper
212 100
106 13.6 10.8 16.2
104 7.6 6.0 9.0



#28
Fluoranthene
Concen: 0.389 ng
RT: 19.236 min Scan# 1731
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50

Tgt Ion:202 Resp: 191694
Ion Ratio Lower Upper
202 100
101 11.4 9.0 13.6
203 17.4 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.365 min Scan# 2014
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 144109
Ion Ratio Lower Upper
240 100
120 13.5 11.4 17.0
236 26.9 21.2 31.8



#30
Pyrene
Concen: 0.388 ng
RT: 19.599 min Scan# 1804
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50

Tgt Ion:202 Resp: 186496
Ion Ratio Lower Upper
202 100
200 20.5 16.5 24.7
203 17.4 13.9 20.9









#31
Terphenyl-d14
Concen: 0.385 ng
RT: 19.812 min Scan# 1847
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:244 Resp: 137424
Ion Ratio Lower Upper
244 100
212 7.4 5.9 8.9
122 11.7 9.7 14.5



#32
Benzo(a)anthracene
Concen: 0.367 ng
RT: 21.355 min Scan# 2013
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50

Tgt Ion:228 Resp: 161855
Ion Ratio Lower Upper
228 100
226 26.2 21.2 31.8
229 19.8 15.6 23.4









#33
Chrysene
Concen: 0.395 ng
RT: 21.408 min Scan# 2018
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:228 Resp: 188327
Ion Ratio Lower Upper
228 100
226 29.0 23.3 34.9
229 19.7 15.7 23.5



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.422 ng
RT: 21.302 min Scan# 2008
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50

Tgt Ion:149 Resp: 47909
Ion Ratio Lower Upper
149 100
167 26.6 21.2 31.8
279 4.0 3.2 4.8









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.710 min Scan# 2493
Delta R.T. -0.004 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:264 Resp: 134831
Ion Ratio Lower Upper
264 100
260 24.5 19.5 29.3
265 23.6 19.2 28.8



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.370 ng
RT: 26.110 min Scan# 3194
Delta R.T. -0.004 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50

Tgt Ion:276 Resp: 176973
Ion Ratio Lower Upper
276 100
138 25.3 19.6 29.4
277 24.9 20.0 30.0









#37
Benzo(b)fluoranthene
Concen: 0.388 ng
RT: 23.002 min Scan# 2286
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 181643
Ion Ratio Lower Upper
252 100
253 22.3 17.7 26.5
125 10.3 8.2 12.4



#38
Benzo(k)fluoranthene
Concen: 0.391 ng
RT: 23.050 min Scan# 2300
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50

Tgt Ion:252 Resp: 189641
Ion Ratio Lower Upper
252 100
253 22.7 17.8 26.6
125 11.1 8.3 12.5









#39
Benzo(a)pyrene
Concen: 0.371 ng
RT: 23.608 min Scan# 2463
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 144370
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.8
125 11.3 9.0 13.6



#40
Dibenzo(a,h)anthracene
Concen: 0.368 ng
RT: 26.130 min Scan# 3200
Delta R.T. -0.000 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50

Tgt Ion:278 Resp: 147082
Ion Ratio Lower Upper
278 100
139 18.4 13.1 19.7
279 24.0 19.1 28.7





#41
Benzo(g,h,i)perylene
Concen: 0.372 ng
RT: 26.842 min Scan# 3408
Delta R.T. -0.004 min
Lab File: BN015768.D
Acq: 03 Aug 2021 14:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:276 Resp: 155188
Ion Ratio Lower Upper
276 100
277 23.9 19.0 28.6
138 20.9 16.6 24.8

