


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080322\
 Data File : BN020973.D
 Acq On : 03 Aug 2022 11:22
 Operator : CG/JU
 Sample : DFTPP273
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP273

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080322.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Aug 03 16:31:11 2022



AutoFind: Scans 2495, 2496, 2497; Background Corrected with Scan 2488

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	46.6	78549	PASS
68	69	0.00	2	1.9	1395	PASS
69	198	0.00	100	42.9	72371	PASS
70	69	0.00	2	0.5	368	PASS
127	198	10	80	53.3	89784	PASS
197	198	0.00	2	0.4	640	PASS
198	198	100	100	100.0	168533	PASS
199	198	5	9	6.7	11374	PASS
275	198	10	60	25.1	42245	PASS
365	198	1	100	3.6	5993	PASS
441	198	0.01	100	13.3	22351	PASS
442	198	50	100	83.5	140691	PASS
443	442	15	24	20.6	28939	PASS

m/z	Abundance
39.90	424.0
44.00	1998.0
49.00	269.0
73.00	306.0
206.90	821.0
280.90	155.0

Instrument :

BNA_N

ClientSampleId :

DFTPP273

m/z	Abundance
36.90	416.0
38.00	1120.0
39.00	6600.0
40.00	635.0
41.10	186.0
44.00	1658.0
44.90	224.0
49.00	814.0
50.00	21984.0
51.00	85024.0
52.00	4379.0
52.90	260.0
55.00	600.0
55.90	2617.0
57.00	5814.0
57.90	264.0
61.00	1163.0
62.00	974.0
63.00	3358.0
64.00	475.0
65.00	1672.0
68.10	1597.0
69.00	79432.0
70.00	429.0
72.90	986.0
74.00	8365.0
75.00	12724.0
76.00	4471.0
77.00	86280.0
78.00	6081.0
79.00	5808.0
80.00	4622.0
81.00	6628.0
82.00	1699.0
83.00	1577.0
83.90	248.0
84.90	1110.0
86.00	1729.0
86.90	899.0
88.00	270.0
88.90	173.0
90.90	1434.0
92.00	1995.0
92.90	11173.0
94.00	669.0
94.90	169.0
96.00	494.0
97.00	208.0
98.00	8657.0
99.00	7367.0
99.90	646.0
101.00	3444.0
102.90	1069.0
103.90	2273.0
105.00	1994.0
106.00	794.0
107.00	26560.0
108.00	4138.0
108.90	1024.0
110.00	48464.0
111.00	7764.0
111.90	1077.0
112.90	349.0
116.00	1721.0
117.00	24216.0
117.90	1753.0
118.90	238.0
120.00	362.0
122.00	1754.0
122.90	2769.0
123.90	1266.0
124.90	1145.0
127.00	91744.0
128.00	7482.0
128.90	36744.0
130.00	2971.0
130.90	488.0
131.90	322.0
132.90	220.0
134.00	1144.0
134.90	3258.0
135.90	1475.0
137.00	1777.0
137.80	375.0
138.90	261.0
139.90	361.0
141.00	4790.0
142.00	1857.0
142.90	1145.0
144.00	270.0
144.90	296.0
145.90	795.0

Instrument :
BNA_N
ClientSampleId :
DFTPP273

147.00	2515.0
148.00	5643.0
148.90	1310.0
150.00	355.0
151.20	788.0
151.80	358.0
152.90	1575.0
153.90	1189.0
154.90	2333.0
156.00	3989.0
157.00	746.0
157.80	792.0
158.90	616.0
159.90	1323.0
161.00	2182.0
162.00	555.0
162.90	197.0
164.90	1832.0
165.90	1498.0
167.00	8792.0
168.00	6166.0
169.00	863.0
169.90	277.0
170.90	404.0
171.90	957.0
173.00	1037.0
173.90	1775.0
175.00	3709.0
175.90	909.0
176.90	1688.0
177.90	752.0
179.00	6972.0
180.00	4933.0
180.90	2367.0
181.90	326.0
183.90	703.0
185.00	2816.0
186.00	23176.0
187.00	7265.0
188.00	997.0
189.00	1858.0
189.80	326.0
191.00	857.0
191.90	1916.0
193.10	2445.0
194.00	490.0
194.90	301.0
196.00	4186.0
196.80	1020.0
197.90	160832.0
198.90	11047.0
199.90	821.0
201.30	584.0
202.90	1492.0
204.00	5944.0
205.00	10518.0
206.00	42528.0
207.00	6299.0
207.90	1707.0
208.90	766.0
210.00	716.0
210.90	1500.0
211.70	492.0
214.90	477.0
216.00	987.0
216.90	11191.0
217.90	1491.0
220.20	322.0
221.00	5909.0
221.90	1626.0
223.00	2376.0
224.00	23528.0
225.00	6147.0
226.00	726.0
227.00	10384.0
227.90	1518.0
229.00	2377.0
229.80	404.0
230.90	799.0
231.90	206.0
234.00	604.0
234.90	626.0
236.00	441.0
237.00	1271.0
238.00	245.0
238.90	491.0
239.80	278.0
241.00	489.0
241.90	1167.0
243.00	1420.0
244.00	16880.0
245.00	2349.0
245.90	3654.0
246.90	877.0
248.00	152.0
248.90	551.0

Instrument :
BNA_N
ClientSampleId :
DFTPP273

250.90	200.0
251.80	156.0
252.90	501.0
254.00	965.0
255.00	86488.0
255.90	11965.0
256.90	1251.0
258.00	5566.0
259.00	799.0
259.80	175.0
260.90	167.0
263.90	192.0
264.90	2123.0
265.90	400.0
270.90	189.0
272.00	276.0
272.90	2360.0
274.00	6907.0
275.00	37472.0
276.00	5236.0
277.00	3730.0
277.90	817.0
280.90	170.0
282.90	464.0
283.90	275.0
284.90	556.0
285.90	156.0
293.00	761.0
293.80	185.0
296.00	10744.0
297.00	1102.0
301.80	170.0
302.00	172.0
303.00	1168.0
303.80	305.0
307.70	152.0
314.00	454.0
315.00	1096.0
316.00	850.0
316.90	172.0
320.90	297.0
321.90	155.0
322.90	2959.0
323.90	466.0
326.00	158.0
326.90	836.0
327.90	329.0
331.90	269.0
333.00	415.0
333.90	2371.0
334.90	626.0
340.90	401.0
345.80	649.0
351.90	1003.0
352.90	669.0
354.00	1253.0
354.90	217.0
364.90	4924.0
365.90	857.0
371.00	306.0
371.90	1631.0
373.00	345.0
382.80	490.0
383.80	227.0
390.00	205.0
390.90	187.0
391.90	169.0
402.00	682.0
402.80	821.0
404.00	369.0
420.90	953.0
421.90	608.0
423.00	5401.0
423.90	2326.0
424.80	350.0
440.90	15037.0
442.00	111352.0
443.00	21984.0
443.90	1903.0

Instrument :
BNA_N
ClientSampleId :
DFTPP273

m/z	Abundance
37.00	425.0
38.00	1283.0
39.00	6777.0
40.00	668.0
40.90	178.0
44.00	1734.0
45.00	230.0
49.00	937.0
50.00	24056.0
51.00	90712.0
52.00	4497.0
53.00	187.0
54.90	374.0
56.00	3059.0
57.00	5854.0
57.90	304.0
61.00	1371.0
62.00	1436.0
63.00	3439.0
64.00	559.0
65.00	1770.0
68.00	1471.0
69.00	84192.0
69.90	385.0
73.00	1064.0
74.00	9010.0
75.00	13595.0
76.00	4724.0
77.00	95624.0
78.00	6616.0
79.00	6901.0
80.00	5181.0
81.00	6901.0
81.90	1714.0
83.00	1610.0
83.90	300.0
85.00	1262.0
86.00	2389.0
87.00	1126.0
87.90	269.0
89.00	246.0
90.90	1656.0
91.90	2071.0
93.00	11434.0
93.90	933.0
95.00	193.0
96.00	428.0
96.90	169.0
98.00	9528.0
99.00	7992.0
100.00	667.0
101.00	4105.0
101.90	229.0
102.90	1298.0
104.00	2576.0
104.90	2477.0
106.00	1106.0
107.00	29704.0
108.00	4698.0
109.00	967.0
110.00	55384.0
110.90	8155.0
111.90	1019.0
112.90	347.0
116.00	1839.0
117.00	26152.0
118.00	1959.0
119.00	229.0
119.90	532.0
121.90	1993.0
122.90	3322.0
123.90	1438.0
125.00	1467.0
127.00	105448.0
128.00	7958.0
128.90	41864.0
130.00	3368.0
130.90	763.0
131.80	467.0
132.70	234.0
133.90	1248.0
135.00	3379.0
135.90	1219.0
137.00	2287.0
137.90	349.0
139.00	244.0
139.90	516.0
141.00	5620.0
141.90	1658.0
142.90	1215.0
144.00	434.0
144.80	340.0

Instrument :
BNA_N
ClientSampleId :
DFTPP273

145.90	1001.0
146.90	2961.0
148.00	6742.0
149.00	1544.0
150.00	383.0
151.00	780.0
151.80	396.0
153.00	1689.0
153.90	1239.0
155.00	3185.0
156.00	4934.0
157.00	860.0
157.90	763.0
158.90	744.0
159.90	1859.0
160.90	2383.0
162.00	887.0
163.00	249.0
163.90	344.0
164.90	1995.0
165.90	1586.0
167.00	10757.0
168.00	6467.0
168.90	967.0
170.00	402.0
170.90	476.0
171.90	976.0
172.90	1233.0
173.90	2311.0
175.00	4372.0
176.00	1055.0
176.90	1767.0
177.90	756.0
178.90	7539.0
180.00	5753.0
181.00	2474.0
182.00	459.0
182.90	177.0
183.90	595.0
184.90	3729.0
186.00	27696.0
187.00	8324.0
187.90	976.0
188.90	1929.0
189.90	314.0
191.00	902.0
192.00	2182.0
193.00	2692.0
194.00	578.0
194.90	334.0
196.00	5221.0
197.90	202304.0
199.00	13754.0
200.00	1093.0
201.50	855.0
202.90	1443.0
204.00	7959.0
205.00	11955.0
206.00	53152.0
207.00	7816.0
207.90	2044.0
209.00	746.0
210.20	847.0
210.80	2144.0
212.00	460.0
212.90	187.0
215.00	845.0
215.90	957.0
216.90	14210.0
217.90	1804.0
221.00	6975.0
221.70	1907.0
223.00	3078.0
224.00	27816.0
225.00	6697.0
225.90	902.0
226.90	11953.0
227.90	1772.0
229.00	2728.0
229.90	395.0
231.00	992.0
232.00	214.0
232.90	270.0
233.80	705.0
234.90	880.0
235.90	642.0
237.00	1276.0
237.90	234.0
238.90	543.0
239.80	352.0
241.00	818.0
241.90	1464.0
243.00	1528.0
244.00	20392.0
245.00	2530.0
245.90	4573.0

Instrument :
BNA_N
ClientSampleId :
DFTPP273

246.90	938.0
247.90	211.0
249.00	713.0
252.00	245.0
253.00	478.0
254.00	931.0
255.00	114208.0
256.00	16776.0
256.90	1421.0
258.00	7000.0
258.90	1047.0
259.90	220.0
261.00	183.0
265.00	2849.0
265.90	449.0
270.90	230.0
271.90	261.0
273.00	3628.0
274.00	9213.0
275.00	51184.0
276.00	6550.0
276.90	4470.0
277.90	809.0
279.10	156.0
281.90	169.0
283.00	505.0
283.90	362.0
285.00	740.0
286.00	163.0
288.80	224.0
291.80	294.0
292.90	934.0
293.90	234.0
295.90	15871.0
297.00	2093.0
298.00	189.0
301.00	199.0
301.80	184.0
302.90	1696.0
303.90	565.0
307.80	205.0
308.90	159.0
309.80	229.0
314.00	655.0
314.90	1506.0
316.00	931.0
321.00	527.0
323.00	4214.0
324.00	795.0
327.00	928.0
327.90	414.0
332.00	402.0
333.00	409.0
334.00	3291.0
334.90	863.0
340.90	511.0
341.90	153.0
346.00	910.0
346.90	240.0
351.90	1283.0
352.90	1040.0
354.00	1363.0
355.00	378.0
364.90	7438.0
365.80	976.0
371.00	335.0
372.00	2429.0
373.00	532.0
382.90	559.0
384.00	377.0
389.90	227.0
390.90	241.0
391.90	226.0
401.90	957.0
402.90	1170.0
404.00	690.0
420.90	1124.0
422.00	1332.0
423.00	7552.0
423.90	2601.0
424.90	465.0
441.00	25336.0
441.90	166272.0
443.00	35712.0
444.00	2935.0

Instrument :
BNA_N
ClientSampleId :
DFTPP273

m/z	Abundance
36.90	202.0
38.00	877.0
39.00	4245.0
39.90	586.0
44.00	1848.0
49.00	668.0
50.00	16104.0
51.00	59912.0
52.00	3080.0
54.90	259.0
56.00	1741.0
57.00	3704.0
61.00	748.0
61.90	751.0
63.00	2560.0
64.00	368.0
65.00	1247.0
68.00	1117.0
69.00	53488.0
69.90	289.0
73.00	805.0
74.00	5846.0
75.00	9407.0
76.00	2947.0
77.00	63240.0
78.00	4158.0
79.00	4301.0
80.00	3107.0
81.00	4819.0
82.00	928.0
83.00	1067.0
84.00	247.0
84.90	940.0
86.00	1333.0
86.90	675.0
87.90	224.0
90.90	1049.0
92.00	1104.0
93.00	7629.0
93.90	557.0
94.90	157.0
96.00	460.0
97.00	154.0
98.00	6292.0
99.00	5065.0
99.90	470.0
101.00	2957.0
102.10	158.0
103.00	771.0
104.00	1733.0
104.90	1661.0
106.00	670.0
107.00	21528.0
107.90	3438.0
109.00	718.0
110.00	37256.0
110.90	5661.0
111.90	733.0
112.90	304.0
116.00	1160.0
117.00	18472.0
118.00	1466.0
119.90	325.0
121.90	1378.0
122.90	2418.0
123.90	1057.0
124.90	878.0
126.10	636.0
127.00	72160.0
128.00	5361.0
129.00	28432.0
130.00	2238.0
131.00	571.0
132.00	237.0
133.00	195.0
133.90	974.0
135.00	2800.0
136.00	903.0
137.00	1466.0
137.90	355.0
139.90	514.0
140.90	3239.0
141.90	1578.0
142.90	784.0
144.00	245.0
145.90	889.0
147.00	2408.0
148.00	4922.0
148.90	910.0
150.00	294.0
150.90	510.0
151.70	334.0

Instrument :

BNA_N

ClientSampleId :

DFTPP273

152.90	1101.0
153.90	943.0
155.00	2228.0
156.00	2987.0
156.90	554.0
157.90	600.0
158.90	483.0
160.00	1451.0
161.00	1715.0
161.90	540.0
164.90	1643.0
166.00	1317.0
167.00	7965.0
167.90	4481.0
169.00	735.0
169.90	163.0
171.00	261.0
171.90	810.0
173.00	906.0
173.90	1626.0
175.00	3155.0
175.90	962.0
177.00	1591.0
177.90	456.0
178.90	5813.0
180.00	4061.0
180.90	1744.0
181.90	288.0
183.90	476.0
185.00	2723.0
186.00	21432.0
187.00	6631.0
188.10	562.0
188.90	1295.0
189.90	277.0
190.90	689.0
192.00	1987.0
193.00	2029.0
194.00	446.0
194.80	212.0
195.90	3988.0
196.70	901.0
197.00	905.0
197.90	142464.0
198.90	9322.0
200.00	722.0
201.40	835.0
202.90	1247.0
204.00	5484.0
205.00	8606.0
206.00	39504.0
207.00	5852.0
207.90	1629.0
208.90	592.0
210.10	725.0
211.00	1373.0
215.00	453.0
216.00	878.0
216.90	10442.0
218.00	1165.0
220.00	233.0
221.00	5047.0
221.80	1346.0
222.90	2304.0
224.00	22136.0
225.00	5630.0
226.10	690.0
227.00	8913.0
227.90	1116.0
228.90	1864.0
229.80	279.0
231.00	821.0
233.80	737.0
234.90	660.0
236.00	440.0
236.90	1057.0
238.90	374.0
239.90	281.0
240.90	516.0
242.00	1122.0
243.00	1197.0
244.00	15552.0
245.00	2505.0
246.00	3309.0
246.80	675.0
247.80	209.0
248.90	556.0
252.00	162.0
252.80	324.0
255.00	88784.0
256.00	12708.0
257.00	814.0
258.00	5459.0
259.00	762.0
260.10	179.0
260.80	181.0

Instrument :
BNA_N
ClientSampleId :
DFTPP273

264.00	193.0
265.00	2249.0
265.90	334.0
271.00	266.0
271.90	282.0
273.00	2531.0
274.00	7487.0
275.00	38080.0
276.00	5046.0
276.90	3464.0
278.00	730.0
278.90	178.0
283.00	427.0
284.00	329.0
284.90	603.0
288.90	156.0
292.90	1113.0
294.10	168.0
294.90	304.0
296.00	12548.0
297.00	1767.0
301.90	222.0
302.90	1291.0
303.90	495.0
307.90	200.0
310.00	191.0
313.90	597.0
315.00	1379.0
315.90	745.0
320.90	502.0
322.00	228.0
323.00	3706.0
324.00	774.0
326.80	768.0
327.80	385.0
331.80	290.0
333.00	411.0
334.00	2407.0
335.00	635.0
340.90	371.0
341.90	241.0
345.90	699.0
346.90	163.0
352.00	758.0
353.00	926.0
353.90	1222.0
355.00	278.0
365.00	5616.0
365.90	691.0
370.90	387.0
371.90	2077.0
372.90	582.0
382.90	561.0
383.90	170.0
389.90	245.0
401.90	862.0
402.90	1074.0
403.90	425.0
420.90	1127.0
422.00	1306.0
423.00	8168.0
424.10	2025.0
424.90	442.0
441.00	26680.0
442.00	144448.0
442.90	29120.0
444.00	2874.0

Instrument :
BNA_N
ClientSampleId :
DFTPP273