

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080324\
 Data File : BN033219.D
 Acq On : 03 Aug 2024 01:51
 Operator : MA/JU
 Sample : P3416-01
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MLS-15-106-121

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/05/2024
 Supervised By :mohammad ahmed 08/07/2024

Quant Time: Aug 03 04:14:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN071124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 02 22:49:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.532	152	3507m	0.400	ng	0.03	
7) Naphthalene-d8	10.297	136	10943	0.400	ng	# 0.01	
13) Acenaphthene-d10	14.158	164	6793	0.400	ng	0.02	
19) Phenanthrene-d10	16.932	188	13060	0.400	ng	# 0.01	
29) Chrysene-d12	21.142	240	8358	0.400	ng	0.00	
35) Perylene-d12	23.320	264	9800	0.400	ng	# 0.00	
System Monitoring Compounds							
4) 2-Fluorophenol	5.184	112	725	0.082	ng	-0.06	
5) Phenol-d6	7.055	99	554m	0.050	ng	0.17	
8) Nitrobenzene-d5	8.931	82	1775m	0.215	ng	0.12	
11) 2-Methylnaphthalene-d10	11.935	152	2497	0.149	ng	0.03	
14) 2,4,6-Tribromophenol	15.728	330	370	0.127	ng	0.01	
15) 2-Fluorobiphenyl	12.800	172	3778	0.149	ng	0.01	
27) Fluoranthene-d10	18.969	212	6238	0.185	ng	0.00	
31) Terphenyl-d14	19.559	244	4992	0.291	ng	0.00	
Target Compounds							
33) Chrysene	21.187	228	732	0.023	ng	# 71	Qvalue
34) Bis(2-ethylhexyl)phtha...	21.008	149	3822	0.261	ng	# 81	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080324\
 Data File : BN033219.D
 Acq On : 03 Aug 2024 01:51
 Operator : MA/JU
 Sample : P3416-01
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MLS-15-106-121

Quant Time: Aug 03 04:14:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN071124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 02 22:49:47 2024
 Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2024
 Supervised By :mohammad ahmed 08/07/2024

