

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080324\
 Data File : BN033223.D
 Acq On : 03 Aug 2024 04:15
 Operator : MA/JU
 Sample : P3426-02
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 927-K1-WS-073124-FD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/05/2024
 Supervised By :mohammad ahmed 08/07/2024

Quant Time: Aug 03 04:44:04 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN071124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 02 22:49:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.510	152	4557	0.400	ng	0.00
7) Naphthalene-d8	10.266	136	15575	0.400	ng	#-0.02
13) Acenaphthene-d10	14.137	164	7686	0.400	ng	0.00
19) Phenanthrene-d10	16.908	188	15096	0.400	ng	-0.01
29) Chrysene-d12	21.125	240	13422	0.400	ng	0.00
35) Perylene-d12	23.300	264	14336	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.221	112	1475m	0.128	ng	-0.02
5) Phenol-d6	6.874	99	2178m	0.152	ng	0.00
8) Nitrobenzene-d5	8.813	82	1975	0.168	ng	0.00
11) 2-Methylnaphthalene-d10	11.871	152	3091	0.130	ng	-0.03
14) 2,4,6-Tribromophenol	15.679	330	497	0.151	ng	-0.04
15) 2-Fluorobiphenyl	12.768	172	4807	0.168	ng	-0.02
27) Fluoranthene-d10	18.955	212	6725	0.172	ng	0.00
31) Terphenyl-d14	19.540	244	15041	0.547	ng	-0.02
Target Compounds						
2) 1,4-Dioxane	3.083	88	626	0.111	ng	# 52
34) Bis(2-ethylhexyl)phtha...	21.008	149	2442	0.104	ng	# 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080324\
Data File : BN033223.D
Acq On : 03 Aug 2024 04:15
Operator : MA/JU
Sample : P3426-02
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
927-K1-WS-073124-FD

Quant Time: Aug 03 04:44:04 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN071124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 02 22:49:47 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/05/2024
Supervised By :mohammad ahmed 08/07/2024

