

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080324\
 Data File : BN033237.D
 Acq On : 03 Aug 2024 14:00
 Operator : MA/JU
 Sample : P3426-02RE
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 927-K1-WS-073124-FDRE

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 08/05/2024
 Supervised By :mohammad ahmed 08/07/2024

Quant Time: Aug 05 03:27:27 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN071124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 02 22:49:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.517	152	4565	0.400	ng	0.01
7) Naphthalene-d8	10.265	136	16642	0.400	ng	#-0.02
13) Acenaphthene-d10	14.137	164	9402	0.400	ng	0.00
19) Phenanthrene-d10	16.908	188	19875	0.400	ng	-0.01
29) Chrysene-d12	21.134	240	14622	0.400	ng	# 0.00
35) Perylene-d12	23.306	264	14874	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.228	112	1472m	0.127	ng	-0.01
5) Phenol-d6	6.874	99	2139m	0.149	ng	0.00
8) Nitrobenzene-d5	8.803	82	2074m	0.166	ng	-0.01
11) 2-Methylnaphthalene-d10	11.871	152	3396	0.133	ng	-0.03
14) 2,4,6-Tribromophenol	15.679	330	615	0.153	ng	-0.04
15) 2-Fluorobiphenyl	12.758	172	5686	0.162	ng	-0.03
27) Fluoranthene-d10	18.950	212	9116	0.177	ng	0.00
31) Terphenyl-d14	19.540	244	19950	0.666	ng	-0.02
Target Compounds						
2) 1,4-Dioxane	3.083	88	642	0.114	ng	# 54
34) Bis(2-ethylhexyl)phtha...	21.008	149	3357	0.131	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080324\
Data File : BN033237.D
Acq On : 03 Aug 2024 14:00
Operator : MA/JU
Sample : P3426-02RE
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
927-K1-WS-073124-FDRE

Quant Time: Aug 05 03:27:27 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN071124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 02 22:49:47 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/05/2024
Supervised By :mohammad ahmed 08/07/2024

