

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080422\
 Data File : BN020991.D
 Acq On : 04 Aug 2022 16:13
 Operator : CG/JU
 Sample : N3900-16
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0CB3

Quant Time: Aug 05 01:21:51 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 03 16:31:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.801	152	2433	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.596	136	7796	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.457	164	4711	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.206	188	9298	0.400	ng/ul	0.00
17) Chrysene-d12	21.430	240	7375	0.400	ng/ul	0.01
23) Perylene-d12	23.768	264	8104	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.207	96	11991	3.852	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.202	152	4147	0.363	ng/ul	0.00
18) Fluoranthene-d10	19.249	212	9912	0.435	ng/ul	0.00
Target Compounds						
					Qvalue	
5) Naphthalene	10.646	128	552	0.023	ng/ul#	68
7) 2-Methylnaphthalene	12.274	142	666	0.046	ng/ul	99
8) 1-Methylnaphthalene	12.494	142	426	0.029	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080422\
Data File : BN020991.D
Acq On : 04 Aug 2022 16:13
Operator : CG/JU
Sample : N3900-16
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0CB3

Quant Time: Aug 05 01:21:51 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 03 16:31:11 2022
Response via : Initial Calibration

