

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080422\
 Data File : BN021001.D
 Acq On : 04 Aug 2022 22:33
 Operator : CG/JU
 Sample : N3900-01
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0C80

Quant Time: Aug 05 01:22:39 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 03 16:31:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.805	152	2488	0.400	ng/ul	0.00
4) Naphthalene-d8	10.596	136	8213	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.457	164	4935	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.205	188	9926	0.400	ng/ul	0.00
17) Chrysene-d12	21.415	240	7869	0.400	ng/ul	0.00
23) Perylene-d12	23.751	264	8791	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.206	96	12042	3.783	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.202	152	4298	0.357	ng/ul	0.00
18) Fluoranthene-d10	19.245	212	10224	0.421	ng/ul	0.00
Target Compounds						
					Qvalue	
5) Naphthalene	10.651	128	1025	0.041	ng/ul#	65
8) 1-Methylnaphthalene	12.499	142	378	0.025	ng/ul	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080422\
Data File : BN021001.D
Acq On : 04 Aug 2022 22:33
Operator : CG/JU
Sample : N3900-01
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0C80

Quant Time: Aug 05 01:22:39 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 03 16:31:11 2022
Response via : Initial Calibration

