

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080422\
 Data File : BN021029.D
 Acq On : 05 Aug 2022 16:31
 Operator : CG/JU
 Sample : N3952-01
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 COCA0

Quant Time: Aug 06 00:28:21 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 03 16:31:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.801	152	2263	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.596	136	7283	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.452	164	4211	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.205	188	8273	0.400	ng/ul	0.00
17) Chrysene-d12	21.421	240	6661	0.400	ng/ul	0.00
23) Perylene-d12	23.756	264	7246	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.206	96	10528	3.636	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.207	152	3467	0.325	ng/ul	0.00
18) Fluoranthene-d10	19.244	212	8298	0.403	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080422\
Data File : BN021029.D
Acq On : 05 Aug 2022 16:31
Operator : CG/JU
Sample : N3952-01
Misc :
ALS Vial : 49 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0CA0

Quant Time: Aug 06 00:28:21 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 03 16:31:11 2022
Response via : Initial Calibration

