

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080422\
 Data File : BN021039.D
 Acq On : 05 Aug 2022 23:27
 Operator : CG/JU
 Sample : PB146768BL
 Misc :
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK768

Quant Time: Aug 06 00:29:08 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 03 16:31:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.801	152	2615	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.596	136	8222	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.452	164	4787	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.206	188	9322	0.400	ng/ul	0.00
17) Chrysene-d12	21.415	240	7036	0.400	ng/ul	0.00
23) Perylene-d12	23.745	264	7784	0.400	ng/ul	#-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.202	96	19710	5.892	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.208	152	4021	0.334	ng/ul	0.00
18) Fluoranthene-d10	19.245	212	8529	0.393	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080422\
Data File : BN021039.D
Acq On : 05 Aug 2022 23:27
Operator : CG/JU
Sample : PB146768BL
Misc :
ALS Vial : 60 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK768

Quant Time: Aug 06 00:29:08 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 03 16:31:11 2022
Response via : Initial Calibration

