

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080422\
 Data File : BN021014.D
 Acq On : 05 Aug 2022 07:14
 Operator : CG/JU
 Sample : N4026-01
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GBJG6

Quant Time: Aug 05 07:54:20 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 03 16:31:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.801	152	3282	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.596	136	10293	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.452	164	5983	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.205	188	11804	0.400	ng/ul	0.00
17) Chrysene-d12	21.418	240	8774	0.400	ng/ul	0.00
23) Perylene-d12	23.751	264	9172	0.400	ng/ul	#-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.206	96	26441	6.297	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.207	152	4758	0.315	ng/ul	0.00
18) Fluoranthene-d10	19.245	212	11533	0.426	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080422\
Data File : BN021014.D
Acq On : 05 Aug 2022 07:14
Operator : CG/JU
Sample : N4026-01
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GBJG6

Quant Time: Aug 05 07:54:20 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 03 16:31:11 2022
Response via : Initial Calibration

