

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080422\
 Data File : BN021040.D
 Acq On : 06 Aug 2022 00:05
 Operator : CG/JU
 Sample : N3951-07
 Misc :
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0CF8

Quant Time: Aug 06 00:54:21 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 03 16:31:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.801	152	2603	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.595	136	8378	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.455	164	4870	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.204	188	9648	0.400	ng/ul	0.00
17) Chrysene-d12	21.413	240	8080	0.400	ng/ul	0.00
23) Perylene-d12	23.745	264	8474	0.400	ng/ul	#-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.207	96	12819	3.849	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.206	152	4008	0.327	ng/ul	0.00
18) Fluoranthene-d10	19.243	212	9751	0.391	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080422\
Data File : BN021040.D
Acq On : 06 Aug 2022 00:05
Operator : CG/JU
Sample : N3951-07
Misc :
ALS Vial : 61 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0CF8

Quant Time: Aug 06 00:54:21 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 03 16:31:11 2022
Response via : Initial Calibration

