

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080422\
 Data File : BN021059.D
 Acq On : 06 Aug 2022 12:28
 Operator : CG/JU
 Sample : N4026-21
 Misc :
 ALS Vial : 81 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GBJJ6

Quant Time: Aug 08 01:38:05 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 03 16:31:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.797	152	2745	0.400	ng/ul	-0.02
4) Naphthalene-d8	10.590	136	9062	0.400	ng/ul	-0.02
9) Acenaphthene-d10	14.451	164	5228	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.200	188	10413	0.400	ng/ul	0.00
17) Chrysene-d12	21.415	240	8555	0.400	ng/ul	0.00
23) Perylene-d12	23.748	264	9143	0.400	ng/ul	#-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.202	96	11874	3.381	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.201	152	3989	0.300	ng/ul	0.00
18) Fluoranthene-d10	19.243	212	8967	0.339	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080422\
Data File : BN021059.D
Acq On : 06 Aug 2022 12:28
Operator : CG/JU
Sample : N4026-21
Misc :
ALS Vial : 81 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GBJJ6

Quant Time: Aug 08 01:38:05 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 03 16:31:11 2022
Response via : Initial Calibration

