

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080525\
 Data File : BN037568.D
 Acq On : 05 Aug 2025 15:03
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 06 01:40:14 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN071525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 19 01:46:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	69	0.00
2	1,4-Dioxane	0.400	0.404	-1.0	67	0.00
3	n-Nitrosodimethylamine	0.400	0.458	-14.5	82	0.00
4 S	2-Fluorophenol	0.400	0.340	15.0	59	0.00
5 S	Phenol-d6	0.400	0.328	18.0	59	0.00
6	bis(2-Chloroethyl)ether	0.400	0.365	8.8	63	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	65	-0.01
8 S	Nitrobenzene-d5	0.400	0.386	3.5	67	-0.01
9	Naphthalene	0.400	0.385	3.8	64	0.00
10	Hexachlorobutadiene	0.400	0.476	-19.0	78	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.352	12.0	61	-0.01
12	2-Methylnaphthalene	0.400	0.354	11.5	60	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	57	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.295	26.3#	47	-0.01
15 S	2-Fluorobiphenyl	0.400	0.406	-1.5	58	0.00
16	Acenaphthylene	0.400	0.376	6.0	55	0.00
17	Acenaphthene	0.400	0.371	7.3	55	-0.01
18	Fluorene	0.400	0.382	4.5	57	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	56	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.433	-8.2	63	0.00
21	4-Bromophenyl-phenylether	0.400	0.361	9.8	54	0.00
22	Hexachlorobenzene	0.400	0.416	-4.0	59	0.00
23	Atrazine	0.400	0.332	17.0	53	-0.01
24	Pentachlorophenol	0.400	0.297	25.8#	49	-0.01
25	Phenanthrene	0.400	0.371	7.3	54	-0.01
26	Anthracene	0.400	0.348	13.0	53	0.00
27 SURR	Fluoranthene-d10	0.400	0.342	14.5	53	0.00
28	Fluoranthene	0.400	0.347	13.3	52	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	54	0.00
30	Pyrene	0.400	0.375	6.3	51	0.00
31 S	Terphenyl-d14	0.400	0.381	4.8	52	0.00
32	Benzo(a)anthracene	0.400	0.358	10.5	50	0.00
33	Chrysene	0.400	0.398	0.5	55	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.374	6.5	56	0.00
35 I	Perylene-d12	0.400	0.400	0.0	54	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.387	3.3	58	0.00
37	Benzo(b)fluoranthene	0.400	0.386	3.5	55	0.00
38	Benzo(k)fluoranthene	0.400	0.415	-3.7	59	0.00
39 C	Benzo(a)pyrene	0.400	0.373	6.8	54	0.00
40	Dibenzo(a,h)anthracene	0.400	0.380	5.0	57	-0.01
41	Benzo(g,h,i)perylene	0.400	0.382	4.5	55	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4