




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN080620\
 Data File : BN011389.D
 Acq On : 06 Aug 2020 21:01
 Operator : CG/JU
 Sample : L3419-15
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Quant Time: Aug 07 02:18:47 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN073020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 07 02:04:45 2020
 Response via : Initial Calibration



TIC: BN011389.D

(21) Benzo(b)fluoranthene

22.557min (0.000) 0.07ng/ul m

response 1415

Ion	Exp%	Act%
252.00	100	100
253.00	25.90	61.87#
125.00	12.10	38.35#
0.00	0.00	0.00