




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080621\
Data File : BN015812.D
Acq On : 06 Aug 2021 17:38
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 07 08:12:17 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 07 08:12:10 2021
Response via : Initial Calibration



TIC: BN015812.D\data.ms

(70) Pentachlorophenol (C)




16.957min (0.000) 74725.83 ng

response 91483

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	63.75
264.00	61.60	62.07
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080621\
Data File : BN015812.D
Acq On : 06 Aug 2021 17:38
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 07 08:12:17 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 07 08:12:10 2021
Response via : Initial Calibration



TIC: BN015812.D\data.ms

(77) Benzidine

19.557min (0.000) 0.00 ng

response 639529

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.74
183.00	13.20	11.96
0.00	0.00	0.00